

# Single-step preparation of selected biological fluids for the high performance liquid chromatographic analysis of fat-soluble vitamins and antioxidants

Lazzarino, Giacomo; Longo, Salvatore; Amorini, Angela Maria; Di Pietro, Valentina; D'Urso, Serena; Lazzarino, Giuseppe; Belli, Antonio; Tavazzi, Barbara

DOI:

[10.1016/j.chroma.2017.10.053](https://doi.org/10.1016/j.chroma.2017.10.053)

License:

Creative Commons: Attribution-NonCommercial-NoDerivs (CC BY-NC-ND)

*Document Version*

Peer reviewed version

*Citation for published version (Harvard):*

Lazzarino, G, Longo, S, Amorini, AM, Di Pietro, V, D'Urso, S, Lazzarino, G, Belli, A & Tavazzi, B 2017, 'Single-step preparation of selected biological fluids for the high performance liquid chromatographic analysis of fat-soluble vitamins and antioxidants', *Journal of Chromatography A*, vol. 1527, pp. 43-52.  
<https://doi.org/10.1016/j.chroma.2017.10.053>

[Link to publication on Research at Birmingham portal](#)

**Publisher Rights Statement:**

<https://doi.org/10.1016/j.chroma.2017.10.053>

**General rights**

Unless a licence is specified above, all rights (including copyright and moral rights) in this document are retained by the authors and/or the copyright holders. The express permission of the copyright holder must be obtained for any use of this material other than for purposes permitted by law.

- Users may freely distribute the URL that is used to identify this publication.
- Users may download and/or print one copy of the publication from the University of Birmingham research portal for the purpose of private study or non-commercial research.
- User may use extracts from the document in line with the concept of 'fair dealing' under the Copyright, Designs and Patents Act 1988 (?)
- Users may not further distribute the material nor use it for the purposes of commercial gain.

Where a licence is displayed above, please note the terms and conditions of the licence govern your use of this document.

When citing, please reference the published version.

**Take down policy**

While the University of Birmingham exercises care and attention in making items available there are rare occasions when an item has been uploaded in error or has been deemed to be commercially or otherwise sensitive.

If you believe that this is the case for this document, please contact [UBIRA@lists.bham.ac.uk](mailto:UBIRA@lists.bham.ac.uk) providing details and we will remove access to the work immediately and investigate.

## Accepted Manuscript

Title: SINGLE-STEP PREPARATION OF SELECTED BIOLOGICAL FLUIDS FOR THE HIGH PERFORMANCE LIQUID CHROMATOGRAPHIC ANALYSIS OF FAT-SOLUBLE VITAMINS AND ANTIOXIDANTS

Authors: Giacomo Lazzarino, Salvatore Longo, Angela Maria Amorini, Valentina Di Pietro, Serafina D'Urso, Giuseppe Lazzarino, Antonio Belli, Barbara Tavazzi

PII: S0021-9673(17)31559-5  
DOI: <https://doi.org/10.1016/j.chroma.2017.10.053>  
Reference: CHROMA 358958

To appear in: *Journal of Chromatography A*

Received date: 19-7-2017  
Revised date: 26-9-2017  
Accepted date: 22-10-2017

Please cite this article as: Giacomo Lazzarino, Salvatore Longo, Angela Maria Amorini, Valentina Di Pietro, Serafina D'Urso, Giuseppe Lazzarino, Antonio Belli, Barbara Tavazzi, SINGLE-STEP PREPARATION OF SELECTED BIOLOGICAL FLUIDS FOR THE HIGH PERFORMANCE LIQUID CHROMATOGRAPHIC ANALYSIS OF FAT-SOLUBLE VITAMINS AND ANTIOXIDANTS, *Journal of Chromatography A* <https://doi.org/10.1016/j.chroma.2017.10.053>

This is a PDF file of an unedited manuscript that has been accepted for publication. As a service to our customers we are providing this early version of the manuscript. The manuscript will undergo copyediting, typesetting, and review of the resulting proof before it is published in its final form. Please note that during the production process errors may be discovered which could affect the content, and all legal disclaimers that apply to the journal pertain.



**SINGLE-STEP PREPARATION OF SELECTED BIOLOGICAL FLUIDS FOR THE HIGH PERFORMANCE LIQUID CHROMATOGRAPHIC ANALYSIS OF FAT-SOLUBLE VITAMINS AND ANTIOXIDANTS**

Giacomo Lazzarino<sup>a, #</sup>, Salvatore Longo<sup>a, b, #</sup>, Angela Maria Amorini<sup>a, \*</sup>, Valentina Di Pietro<sup>c</sup>,  
Serafina D'Urso<sup>b, d</sup>, Giuseppe Lazzarino<sup>b, d, \*</sup>, Antonio Belli<sup>c, e</sup>, Barbara Tavazzi<sup>a</sup>

<sup>a</sup>Institute of Biochemistry and Clinical Biochemistry, Catholic University of Rome, Largo F. Vito 1, 00168 Rome, Italy

<sup>b</sup>LTA-Biotech srl, Viale Don Orione, 3D, 95047 Paternò (CT), Italy

<sup>c</sup>Neuroscience and Ophthalmology group, School of Clinical and Experimental Medicine, College of Medical and Dental Sciences, University of Birmingham, Edgbaston B15 2TT, Birmingham, UK.

<sup>d</sup>Department of Biomedical and Biotechnological Sciences, Division of Medical Biochemistry, University of Catania, Viale A. Doria 6, 95125 Catania, Italy.

<sup>e</sup>National Institute for Health Research Surgical Reconstruction and Microbiology Research Centre, Queen Elizabeth Hospital, Edgbaston B15 2TH, Birmingham, UK.

<sup>#</sup>These Authors equally contributed to this work

**\*Correspondence should be addressed to:** Dr. Angela Maria Amorini

Institute of Biochemistry and Clinical Biochemistry

Catholic University of Rome, Largo F. Vito 1, 00168 Rome, Italy

Tel.: +39-0630155135; Fax: +39-0630155135; E-mail: [angelamaria.amorini@unicatt.it](mailto:angelamaria.amorini@unicatt.it)

Prof. Giuseppe Lazzarino, Department of Biomedical and Biotechnological Sciences, Division of Medical Biochemistry, University of Catania, Viale A. Doria 6, 95125 Catania, Italy

Tel.: +39-0957384095; Fax: +39-095337036; E-mail: [lazzarig@unict.it](mailto:lazzarig@unict.it)

## Highlights

- Fat-soluble vitamin and antioxidant extraction from biological fluids is currently uneasy
- We set up a single step, single solvent extraction of these compounds from human fluids
- Fat-soluble vitamins and antioxidants were separated by HPLC with UV-DAD
- Extraction has been applied to human serum and seminal plasma from healthy controls
- Recovery, rapidity and simplicity are main advantages over existing methods

## ABSTRACT

Fat-soluble vitamins and antioxidants are of relevance in health and disease. Current methods to extract these compounds from biological fluids mainly need use of multi-steps and multi organic solvents. They are time-consuming and difficult to apply to treat simultaneously large sample number.

We here describe a single-step, one solvent extraction of fat-soluble vitamins and antioxidants from biological fluids, and the chromatographic separation of *all-trans*-retinoic acid, 25-hydroxycholecalciferol, *all-trans*-retinol, astaxanthin, lutein, zeaxanthin, *trans*- $\beta$ -apo-8'-carotenal,  $\gamma$ -tocopherol,  $\beta$ -cryptoxanthin,  $\alpha$ -tocopherol, phylloquinone, lycopene,  $\alpha$ -carotene,  $\beta$ -carotene and coenzyme Q<sub>10</sub>. Extraction is obtained by adding one volume of biological fluid to two acetonitrile volumes, vortexing for 60 seconds and incubating for 60 minutes at 37 °C under agitation. HPLC separation occurs in 30 minutes using Hypersil C18, 100 x 4.6 mm, 5  $\mu$ m particle size column, gradient from 70% methanol + 30% H<sub>2</sub>O to 100% acetonitrile, flow rate of 1.0 mL/min and 37 °C column temperature. Compounds are revealed using highly sensitive UV-VIS diode array detector. The HPLC method suitability was assessed in terms of sensitivity, reproducibility and recovery. Using the present extraction and chromatographic conditions we obtained values of the fat-soluble vitamins and antioxidants in serum from 50 healthy controls similar to those found in literature.

Additionally, the profile of these compounds was also measured in seminal plasma from 20 healthy fertile donors.

Results indicate that this simple, rapid and low cost sample processing is suitable to extract fat-soluble vitamins and antioxidants from biological fluids and can be applied in clinical and nutritional studies.

## KEYWORDS

Fat-soluble vitamins; fat-soluble antioxidants; carotenoids; reversed-phase HPLC; serum; seminal plasma.

## 1 INTRODUCTION

Fat-soluble vitamins are compounds with multiple biological roles and are essential to ensure correct cell functioning that strictly depends on their circulating levels. With the exception of vitamin D<sub>3</sub>, humans are not able to synthesize fat-soluble vitamins (A, E and K) that are generally present in different amounts as pro-vitamins in various types of foods, mostly of vegetal origin. After their ingestion, pro-vitamins usually need one or more biotransformation step to generate the corresponding final active vitamin forms [1, 2]. To ensure adequate levels within the body, a constant daily ingestion of pro-vitamin/vitamin-rich foods with diet is necessary. In this, the so called Mediterranean diet is considered one of the best dietary regimen, ensuring the consumption of a large variety of fruits, vegetables and plant-derived products (such as extra-virgin olive oil) containing relevant amounts of most fat-soluble pro-vitamins/vitamins [3].

It is important to stress that the Mediterranean diet also allows the intake of fat-soluble antioxidants, among which carotenoids, such as lutein, zeaxanthin, astaxanthin,  $\alpha$ -carotene,  $\beta$ -carotene, lycopene, etc., are certainly one of the most well represented groups in foods of vegetal origin [4]. Several studies indicated that these compounds, due to their capacity to scavenge reactive oxygen species

(ROS) [5], may exert beneficial effects in different pathological states characterized by increased ROS production, such as neurodegenerations [6], myocardial ischemia [7], diabetic retinopathy [8]. In the last decades, a renovated scientific interest has been developed not only towards fat-soluble antioxidants, but also towards fat-soluble vitamins. It was discovered that specific fat-soluble vitamins are either useful adjuvant in the pharmacological therapies of several pathologies [9, 10] or possibly implicated in some chronic neurodegenerations [11-13]. These facts strongly increased the importance in the clinic setting of a correct identification and quantification of fat-soluble vitamins and antioxidants, often present in very low concentrations in body fluids [14].

Several methods have been proposed for the separation of hydrophobic vitamins and antioxidants in different biological matrices, mainly based on reversed-phase HPLC and using various highly sensitive detectors to quantify their concentrations at the nanomolar levels (fluorometric, electrochemical, MS detectors) [15-19]. Due to the chemical nature (moderately to highly hydrophobic) of fat-soluble vitamins and antioxidants, and to the needs of their quantification in highly polar and complex biological samples (serum, plasma, seminal plasma), specific and efficient sample processing are required. To date, sample treatment for the extraction of fat-soluble vitamins are mainly based on complex multi-steps procedures [20]. Besides using large amounts of highly toxic organic solvents (chloroform, n-hexane), these methods have the disadvantage to be time-consuming and expensive, rendering problematic their application for the routine preparation of high number of samples in one shot.

In this work, we describe a novel procedure for the efficient fat-soluble vitamin and antioxidant extraction from biological fluids based on a single-step, single-organic solvent treatment of samples, coupled to a simple and fast reversed phase HPLC method for the simultaneous separation of *all-trans*-retinoic acid, 25(OH)-D<sub>3</sub>, *all-trans*-retinol, astaxanthin, lutein, zeaxanthin, *trans*- $\beta$ -apo-8'-carotenal,  $\gamma$ -tocopherol,  $\beta$ -cryptoxanthin,  $\alpha$ -tocopherol, lycopene,  $\alpha$ -carotene,  $\beta$ -carotene, phylloquinone and coenzyme Q<sub>10</sub>. The method was validated in terms of sensitivity, reproducibility

and recovery and was successfully applied to measure the aforementioned compounds in two biological matrices, serum and seminal plasma, obtained from two groups of healthy donors.

## 2 MATERIALS AND METHODS

### 2.1 Chemicals

Standards of *all-trans*-retinoic acid, 25(OH)-D<sub>3</sub>, *all-trans*-retinol, astaxanthin, lutein, zeaxanthin, *trans*- $\beta$ -apo-8'-carotenal,  $\gamma$ -tocopherol,  $\beta$ -cryptoxanthin,  $\alpha$ -tocopherol, phylloquinone, lycopene,  $\alpha$ -carotene,  $\beta$ -carotene and coenzyme Q<sub>10</sub> were purchased from Sigma (Sigma-Aldrich, St. Louis, MO, USA) at the highest purity available. HPLC-grade solvents (methanol, acetonitrile, chloroform and ethanol) were supplied from Carlo Erba Reagenti (Milano, Italy). Ultrapure water was obtained from a Milli-Q system (Millipore, Millford, MA, USA).

### 2.2 Sampling of serum and seminal plasma

Peripheral venous blood samples were obtained from 50 healthy volunteers (33 males, 27 females;  $48.6 \pm 19.3$  years of age) recruited among the personnel of the Catholic University of Rome. Assumption of dietary supplements, rich in fat-soluble vitamins or antioxidants, was used as the only exclusion criterion. Whole blood was collected from the antecubital vein into a VACUETTE® polypropylene tube containing serum separator and clot activator (Greiner-Bio One GmbH, Kremsmunster, Austria) and immediately protected from light. After 30 minutes at room temperature in the dark, samples were centrifuged at  $1890 \times g$  for 10 min at 10 °C and 250  $\mu$ l of the separated sera were immediately withdrawn and processed for the extraction of fat-soluble vitamins and antioxidants.

Seminal plasma was prepared from ejaculates obtained from 20 fertile healthy volunteers ( $44.3 \pm 12.6$  years of age) recruited among the personnel of the University of Catania. The ascertained fertility (presence of offspring) was used as the only inclusion criterion. Within 60 minutes from ejaculation, the light-protected liquefied semen samples were centrifuged at  $1480 \times g$  for 10 min at

10 °C and the upper seminal plasma was immediately withdrawn and processed for fat-soluble vitamin and antioxidant extraction.

### ***2.3 Single-step extraction of fat-soluble vitamins and antioxidants***

An aliquot of each serum or seminal plasma sample (250 µl) was added to 500 µl of HPLC-grade CH<sub>3</sub>CN. After vigorous vortexing for 60 seconds, these mixtures were incubated at 37 °C for 1 hour in a water bath under agitation (to allow full extraction of lipid soluble compounds) and then centrifuged at 20,690 x g for 15 min at 4 °C to precipitate proteins. Clear supernatants were directly used for the reversed phase HPLC analysis of fat-soluble vitamins. All the aforementioned procedures were carried out by protecting samples from light, in order to avoid degradation of photo-sensitive molecules.

To assess recovery, aliquots of 5 randomly selected serum samples were either extracted with no addition or spiked with low (five samples) or high (five samples) concentrations of standard mixtures of fat-soluble vitamins and antioxidants, extracted and deproteinized according to what afore described and then analyzed by HPLC.

Additionally, in order to compare efficiency of this new sample processing with a previously well established method (using a multi-step, multi-organic solvent treatment of samples), five randomly selected serum samples were treated for fat-soluble vitamin and antioxidant extraction according to Granado-Lorencio (Granado-Lorencio 2006 [15]).

### ***2.4 Reversed phase HPLC assay of fat-soluble vitamins and antioxidants***

Stock solutions of *all-trans*-retinoic acid, *all-trans*-retinol, 25(OH)-D<sub>3</sub>, α-tocopherol, γ-tocopherol and phylloquinone were prepared in ethanol to get a 1 mM final concentration. Lutein, zeaxanthin, astaxanthin, β-cryptoxanthin, α-carotene, β-carotene, *trans*-β-apo-8'-carotenal, lycopene and coenzyme Q<sub>10</sub> were dissolved as 1 mM stock solutions in chloroform. All standards were carefully protected from light.



Stock solutions were stored at  $-80^{\circ}\text{C}$  for up to one month with a loss  $\leq 1\%$ . Proper standard mixtures with known concentrations used to validate the HPLC method (linearity, sensitivity, reproducibility) were prepared daily by diluting stock solutions with HPLC-grade  $\text{CH}_3\text{CN}$ . Before HPLC analysis, mixtures were vortexed for 60 seconds and then incubated for 1 hour at  $37^{\circ}\text{C}$  centigrade, using the same conditions adopted for biological fluid extraction.

### ***2.5 HPLC apparatus and chromatographic conditions for fat-soluble vitamin and antioxidant separation***

The HPLC apparatus consisted of a Spectra System P4000 pump system (Thermo Fisher Scientific, Rodano, Milan, Italy) equipped with a highly-sensitive, 5 cm light path flow cell, UV6000LP diode array detector (Thermo Fisher Scientific, Rodano, Milan, Italy), setup for acquisition between 200 and 550 nm wavelengths. Data acquisition and analysis were performed using the ChromQuest software package provided by the HPLC manufacturer. Chromatographic separation was performed using a 100 x 4.6 mm, 5  $\mu\text{m}$  particle size Hypersil Gold RP C18 column provided with its own guard column (Thermo Fisher Scientific, Milan, Italy). A gradient from solvent A (70% methanol 30%  $\text{H}_2\text{O}$ ) to solvent B (100% acetonitrile) was formed as follows: 0.5 min at 100% A; 8 min at up to 100% B (hold for additional 27 minutes). A flow rate of 1.0 mL/min and a column temperature of  $37^{\circ}\text{C}$  were maintained constant throughout the analysis, while a flow rate of 2 ml/min was used during column washing with 100% B (5 minutes) and the re-equilibrating time with 100% A before the next run (5 minutes). A thermostated autosampler AS3000 (SpectraSystem, Thermo Fisher Scientific, Rodano, Milan, Italy) was connected to the chromatographic system and programmed to inject a volume of 200  $\mu\text{l}$  onto the column. Identification and quantification of fat-soluble vitamins and antioxidants in biological samples were obtained by comparing retention times and absorption spectra of different peaks to those of standard mixtures containing true compounds with known concentrations. The concentration of the different compounds in both serum and seminal plasma extracts were calculated from the standard run data at the following wavelengths: *all-trans*-retinoic

acid = 345 nm; 25(OH)-D<sub>3</sub> = 265 nm; *all-trans*-retinol = 327 nm; astaxanthin = 479 nm; lutein = 450 nm; zeaxanthin = 445 nm; *trans*- $\beta$ -apo-8'-carotenal = 458 nm;  $\gamma$ -tocopherol = 300 nm;  $\beta$ -cryptoxanthin = 480 nm;  $\alpha$ -tocopherol = 295 nm; phylloquinone = 245 nm; lycopene = 475 nm;  $\alpha$ -carotene = 477 nm;  $\beta$ -carotene = 481 nm; coenzyme Q<sub>10</sub> = 274 nm.

## 2.6 Statistics

Differences across groups were estimated by the two-way ANOVA for repeated measures. Fisher's protected least square was used as the post hoc test. Only two-tailed p-values of less than 0.05 were considered statistically significant.

## 3 RESULTS

### 3.1 HPLC separation of standard mixtures of fat-soluble vitamins

In Figure 1 a representative chromatogram of a standard mixture containing 0.1-5  $\mu$ M of *all-trans*-retinoic acid, 25(OH)-D<sub>3</sub>, *all-trans*-retinol, astaxanthin, lutein, zeaxanthin, *trans*- $\beta$ -apo-8'-carotenal,  $\gamma$ -tocopherol,  $\beta$ -cryptoxanthin,  $\alpha$ -tocopherol, phylloquinone, lycopene,  $\alpha$ -carotene,  $\beta$ -carotene and coenzyme Q<sub>10</sub> is shown at three wavelengths (265 nm, panel A; 327 nm, panel B, 450 nm, panel C). The different  $\lambda_{\text{max}}$  of the absorption spectra of each substance are summarized in Table 1. By properly combining the differences in  $\lambda_{\text{max}}$  and retention times of the various compounds, it is possible to observe from Figure 1 that these HPLC conditions allow the separation and subsequent quantification of all the fat-soluble vitamins and antioxidants forming the standard mixture, even though the peaks of the isomeric carotenoids lutein and zeaxanthin (peaks 5 and 6 of panel C) were only partially resolved. It is worth underlining that standard mixtures underwent the same period of incubation at 37 °C occurring to selected biological fluids, prior to the HPLC analysis. Comparing areas of the peaks of chromatographic runs of mixtures analyzed with no incubation with those of runs of incubated mixtures, it was possible to establish that one hour at 37 °C did not cause any

appreciable degradation in any of the compounds of interest (values were comprised within the inter-assay coefficient of variation for peak areas reported in Table 3, data not shown).

The parameters to characterize the HPLC separation of these 15 fat-soluble vitamins and antioxidants, i.e. lower limit of detection (LLOD, evaluated with a signal to noise ratio  $> 3$ ), lower limit of quantification (LLOQ, evaluated with a signal to noise ratio  $> 10$ ) and linearity are summarized in Table 2. Minimal and maximal LLOQ values were observed for zeaxanthin (0.3 nM), and  $\alpha$ -tocopherol (30 nM), respectively. Linearity was observed for all of the compounds in the concentration range between LLOQ and 3000 x LLOQ (five mixtures with the intermediate concentrations of 10, 20, 50, 500 and 1500 x LLOQ for each analyte were assayed), with high values of the respective correlation coefficients.

Reproducibility of the analysis, determined by calculating intra-assay (five consecutive chromatographic runs of the same mixture) and inter-assay (five chromatographic runs of five mixtures, freshly prepared on five consecutive days) coefficients of variations (CV) of peak areas and retention times, are reported in Table 3. Low values for the CVs (both intra- and inter-assay) of peak areas and retention times were recorded for these chromatographic conditions. It is worth recalling that to perform these tests all standard mixtures underwent the same extraction process used for biological samples, prior to the HPLC analysis.

### ***3.2 Evaluation of the extraction procedure of biological fluids for the HPLC analysis of fat-soluble vitamins and antioxidants***

The suitability of the new extraction procedure of fat-soluble vitamins and antioxidants from biological fluids, as well as that of the HPLC conditions applied for their determination, was preliminarily studied by spiking five randomly selected serum samples with a mixture containing either low (10 x LLOQ) or high vitamin and antioxidant concentrations (500 x LLOQ). The non-enriched and enriched serum samples were then processed and assayed as previously described. Recovery, reported in Tables 4a and 4b, demonstrates a high efficiency of the new single-step

method to extract fat-soluble vitamins and antioxidants, when these compounds are added to biological fluids at either low or high concentrations. The minimal recovery of 80% was recorded when zeaxanthin was added at low concentration (Table 4a). It is worth underlining that since  $\beta$ -tocopherol could not be separated from  $\gamma$ -tocopherol using these chromatographic conditions, and since this position isomer is present in cereals [21] and might therefore be detectable in biological samples, hereinafter these isomers are cumulatively indicated in sample extracts as  $\beta$ + $\gamma$ -tocopherols. To further corroborate the suitability of the present biological fluid processing finalized to measure fat-soluble vitamins and antioxidants by HPLC with UV-VIS-DAD detector, the same five serum samples were also processed using one of the conventional multi-steps method reported in literature [15]. Results summarized in Table 5, strongly demonstrate that a single-step extraction of serum with one organic solvent only (acetonitrile for 60 min at 37 °C) has the same efficiency than that of a multi-steps, multi-organic solvents, time consuming extraction process [15, 14]. The statistical comparison of the results, performed according to the analysis of variance for repeated measures, showed no differences in the concentrations of the various compounds determined in the same serum samples extracted with the two methods.

### ***3.3 Fat-soluble vitamins and antioxidants in serum and seminal plasma***

After having demonstrated that the single-step procedure efficiently extracts lipid soluble vitamins and antioxidants from serum and that the reversed phase HPLC conditions we set up had the correct characteristics of sensitivity, reproducibility and linearity, we extracted and analyzed by HPLC serum samples of 50 healthy controls to determine their circulating concentrations in lipid soluble vitamins and antioxidants.

A representative chromatogram of the separation of fat-soluble vitamins and antioxidants in serum of a control healthy subject and in seminal plasma of a fertile donor, subsequent to the single-step  $\text{CH}_3\text{CN}$  deproteinization/extraction, are shown, respectively, in Figures 2 and 3 at three different wavelengths (265 nm, panel A; 327 nm, panel B; 450 nm, panel C). At 265 nm wavelength,

25(OH)-D<sub>3</sub> is clearly identified in serum but not in seminal plasma. Either  $\beta+\gamma$ - and  $\alpha$ -tocopherol are visible at this wavelength in both extracts, as well as coenzyme Q<sub>10</sub>, the peak of which is shown at 274 nm wavelength in the inserts of panel A of Figures 2 and 3. A very small peak of phylloquinone (lower than its LLOQ) was also found in serum extract (not shown). When observing the same chromatograms at 327 nm wavelength, a clearly visible peak of vitamin A (peak number 3) and a small peak of *all-trans*-retinoic acid (peak number 1) are identifiable in both fluids. Observing these two chromatographic runs at a wavelength typical of carotenoids (450 nm), it is possible to note the incomplete separation of lutein and zeaxanthin, and that astaxanthin, lutein, zeaxanthin, *trans*- $\beta$ -apo-8'-carotenal,  $\beta$ -cryptoxanthin, lycopene,  $\alpha$ -carotene,  $\beta$ -carotene are all easily measurable in both serum and seminal plasma extracts.

The mean serum concentrations of the 15 compounds of interest measured in this group of healthy donors are summarized in Table 6. It is possible to note that we could not determine phylloquinone (vitamin K<sub>1</sub>) concentration. As also evident from the chromatogram shown in Figure 2 (Panel A), the values of the peak area of phylloquinone in all serum samples assayed were below its LLOQ (6 nM), confirming that the physiological values of this substance in serum (0.5-1 nM) [22, 23] are below the sensitivity of our highly sensitive diode array detector. That is, the single-step process effectively extracts even phylloquinone (see Tables 4a and 4b reporting the efficiency of the extraction in serum samples spiked with standard mixture containing measurable levels of phylloquinone) but a more sensitive detector (mass spectrometer) for its quantification in serum is required.

The other 14 fat-soluble compounds were always measurable in all serum samples and showed a moderate range of variability. The lowest and highest concentrations of lipid soluble vitamins and antioxidants in serum were found, respectively, for astaxanthin and zeaxanthin ( $0.004 \pm 0.002$   $\mu\text{mol/l}$  serum for both compounds) and  $\alpha$ -tocopherol ( $28.51 \pm 7.08$   $\mu\text{mol/l}$  serum).

The applicability of the present single-step method was also tested to extract fat-soluble vitamins and antioxidants from seminal plasma obtained from 20 healthy, fertile control donors and

subsequently analyzing these extracts using the above described HPLC conditions. Data reported in Table 7 indicate that, under the present HPLC conditions, seminal plasma does not contain measurable levels of 25(OH)-D<sub>3</sub>, astaxanthin and phylloquinone. Quantifiable concentrations of the remaining 12 compounds (*all-trans*-retinoic acid, *all-trans*-retinol, lutein, zeaxanthin, *trans*-β-apo-8'-carotenal, β+γ-tocopherol, β-cryptoxanthin, α-tocopherol, lycopene, α-carotene, β-carotene and coenzyme Q<sub>10</sub>) were found in all seminal plasma extracts, indicating that the present extraction method and HPLC separation can successfully be applied to monitor the status of lipid soluble antioxidants in human ejaculate. As it occurs in serum, α-tocopherol is the lipid soluble compound with the maximal concentration in seminal plasma ( $3.06 \pm 0.85 \mu\text{mol/l}$ ).

## 4 DISCUSSION

Results reported in the present study allowed to validate a single-step, efficient, simple, rapid and low cost method to extract fat-soluble vitamins and antioxidants from biological fluids which, coupled to a proper reversed phase HPLC analysis, is suitable to determine the concentrations of *all-trans*-retinoic acid, 25(OH)-D<sub>3</sub>, *all-trans*-retinol, astaxanthin, lutein, zeaxanthin, *trans*- $\beta$ -apo-8'-carotenal,  $\beta$ + $\gamma$ -tocopherol,  $\beta$ -cryptoxanthin,  $\alpha$ -tocopherol, phylloquinone, lycopene,  $\alpha$ -carotene,  $\beta$ -carotene and coenzyme Q<sub>10</sub> in human serum and seminal plasma. Values of these fat-soluble compounds we found in serum of control healthy subjects (Table 6) are in accordance with previous studies in which different extraction and analytical methods were adopted [14, 19, 24-28], thereby strongly demonstrating the suitability of both extraction method and HPLC conditions.

The classical multi-steps, multi-solvents procedures to prepare biological fluids for subsequent analysis of fat-soluble vitamins and antioxidants [14, 15] are based on the initial deproteinization of samples (first step) with mildly polar organic solvents (methanol, acetonitrile, iso-propanol), followed by the extraction of the protein-free biological sample (second step) with a highly hydrophobic organic solvent (prevalently n-hexane). This highly hydrophobic phase, containing all fat-soluble compounds including vitamins and antioxidants, is recovered and evaporated to dryness (third step) and resuspended with a proper amount of mildly polar organic solvent (fourth step). This extract is lastly ready to be injected onto an HPLC column for the analysis of fat-soluble vitamins and antioxidants [14, 25]. Alternatively, solid-phase extraction (SPE) with proper cartridges is the most utilized procedure to process biological fluids for fat-soluble vitamin and antioxidant determination [29, 30]. However, even in this case, various steps (equilibration of the SPE cartridge and elution of samples) and solvents are necessary. It is therefore evident that any of these multi-steps, multi-solvents extraction procedures, as well as the SPE procedure, can hardly be applied to process simultaneously a large number of biological fluid samples. Hence, these processing protocols are relatively easy to be applied in research laboratories, but very difficult to

be applied in the routine of clinical-biochemistry laboratories, where hundreds of samples/day have to be treated.

All these limitations are avoided by applying the single-step, one organic solvent treatment of biological fluids to efficiently extract fat-soluble vitamins and antioxidants described in the present study. Differently from the traditional extraction methods, the use of which renders difficult the simultaneous preparation of several samples, the present single-step deproteinization/extraction protocol with acetonitrile only, can concomitantly be applied to tenths of samples with nearly the same working time. In our personal experience during this study, we received in one morning 34 serum samples from healthy controls: in about 2.5 hours, the whole process was completed for all 34 sera, and the first sample was placed in the autosampler and started to be analysed by HPLC to measure its fat-soluble vitamin and antioxidant levels.

Additional great advantages of this method are the minimal manipulation of samples, the reduction of costs-per-sample thanks to the use of moderately expensive solvent (acetonitrile), the elimination in the use of highly toxic organic solvents (n-hexane), the use of common laboratory equipments only (thermostated water bath, thermostated top bench centrifuge). It is worth recalling that the use of acetonitrile to extract serum samples for the analysis of fat-soluble vitamins has been proposed in the past [30]. According to that sample treatment, acetonitrile was supplemented with SDS, presumably to optimize extraction of lipid soluble vitamins potentially associated with albumin, and extracted samples in acetonitrile-SDS were passed through proper SPE cartridges before the HPLC assay [31]. Our results indicate that either SDS or SPE cartridges for sample cleaning are unnecessary, since a single-step extraction at 37 °C for one hour with acetonitrile only, efficiently extracts nearly the 100% of the serum fat-soluble vitamin and antioxidant content. This is clearly demonstrated by data summarized in Tables 4a and 4b reporting the percent recovery of serum samples spiked with low or high concentrations of standard mixtures of fat-soluble vitamins and antioxidants, and by data reported in Table 5 showing no statistical differences in the extraction efficiency of this single-step and the established multi-steps processing of samples. It is also worth



underlining that the present extraction method allows to obtain clean samples directly injectable onto the HPLC column. The present study has been realized using the same chromatographic column and guard column for about 400 runs with no significant change in its separation efficiency. Using this HPLC conditions, the separation of the compounds of interest is obtained in about 30 minutes, with a run-to-run time interval of about 45 minutes, i.e. a maximum of 32 samples/day can be assayed. As demonstrated by data summarized in Tables 2, 3, 4a and 4b this HPLC method is characterized by high sensitivity, reproducibility and linearity in a wide range of concentrations. Referring to sensitivity, the use of a highly sensitive diode array detector (equipped with a 5 cm light path flow cell) consent to detect and to quantify the various compounds at the nanomolar level. This characteristic contributes to decrease the analytical costs since a moderately expensive HPLC apparatus is required, particularly when compared to HPLC methods based on the use of MS detectors [16, 19, 20, 29]. It is important to recall that, notwithstanding the high sensitivity of the method, we could not quantify phylloquinone in serum since the area of the peak in all samples was always below its LLOQ, that is, MS detectors are needed for the quantification of phylloquinone in serum [22, 23, 32, 33].

Since a renovated interest for the determination of fat-soluble vitamins and antioxidants has been developed in recent years, due to their possible involvement in different pathological states [9, 34-36], we verified the suitability of the single-step extraction method, coupled with the presently characterized HPLC separation, to measure these fat-soluble compounds not only in serum samples of controls but also in seminal plasma of healthy fertile subjects. To our knowledge, this is the first time that a comprehensive profile of fat-soluble vitamins and antioxidants has been measured in extracts of seminal plasma. A number of previous studies correlated various plasma antioxidants with infertility [37, 38]); very few measured antioxidants directly in seminal plasma [27, 39]; none measured the pattern we assayed in our seminal plasma samples [40]. Our results indicate that fat-soluble vitamins and antioxidants are present in relatively low amounts in ejaculates of healthy fertile controls (Table 7), also indicating that 25(OH)-D<sub>3</sub>, astaxanthin and phylloquinone are below

their respective LLOQ and LLOD (no peaks corresponding to these compounds could be observed in any seminal plasma extract). However, the mean cumulative concentration of carotenoids in our seminal plasma (lutein + zeaxanthin + *trans*- $\beta$ -apo-8'-carotenal +  $\beta$ -cryptoxanthin + lycopene +  $\alpha$ -carotene +  $\beta$ -carotene) was  $0.180 \pm 0.014 \mu\text{mol/l}$ , representing about the 6% of the  $\alpha$ -tocopherol concentration, i.e. of the fat-soluble antioxidant found in the highest concentration in this biological fluid. That is, this group of lipid soluble compounds certainly exerts a significant protective action in human seminal plasma (under physiological conditions) towards biochemical damages induced by oxidative/nitrosative stress. Since most of these carotenoids are present only in certain types of food, it is unquestionably possible that both their concentrations and pattern may have significant differences in populations with different nutritional habits. However, having established values of normality in this group of fertile controls, it will now be possible to verify whether a decrease in the concentration of one or more of these lipid soluble vitamins and antioxidants occurs in pathological seminal plasma, i.e. in seminal plasma from infertile subjects.

#### **4.1 Conclusions**

In conclusion, the single-step, single-organic solvent extraction of fat-soluble vitamins and antioxidants from biological fluids reported in the present study is a valid innovation over the existing methods [14, 15], particularly in terms of time-for-sample processing and costs of reagents. Coupled with the HPLC conditions purposely set up, which do not however allow separation of  $\beta$ - and  $\gamma$ -tocopherol and do not permit a complete separation of the lutein and zeaxanthin peaks, this procedure is suitable to extract and measure fat-soluble vitamins and antioxidants in a high number of human biological fluids (serum, plasma, seminal plasma, seminal fluid, urine, cerebrospinal fluid, saliva, microdialysis, synovial fluid). Although further confirmations are certainly needed, this method might be applied also in the routine of clinical-biochemistry laboratories, where a potentially high number of samples/day has to be processed and assayed. Therefore, the extraction and HPLC methods described in the present study might successfully be applied in clinical and

nutritional studies, when supplementation of vitamins A, D, E, and/or carotenoids and/or antioxidant (coenzyme Q<sub>10</sub>) is a powerful strategy for several pathologies. Monitoring of the aforementioned compounds in selected biological fluids is obviously necessary [9, 10, 14, 36, 41] and require a suitable, low-cost, time saving method to process large number of samples such as the new single-step, single-solvent procedure described in the present study.

## **ACKNOWLEDGEMENTS**

We would like to thank Mr. Salvatore Meo for his technical assistance in preparing the artworks. This work has been supported by research funds of the Catholic University of Rome and of the University of Catania.

**REFERENCES**

- [1] J. Von Lintig, Provitamin A metabolism and functions in mammalian biology, *Am J Clin Nutr.* 96 (2012) 1234S-1244S.
- [2] J.A. Olson, Provitamin A function of carotenoids: the conversion of beta-carotene into vitamin A, *J. Nutr.* 119 (1989) 105-108.
- [3] M. Gerber, Biofactors in the Mediterranean diet, *Clin. Chem. Lab. Med.* 41 (2003) 999-1004.
- [4] D. Znidarčič, D. Ban, H. Sircelj, Carotenoid and chlorophyll composition of commonly consumed leafy vegetables in Mediterranean countries, *Food Chem.* 129 (2011) 1164-1168.
- [5] A. Mortensen, L.H. Skibsted, T.G. Truscott, The interaction of dietary carotenoids with radical species, *Arch. Biochem. Biophys.* 385 (2001) 13–19.
- [6] R. Sandhir, A. Mehrotra, S.S. Kamboj, Lycopene prevents 3-nitropropionic acid-induced mitochondrial oxidative stress and dysfunctions in nervous system, *Neurochem. Int.* 57 (2010) 579-587.
- [7] C. Tong, C. Peng, L. Wang, L. Zhang, X. Yang, P. Xu, J. Li, T. Delplancke, H. Zhang, H. Qi, Intravenous Administration of Lycopene, a Tomato Extract, Protects against Myocardial Ischemia-Reperfusion Injury, *Nutrients.* 8 (2016) 138 doi: 10.3390/nu8030138.

- [8] K. Neelam, C.J. Goenadi, K. Lun, C.C. Yip, K.G. Au Eong, Putative protective role of lutein and zeaxanthin in diabetic retinopathy, *Br. J. Ophthalmol.* (2017) doi: 10.1136/bjophthalmol-2016-309814.
- [9] Z. Asemi, F. Raygan, F. Bahmani, Z. Rezavandi, H.R. Talari, M. Rafiee, S. Poladchang, M. Darooghegi Mofrad, S. Taheri, A.A. Mohammadi, A. Esmailzadeh, The effects of vitamin D, K and calcium co-supplementation on carotid intima-media thickness and metabolic status in overweight type 2 diabetic patients with CHD, *Br. J. Nutr.* 116 (2016) 286-293.
- [10] R. Zakariaeeabkoo, Are vitamins A and D important in the development of food allergy and hoe are they best measured?, *Clin. Biochem.* 47 (2014) 804-811.
- [11] C.K. Sen, C. Rink, S. Khanna, Palm oil-derived natural vitamin E alpha-tocotrienol in brain health and disease, *J. Am. Coll. Nutr.* 29 (3 Suppl) (2010) 314S-323S.
- [12] G. Raszewski, R. Chwedorowicz, A. Chwedorowicz, K. Gustaw Rothenberg. Homocysteine, antioxidant vitamins and lipids as biomarkers of neurodegeneration in Alzheimer's disease versus non-Alzheimer's dementia, *Ann. Agric. Environ. Med.* 23 (2016) 193-196.
- [13] G.T. Vatassery, Vitamin E and other endogenous antioxidants in the central nervous system, *Geriatrics.* 53 Suppl 1 (1998) S25-S27.
- [14] J.M. Mata-Granados, M.D. Luque de Castro, J.M. Quesada Gomez, Inappropriate serum levels of retinol,  $\alpha$ -tocopherol, 25-hydroxyvitamin D<sub>3</sub> and 24, 25-dihydroxyvitamin D<sub>3</sub> levels in healthy Spanish adults: Simultaneous assessment by HPLC, *Clin. Biochem.* 41 (2008) 676-680.

- [15] F. Granado-Lorencio, B. Olmedilla-Alonso, C. Herrero-Barbudo, I. Blanco-Navarro, S. Blázquez-García, B. Pérez-Sacristán, Simultaneous determination of vitamins A, E and 25-OH-vitamin D: Application in clinical assessments, *Clin. Biochem.* 39 (2006) 180-182.
- [16] R. Andreoli, P. Manini, D. Poli, E. Bergamaschi, A. Mutti, W.M. Niessen, Development of a simplified method for the simultaneous determination of retinol, alpha-tocopherol, and beta-carotene in serum by liquid chromatography-tandem mass spectrometry with atmospheric pressure chemical ionization, *Anal. Bioanal. Chem.* 378 (2004) 987-994.
- [17] M.G. Ferruzzi, M.L. Nguyen, L.C. Sander, C.L. Rock, S.J. Schwartz, Analysis of lycopene geometrical isomers in biological microsamples by liquid chromatography with coulometric array detection, *J. Chromatogr. B Biomed. Sci. Appl.* 760 (2001) 289-299.
- [18] W.A. MacCrehan, Schönberger E. Determination of retinol, alpha-tocopherol, and beta-carotene in serum by chromatography with absorbance and electrochemical detection, *Clin. Chem.* 33 (1987) 1585-1592.
- [19] S.L.M. Arnold, J.K. Amory, T.J. Walsh, N. Isoherranen, A sensitive and specific method for measurement of multiple retinoids in human serum with UHPLC-MS/MS, *J. Lipid Res.* 53 (2012) 587-598.
- [20] A.A. Albahrani, V. Rotarou, P.J. Roche, R.F. Greaves, A simultaneous quantitative method for vitamins A, D and E in human serum using liquid chromatography-tandem mass spectrometry, *J. Steroid Biochem. Mol. Biol.* 159 (2016) 41-53. doi: 10.1016/j.jsbmb.2016.02.019.

- [21] M.N. Irakli, V.F. Samanidou, I.N. Papadoyannis, Optimization and validation of the reversed-phase high-performance liquid chromatography with fluorescence detection method for the separation of tocopherol and tocotrienol isomers in cereals, employing a novel sorbent material, *J. Agric. Food Chem.* 60 (2012) 2076-2082.
- [22] R. Paroni, E.M. Faioni, C. Razzari, G. Fontana, M. Cattaneo, Determination of vitamin K<sub>1</sub> in plasma by solid phase extraction and HPLC with fluorescence detection, *J. Chromatogr. B.* 877 (2009) 351-354.
- [23] V. Ducros, M. Pollicand, F. Laporte, A. Favier, Quantitative determination of plasma vitamin K<sub>1</sub> by high-performance liquid chromatography coupled to isotope dilution tandem mass spectrometry, *Anal. Biochem.* 401 (2010) 7-14.
- [24] P. Borel, M. Moussa, E. Reboul, B. Lyan, C. Defoort, S. Vincent-Baudry, M. Maillot, M. Gastaldi, M. Darmon, H. Portugal, R. Planells, D. Lairon, Human plasma levels of vitamin E and carotenoids are associated with genetic polymorphisms in genes involved in lipid metabolism, *J. Nutr.* 137 (2007) 2653-2659.
- [25] J. Karppi, T. Nurmi, B. Olmedilla-Alonso, F. Granada-Lorencio, K. Nyssönen, Simultaneous measurement of retinol,  $\alpha$ -tocopherol and six carotenoids in human plasma by using an isocratic reversed-phase HPLC method, *J. Chromatogr. B.* 867 (2008) 226-232.
- [26] C.K. Schmidt, A. Brouwer, H. Nau, Chromatographic analysis of endogenous retinoids in tissues and serum, *Anal. Biochem.* 315 (2003) 36-48.

- [27] A.E. Omu, T. Fatinikun, N. Mannazhath, S. Abraham, Significance of simultaneous determination of serum and seminal plasma alpha-tocopherol and retinol in infertile men by high-performance liquid chromatography, *Andrologia*. 31 (1999) 347-54.
- [28] B.L. Lee, C.N. Ong, Comprehensive high-performance liquid chromatographic method for the measurements of lipophilic antioxidants in human plasma, *J. Chromatogr. A*. 1216 (2009) 3131-3137.
- [29] F. Priego Capote, J.R. Jiménez, J.M. Granados, M.D. de Castro, Identification and determination of fat-soluble vitamins and metabolites in human serum by liquid chromatography/triple quadrupole mass spectrometry with multiple reaction monitoring, *Rapid Commun, Mass Spectrom.* 21 (2007) 1745-1754.
- [30] P.F. Chatzimichalakis, V.F. Samanidou, I.N. Papadoyannis, Development of a validated liquid chromatography method for the simultaneous determination of eight fat-soluble vitamins in biological fluids after solid-phase extraction, *J. Chromatogr. B. Analyt. Technol. Biomed. Life Sci.* 805 (2004) 289-296.
- [31] J.M. Mata-Granados, J.M. Quesada Gómez, M.D. Luque de Castro, Fully automatic method for the determination of fat soluble vitamins and vitamin D metabolites in serum, *Clin. Chim. Acta* 403 (2009) 126-130.
- [32] Q. Song, A. Wen, L. Ding, L. Dai, L. Yang, X. Qi, HPLC-APCI-MS for the determination of vitamin K<sub>1</sub> in human plasma: Method and clinical application, *J. Chromatogr. B*. 875 (2008) 541-545.



- [33] A. Gentili, A. Cafolla, T. Gasperi, S. Bellante, F. Caretti, R. Curini, V. Pérez Fernández, Rapid, high performance method for the determination of vitamin K<sub>1</sub>, menaquinone-4 and vitamin K<sub>1</sub> 2,3-epoxide in human serum and plasma using liquid chromatography-hybrid quadrupole linear ion trap mass spectrometry, *J. Chromatogr. A.* 1338 (2014) 102-110.
- [34] S.J. Hodges, M.J. Pilkington, M.J. Shearer, L. Bitensky, J. Chayen, Age-related changes in the circulating levels of congeners of vitamin K<sub>2</sub>, menaquinone-7 and menaquinone-8, *Clin. Sci. (Lond).* 78 (1990) 63-66.
- [35] A.G. Scrimgeour, M.L. Condlin, Nutritional treatment for traumatic brain injury, *J. Neurotrauma.* 31 (2014) 989-999.
- [36] C.R. Polachini, R.M. Spanevello, D. Zanini, J. Baldissarelli, L.B. Pereira, M.R. Schetinger, I.B. da Cruz, C.E. Assmann, M.D. Bagatini, V.M. Morsch, Evaluation of Delta-Aminolevulinic Dehydratase Activity, Oxidative Stress Biomarkers, and Vitamin D Levels in Patients with Multiple Sclerosis, *Neurotox. Res.* 29 (2016) 230-242.
- [37] L. Keskes-Ammar, N. Feki-Chakroun, T. Rebai, Z. Sahnoun, H. Ghazzi, S. Hammami, K. Zghal, H. Fki, J. Damak, A. Bahloul, Sperm oxidative stress and the effect of an oral vitamin E and selenium supplement on semen quality in infertile men. *Arch. Androl.* 49 (2003) 83-94.
- [38] A.O. Hammoud, A.W. Meikle, C.M. Peterson, J. Stanford, M. Gibson, D.T. Carrell, Association of 25-hydroxy-vitamin D levels with semen and hormonal parameters. *Asian J. Androl.* 14 (2012) 855-859.


[39] R. Kand'ár, P. Drábková, R. Hampl, The determination of ascorbic acid and uric acid in human seminal plasma using an HPLC with UV detection, *J. Chromatogr. B Analyt. Technol. Biomed. Life Sci.* 879 (2011) 2834-2839.

[40] S.I. Benedetti, M.C. Tagliamonte, S. Catalani, M. Primiterra, F. Canestrari, S. De Stefani, S. Palini, C. Bulletti, Differences in blood and semen oxidative status in fertile and infertile men, and their relationship with sperm quality, *Reprod. Biomed. Online.* 25 (2012) 300-306. doi: 10.1016/j.rbmo.2012.05.011.


[41] B. Walther, J.P. Karl, S.L. Booth, P. Boyaval, Menaquinones, bacteria, and the food supply: the relevance of dairy and fermented food products to vitamin K requirements, *Adv.. Nutr.* 4 (2013) 463-473.

## LEGEND TO FIGURES


**Figure 1.** Representative chromatogram of a standard mixture containing 0.1-5  $\mu\text{M}$  of fat-soluble vitamins and antioxidants, shown at three different wavelengths. 1 = *all-trans*-retinoic acid; 2 = 25(OH)-D<sub>3</sub>; 3 = *all-trans*-retinol; 4 = astaxanthin; 5 = lutein; 6 = zeaxanthin; 7 = *trans*- $\beta$ -apo-8'-carotenal; 8 =  $\gamma$ -tocopherol; 9 =  $\beta$ -cryptoxanthin; 10 =  $\alpha$ -tocopherol; 11 = phylloquinone; 12 = lycopene; 13 =  $\alpha$ -carotene; 14 =  $\beta$ -carotene; 15 = coenzyme Q<sub>10</sub>. Panel A = 265 nm; panel B = 327 nm; panel C = 450 nm. Chromatographic conditions are fully described under Materials and Methods.



**Figure 2.** Representative chromatogram of the separation of fat-soluble vitamins and antioxidants in serum of a control healthy subject, extracted according to the new single-step extraction, is illustrated at three different wavelengths (265 nm, panel A; 327 nm, panel B; 450 nm, panel C). Peak numbers correspond to the different compounds as indicated in the legend to Figure 1, with the exception of peak 8 that in biological fluid extracts is =  $\beta$ + $\gamma$ -tocopherol. In the insert of panel A, the chromatogram was magnified and replotted at 274 nm wavelength (the maximum of coenzyme Q<sub>10</sub> absorption) to allow a better identification of the peak referring to coenzyme Q<sub>10</sub>. In the insert of panel C, the chromatogram was magnified to better appreciate that the incomplete separation of lutein and zeaxanthin did not prevent correct calculation of both compounds. Sample processing and chromatographic conditions are fully described under Materials and Methods.



**Figure 3.** Representative chromatogram of the separation of fat-soluble vitamins and antioxidants in seminal plasma of a control fertile subject, extracted according to the new single-step extraction, is illustrated at three different wavelengths (265 nm, panel A; 327 nm, panel B; 450 nm, panel C). Peak numbers correspond to the different compounds as indicated in the legend to Figure 1, with the exception of peak 8 that in biological fluid extracts is =  $\beta$ + $\gamma$ -tocopherol. In the insert of panel A, the chromatogram was magnified and replotted at 274 nm wavelength to allow a better identification of the peak referring to coenzyme Q<sub>10</sub>. Sample processing and chromatographic conditions are fully described under Materials and Methods.



**Table 1.** Maxima of absorption recorded between 200 and 500 nm wavelengths for 15 fat soluble vitamins and antioxidants.

	<b>Wavelengths of maxima of adsorption (nm)</b>
<i>all trans</i> -retinoic acid	345
25-hydroxycholecalciferol	265
<i>all trans</i> -retinol	327
Astaxanthin	301 - 479
Lutein	270 - 423 - 450 - 475
Zeaxanthin	335 - 445
<i>trans</i> - $\beta$ -apo-8'-carotenal	218 - 270 - 458
$\gamma$ -tocopherol	216 - 300
$\beta$ -cryptoxanthin	275 - 456 - 480
$\alpha$ -tocopherol	295
Phylloquinone	245 - 266 - 330
Lycopene	296 - 448 - 475 - 503
$\alpha$ -carotene	265 - 450 - 477
$\beta$ -carotene	454 - 481
Coenzyme Q <sub>10</sub>	274

Absorption spectra were acquired at the maximal heights of the chromatographic peaks of each compound, after baseline subtraction.

**Table 2.** Lower limit of detection, lower limit of quantification and linearity of the reversed phase HPLC method for the detection of fat soluble vitamins and antioxidants in biological fluids.

	<sup>a</sup> LLOD (nM)	<sup>b</sup> LLOQ (nM)	3000 x LLOQ ( $\mu$ M)	<sup>c</sup> Correlation coefficients of linearity straight lines
<i>all-trans-retinoic acid</i>	1.2	2	6	0.996
<b>25-hydroxycholecalciferol</b>	2	3	9	0.993
<i>all-trans-retinol</i>	2	3	9	0.999
<b>Astaxanthin</b>	0.5	0.8	2.4	0.999
<b>Lutein</b>	0.5	0.8	2.4	0.999
<b>Zeaxanthin</b>	0.2	0.3	0.9	0.999
<i>trans-<math>\beta</math>-apo-8'-carotenal</i>	2	3	9	0.996
$\gamma$ -tocopherol	10	16	48	0.998
$\beta$ -cryptoxanthin	0.4	0.6	1.8	0.999
$\alpha$ -tocopherol	20	30	90	0.998
<b>Phylloquinone</b>	1	1.6	4.8	0.997
<b>Lycopene</b>	0.4	0.6	1.8	0.999
$\alpha$ -carotene	2	3	9	0.999
$\beta$ -carotene	0.3	0.5	1.5	0.999
<b>Coenzyme Q<sub>10</sub></b>	2	3	9	0.999

<sup>a</sup>LLOD = Lower limit of detection, evaluated with a signal to noise ratio > 3.

<sup>b</sup>LOQ = limit of quantification, evaluated with a signal to noise ratio > 10.

<sup>c</sup>Linearity was determined by assaying standard mixtures with the following concentrations:

LLOQ, 10 x LLOQ, 20 x LLOQ, 50 x LLOQ, 500 x LLOQ, 1500 x LLOQ and 3000 x LLOQ.

**Table 3.** Reproducibility of the HPLC method for the separation of fat soluble vitamins and antioxidants.

	<sup>a</sup> Intra-assay coefficient of variation of retention times	<sup>a</sup> Intra-assay coefficient of variation of peak areas	<sup>b</sup> Inter-assay coefficient of variation of retention times	<sup>b</sup> Inter-assay coefficient of variation of peak areas
<i>all-trans-retinoic acid</i>	0.12 ± 0.07	1.65 ± 0.12	0.52 ± 0.12	2.12 ± 0.50
<b>25-hydroxycholecalciferol</b>	0.18 ± 0.02	1.10 ± 0.10	0.24 ± 0.09	2.24 ± 0.41
<i>all-trans-retinol</i>	0.14 ± 0.02	0.67 ± 0.06	0.39 ± 0.08	1.23 ± 0.48
<b>Astaxanthin</b>	0.15 ± 0.03	1.01 ± 0.08	0.59 ± 0.07	1.99 ± 0.86
<b>Lutein</b>	0.17 ± 0.05	0.84 ± 0.09	0.22 ± 0.01	1.13 ± 0.21
<b>Zeaxanthin</b>	0.13 ± 0.02	0.99 ± 0.10	0.77 ± 0.08	1.83 ± 0.34
<i>trans-β-apo-8'-carotenal</i>	0.12 ± 0.01	0.47 ± 0.02	0.28 ± 0.02	1.67 ± 0.23
<b>γ-tocopherol</b>	0.16 ± 0.04	1.25 ± 0.11	0.51 ± 0.03	1.05 ± 0.17
<b>β-cryptoxanthin</b>	0.15 ± 0.02	0.88 ± 0.09	0.32 ± 0.09	1.34 ± 0.25
<b>α-tocopherol</b>	0.11 ± 0.03	0.56 ± 0.05	0.44 ± 0.07	1.11 ± 0.10
<b>Phylloquinone</b>	0.19 ± 0.03	1.40 ± 0.20	0.58 ± 0.08	2.01 ± 0.42
<b>Lycopene</b>	0.12 ± 0.03	0.53 ± 0.04	0.34 ± 0.02	1.09 ± 0.21
<b>Alfa-carotene</b>	0.13 ± 0.01	0.87 ± 0.02	0.25 ± 0.01	1.03 ± 0.26
<b>Beta-carotene</b>	0.18 ± 0.02	0.77 ± 0.03	0.37 ± 0.04	1.02 ± 0.12
<b>Coenzyme Q<sub>10</sub></b>	0.21 ± 0.01	0.37 ± 0.02	0.55 ± 0.06	1.46 ± 0.39

<sup>a</sup>Each point is the mean ± SD of five consecutive chromatographic runs of the same standard mixture.

<sup>b</sup>Each point is the mean ± SD of five chromatographic runs of five different standard mixtures assayed in five consecutive days.



**Table 4a.** Recovery of the method for the detection of lipid soluble vitamins in biological fluids.

	Mean values in control serum samples ( $\mu\text{mol/l}$ serum)	Concentration added = 10 x LLOQ ( $\mu\text{mol/l}$ serum)	Expected mean values ( $\mu\text{mol/l}$ serum)	Mean measured values ( $\mu\text{mol/l}$ serum)	Mean recovery (%)	%RSD
<i>all-trans-retinoic acid</i>	0.005 $\pm$ 0.001	0.02	0.025	0.022 $\pm$ 0.003	88.0	13.6
<b>25-hydroxycholecalciferol</b>	0.067 $\pm$ 0.022	0.03	0.097	0.087 $\pm$ 0.007	89.6	8.0
<i>all-trans-retinol</i>	5.24 $\pm$ 1.09	0.03	5.27	4.99 $\pm$ 0.09	94.7	1.8
<b>Astaxanthin</b>	0.005 $\pm$ 0.001	0.008	0.013	0.014 $\pm$ 0.001	107.7	7.1
<b>Lutein</b>	0.670 $\pm$ 0.200	0.008	0.678	0.66 $\pm$ 0.01	97.3	1.5
<b>Zeaxanthin</b>	0.002 $\pm$ 0.001	0.003	0.005	0.003 $\pm$ 0.0004	80	13.3
<i>trans-<math>\beta</math>-apo-8'-carotenal</i>	0.114 $\pm$ 0.031	0.03	0.117	0.127 $\pm$ 0.011	108.5	8.7
<b><math>\beta</math>+<math>\gamma</math>-tocopherol</b>	1.24 $\pm$ 0.54	0.16	1.40	1.30 $\pm$ 0.08	92.8	6.2
<b><math>\beta</math>-cryptoxanthin</b>	0.006 $\pm$ 0.004	0.006	0.012	0.010 $\pm$ 0.001	83.3	10.0
<b><math>\alpha</math>-tocopherol</b>	22.44 $\pm$ 3.13	0.3	22.74	21.67 $\pm$ 0.51	95.3	2.3
<b>Phylloquinone</b>	N.D.	0.016	0.016	0.017 $\pm$ 0.001	106.3	5.9
<b>Lycopene</b>	0.178 $\pm$ 0.046	0.006	0.184	0.168 $\pm$ 0.014	91.3	8.3
<b><math>\alpha</math>-carotene</b>	0.011 $\pm$ 0.005	0.03	0.041	0.047 $\pm$ 0.006	114.6	12.8
<b><math>\beta</math>-carotene</b>	0.258 $\pm$ 0.043	0.005	0.263	0.249 $\pm$ 0.019	94.7	7.6
<b>Coenzyme Q<sub>10</sub></b>	0.055 $\pm$ 0.023	0.03	0.085	0.086 $\pm$ 0.004	101.2	4.7

Each value is the mean  $\pm$  SD of five serum samples. Five control serum samples were extracted and analysed with no addition. Aliquots of each serum were spiked, before extraction, with a standard mixture at low concentration (10 x LLOQ) of fat-soluble vitamins and antioxidants. N.D. = Not Detectable. Sample processing and chromatographic conditions are fully described in the Materials and Methods section.

**Table 4b.** Recovery of the method for the detection of lipid soluble vitamins in biological fluids.

	Mean values in control serum samples ( $\mu\text{mol/l}$ serum)	Concentrati on added = 500 x LLOQ ( $\mu\text{mol/l}$ serum)	Expected values ( $\mu\text{mol/l}$ serum)	Mean measured values ( $\mu\text{mol/l}$ serum)	Mean recovery (%)	%RSD
<i>all-trans-retinoic acid</i>	0.005 $\pm$ 0.001	1	1.005	1.022 $\pm$ 0.095	101.7	9.3
<b>25-hydroxycholecalciferol</b>	0.067 $\pm$ 0.022	1.5	1.567	1.499 $\pm$ 0.110	95.7	7.3
<i>all-trans-retinol</i>	5.24 $\pm$ 1.09	1.5	6.74	6.66 $\pm$ 0.54	98.9	8.1
<b>Astaxanthin</b>	0.005 $\pm$ 0.001	0.4	0.405	0.398 $\pm$ 0.041	98.3	10.3
<b>Lutein</b>	0.670 $\pm$ 0.200	0.4	1.07	1.06 $\pm$ 0.09	99.1	8.5
<b>Zeaxanthin</b>	0.002 $\pm$ 0.001	0.15	0.152	0.151 $\pm$ 0.018	99.3	11.9
<i>trans-<math>\beta</math>-apo-8'-carotenal</i>	0.114 $\pm$ 0.031	1.5	1.614	1.508 $\pm$ 0.14	93.4	9.3
<b><math>\beta</math>+<math>\gamma</math>-tocopherol</b>	1.24 $\pm$ 0.54	8	9.24	9.18 $\pm$ 0.15	99.4	1.6
<b><math>\beta</math>-cryptoxanthin</b>	0.006 $\pm$ 0.004	0.3	0.306	0.323 $\pm$ 0.018	105.6	5.6
<b><math>\alpha</math>-tocopherol</b>	22.44 $\pm$ 3.13	15	37.44	37.04 $\pm$ 2.42	98.9	6.5
<b>Phylloquinone</b>	N.D.	0.8	0.8	0.79 $\pm$ 0.07	98.8	8.9
<b>Lycopene</b>	0.178 $\pm$ 0.046	0.3	0.478	0.459 $\pm$ 0.32	96.0	7.0
<b><math>\alpha</math>-carotene</b>	0.011 $\pm$ 0.005	1.5	1.511	1.459 $\pm$ 0.088	96.6	6.0
<b><math>\beta</math>-carotene</b>	0.258 $\pm$ 0.043	0.25	0.508	0.518 $\pm$ 0.013	102.0	2.5
<b>Coenzyme Q<sub>10</sub></b>	0.055 $\pm$ 0.023	1.5	1.555	1.571 $\pm$ 0.048	101.2	3.1

Each value is the mean  $\pm$  SD of five serum samples. Five control serum samples were extracted and analysed with no addition. Aliquots of each serum were spiked, before extraction, with a standard mixture at high concentration (500 x LLOQ) of fat-soluble vitamins and antioxidants. N.D. = Not Detectable. Sample processing and chromatographic conditions are fully described in the Materials and Methods section.

**Table 5.** Comparison in the efficiency of the new single-step, single-organic solvent extraction of fat-soluble vitamins from control sera with that obtained using a well-established multi-step, multi-organic solvent protocol.

	<b>Lipid-soluble vitamins and antioxidants in serum samples after a new single-step extraction (<math>\mu\text{mol/l}</math> serum)</b>	<b>Lipid-soluble vitamins and antioxidants in serum samples after a conventional multi-step extraction (<math>\mu\text{mol/l}</math> serum)</b>
<i>all-trans</i> -retinoic acid	0.005 $\pm$ 0.001	0.006 $\pm$ 0.002
25-hydroxycholecalciferol	0.067 $\pm$ 0.022	0.049 $\pm$ 0.014
<i>all-trans</i> -retinol	5.24 $\pm$ 1.09	5.81 $\pm$ 1.34
Astaxanthin	0.005 $\pm$ 0.001	0.007 $\pm$ 0.002
Lutein	0.670 $\pm$ 0.200	0.598 $\pm$ 0.176
Zeaxanthin	0.002 $\pm$ 0.001	0.004 $\pm$ 0.002
<i>trans</i> - $\beta$ -apo-8'-carotenal	0.114 $\pm$ 0.031	0.142 $\pm$ 0.060
$\beta$ + $\gamma$ -tocopherol	1.24 $\pm$ 0.54	1.09 $\pm$ 0.48
$\beta$ -cryptoxanthin	0.006 $\pm$ 0.004	0.009 $\pm$ 0.004
$\alpha$ -tocopherol	22.44 $\pm$ 3.13	20.39 $\pm$ 5.01
Phylloquinone	N.D.	N.D.
Lycopene	0.178 $\pm$ 0.046	0.237 $\pm$ 0.079
$\alpha$ -carotene	0.011 $\pm$ 0.005	0.017 $\pm$ 0.006
$\beta$ -carotene	0.258 $\pm$ 0.043	0.236 $\pm$ 0.031
Coenzyme Q <sub>10</sub>	0.055 $\pm$ 0.023	0.058 $\pm$ 0.016

Each value is the mean  $\pm$  S.D. of five control serum samples processed using either the new protocol for single-step, single solvent extraction of fat-soluble vitamins and antioxidants from biological fluids, or according to a conventional multi-step, multi solvent method as described by (Granado-Lorencio, 2006). No statistical differences in any of the compounds assayed were observed between the two extraction methods. N.D. = Not Detectable. Extraction procedures and HPLC conditions are fully described under Materials and Methods.

**Table 6.** Circulating values of fat soluble vitamins and antioxidants found in serum of 50 healthy controls.

	<b>Concentration (<math>\mu\text{mol/l}</math> serum)</b>
<i>all-trans-retinoic acid</i>	$0.006 \pm 0.003$
<b>25-hydroxycholecalciferol</b>	$0.074 \pm 0.029$
<i>all-trans-retinol</i>	$5.69 \pm 1.89$
<b>Astaxanthin</b>	$0.004 \pm 0.002$
<b>Lutein</b>	$0.593 \pm 0.211$
<b>Zeaxanthin</b>	$0.004 \pm 0.002$
<i>trans-<math>\beta</math>-apo-8'-carotenal</i>	$0.152 \pm 0.054$
<b><math>\beta</math>+<math>\gamma</math>-tocopherol</b>	$1.68 \pm 0.76$
<b><math>\beta</math>-cryptoxanthin</b>	$0.008 \pm 0.003$
<b><math>\alpha</math>-tocopherol</b>	$28.51 \pm 7.08$
<b>Phylloquinone</b>	N.D.
<b>Lycopene</b>	$0.211 \pm 0.096$
<b><math>\alpha</math>-carotene</b>	$0.010 \pm 0.006$
<b><math>\beta</math>-carotene</b>	$0.274 \pm 0.099$
<b>Coenzyme Q<sub>10</sub></b>	$0.038 \pm 0.028$

Values are the mean  $\pm$  S.D. of 50 serum samples obtained from healthy controls and processed using the new protocol for single-step, single solvent extraction of fat-soluble vitamins and antioxidants. N.D. = Not Detectable. Extraction procedures and HPLC conditions are fully described under Materials and Methods.

**Table 7.** Concentrations of fat soluble vitamins and antioxidants in seminal plasma obtained from a group of 20 fertile controls and detected by HPLC.

	<b>Concentration (<math>\mu\text{mol/l}</math> seminal plasma)</b>
<i>all trans</i> retinoic acid	0.001 $\pm$ 0.001
25-hydroxycholecalciferol	N.D.
<i>all trans</i> retinol	0.068 $\pm$ 0.028
Astaxanthin	N.D.
Lutein	0.067 $\pm$ 0.032
Zeaxanthin	0.001 $\pm$ 0.001
<i>trans</i> - $\beta$ -apo-8'-carotenal	0.022 $\pm$ 0.005
$\beta$ + $\gamma$ -tocopherol	0.066 $\pm$ 0.024
$\beta$ -cryptoxanthin	0.003 $\pm$ 0.001
$\alpha$ -tocopherol	3.06 $\pm$ 0.85
Phylloquinone	N.D.
Lycopene	0.003 $\pm$ 0.001
$\alpha$ -carotene	0.004 $\pm$ 0.002
$\beta$ -carotene	0.008 $\pm$ 0.005
Coenzyme Q <sub>10</sub>	0.006 $\pm$ 0.003

Values are the mean  $\pm$  S.D. of 20 seminal plasma samples processed using the new protocol for single-step, single solvent extraction of fat-soluble vitamins and antioxidants. N.D. = Not Detectable. Extraction procedures and HPLC conditions are fully described under Materials and Methods.