

An air suspension to demonstrate the properties of torsion balances with fibers of zero length

Gettings, Conner; Speake, Clive

DOI:

[10.1063/1.5142373](https://doi.org/10.1063/1.5142373)

License:

Other (please specify with Rights Statement)

Document Version

Publisher's PDF, also known as Version of record

Citation for published version (Harvard):

Gettings, C & Speake, C 2020, 'An air suspension to demonstrate the properties of torsion balances with fibers of zero length', *Review of Scientific Instruments*, vol. 91, no. 2, 025108, pp. 025108-1 - 025108-7.
<https://doi.org/10.1063/1.5142373>

[Link to publication on Research at Birmingham portal](#)

Publisher Rights Statement:

This article may be downloaded for personal use only. Any other use requires prior permission of the author and AIP Publishing. This article appeared in *Rev. Sci. Instrum.* 91, 025108 (2020) and may be found at <https://aip.scitation.org/doi/10.1063/1.5142373>

General rights

Unless a licence is specified above, all rights (including copyright and moral rights) in this document are retained by the authors and/or the copyright holders. The express permission of the copyright holder must be obtained for any use of this material other than for purposes permitted by law.

- Users may freely distribute the URL that is used to identify this publication.
- Users may download and/or print one copy of the publication from the University of Birmingham research portal for the purpose of private study or non-commercial research.
- User may use extracts from the document in line with the concept of 'fair dealing' under the Copyright, Designs and Patents Act 1988 (?)
- Users may not further distribute the material nor use it for the purposes of commercial gain.

Where a licence is displayed above, please note the terms and conditions of the licence govern your use of this document.

When citing, please reference the published version.

Take down policy

While the University of Birmingham exercises care and attention in making items available there are rare occasions when an item has been uploaded in error or has been deemed to be commercially or otherwise sensitive.

If you believe that this is the case for this document, please contact UBIRA@lists.bham.ac.uk providing details and we will remove access to the work immediately and investigate.

An air suspension to demonstrate the properties of torsion balances with fibers of zero length

Cite as: Rev. Sci. Instrum. **91**, 025108 (2020); <https://doi.org/10.1063/1.5142373>

Submitted: 12 December 2019 . Accepted: 27 January 2020 . Published Online: 07 February 2020

C. Gettings , and C. C. Speake 



View Online



Export Citation



CrossMark

ARTICLES YOU MAY BE INTERESTED IN

[A simple vacuum suitcase for enabling plasma facing component characterization in fusion devices](#)

Review of Scientific Instruments **91**, 026104 (2020); <https://doi.org/10.1063/1.5119166>

[Research on fault diagnosis and state assessment of vacuum pump based on acoustic emission sensors](#)

Review of Scientific Instruments **91**, 025107 (2020); <https://doi.org/10.1063/1.5125639>

[Cryogenic frequency-domain electron spin resonance spectrometer based on coplanar waveguides and field modulation](#)

Review of Scientific Instruments **91**, 025106 (2020); <https://doi.org/10.1063/1.5141461>

Lock-in Amplifiers

Find out more today



 Zurich Instruments

An air suspension to demonstrate the properties of torsion balances with fibers of zero length

Cite as: Rev. Sci. Instrum. 91, 025108 (2020); doi: 10.1063/1.5142373

Submitted: 12 December 2019 • Accepted: 27 January 2020 •

Published Online: 7 February 2020



View Online



Export Citation



CrossMark

C. Gettings^{a)}  and C. C. Speake 

AFFILIATIONS

Institute for Gravitational Wave Astronomy, Astrophysics and Space Research Group, School of Physics and Astronomy, University of Birmingham, Edgbaston, Birmingham B15 2TT, United Kingdom

^{a)} Author to whom correspondence should be addressed: cgettings@star.sr.bham.ac.uk

ABSTRACT

We report on the design and characterization of an air-bearing suspension that has been constructed to highlight the properties of torsion balances with fibers of zero length. A float is levitated on this suspension, and its rotational and translational motion in the horizontal plane of the laboratory is controlled using magnetic actuators. We demonstrate the *in situ* electromagnetic tuning of the float's center-of-buoyancy to an accuracy of ± 0.3 mm, which was limited by the noise in the air bearing. The rotational stiffness of the float, which is approximately zero by design, was also measured. We compare the observed behavior of the float with the predictions of a detailed model of the statics of the float-actuator system. Finally, we briefly discuss the application of these ideas and results to the construction of sensitive devices for the measurement of weak forces with short ranges.

Published under license by AIP Publishing. <https://doi.org/10.1063/1.5142373>

I. INTRODUCTION

Torsion balances have a long history in experimental science, stretching back to the time of Cavendish.¹ In modern physics, torsion balances are used extensively in tests of the inverse square law of gravity,^{2,3} Casimir force measurements,⁴ and tests of the weak equivalence principle.⁵ The advantage of the torsion balance is that well-manufactured fibers have very low rotational stiffness giving a high sensitivity, and well-designed balances can be made such that, to some degree, tilt or horizontal acceleration due to seismic noise will not couple to rotational motion of the suspended mass (or bob).⁶ Despite these advantages, the traditional torsion balance design has some limitations, particularly when it is employed to detect forces within sub-millimeter ranges. Due to the vertical distance of the center-of-mass from the point of attachment, horizontal accelerations, due to micro-seismic motion, for example, can couple strongly to the simple pendulum mode. This makes control of the torsion bob difficult. At some level, there will always be some coupling of tilt to rotational motion.⁷ Issues also face low-frequency torsion pendulum experiments where ground vibration and other sources of Newtonian noise become increasingly problematic.^{8,9} Tilt-rotational mode coupling is also a concern for seismic inertial sensors.¹⁰ In addition to weak force measurements, the co-location of the center-of-mass

and center-of-buoyancy of a suspended mass is a crucial feature of horizontal accelerometers and tiltmeters.^{11,12} This is currently achieved only by the adjustment of small masses such as lockable screws.

The goal then is to create a device that shares the advantages of the torsion balance but is not limited by the drawbacks mentioned above, where necessary adjustments to the center-of-buoyancy can be achieved accurately and remotely irrespective of the device's environment. In Ref. 13, we showed how the stiffness of the actuators acting on a levitated object (referred to as a float) could be tuned *in situ* in such a way that the center-of-buoyancy of the levitation system could be altered to lie at the center-of-mass of the float and that the rotational stiffness could be tuned, ideally, to zero. This could all be achieved while simultaneously controlling the translational degrees of freedom. The center-of-mass is the point where inertial forces act, whereas the center-of-buoyancy is the location of the resultant of the forces that are applied to levitate the float and control its position. In the general case, the center-of-mass will not lie at the center-of-buoyancy due to manufacturing imperfections and so horizontal accelerations and tilts will couple to the rotational mode of the device. The classical torsion balance has an in-built low sensitivity to tilt and horizontal acceleration as the center-of-buoyancy of the torsion bob can lie to a good approximation on the rotational

axis. In Ref. 13, we presented some initial results of measurements of the tuning of the period of oscillation of a float suspended by perfect diamagnetism (superconductivity). In this paper, we focus on the demonstration of the precise tuning of the center-of-buoyancy of a float.

We have constructed an air suspension, referred to here as the air bearing,¹⁴ which levitates the float. The float is then controlled in the horizontal plane of the laboratory by magnetic actuators that consist of coil-magnet pairs. By changing the currents in the coils, we can tune the center-of-buoyancy of the float. We present measurements that support this concept, suggesting that torsion balances with fibers of zero length can indeed be tuned *in situ* to be rotationally decoupled from ground tilt and horizontal accelerations. The actuators were designed such that the rotational stiffness of the float was nominally zero. Without further tuning, the magnitude of the rotational stiffness was experimentally found to be lower than that of the torsion balance used in a recent determination of Newton's constant of gravitation.¹⁵

II. THEORY

In our previous publication, we derived expressions for the rotational stiffness and center-of-buoyancy shift for superconducting and electrostatic suspensions.¹³ Here, we give the corresponding expressions for a system that uses electromagnetic actuators.

Consider Fig. 1, where a magnet with dipole moment, m , lies in a magnetic field produced by a coil with current I . The magnetic field on the axis of a coil of negligible cross section can be described by the following¹⁶ equation:

$$B_{\zeta} = \frac{\mu_0 R^2 I}{2(R^2 + \zeta^2)^{\frac{3}{2}}}, \quad (1)$$

where μ_0 is the vacuum permeability constant, R is the radius of the coil, and ζ is the axial distance between the coil and magnet centers, as described by the coordinates in Fig. 1. We can integrate this expression over the dimensions of the real coils to find the field and its derivatives. If we assume that the magnetic field is uniform over dimensions of the magnet, we can write its potential energy in terms of its magnetic dipole moment,




FIG. 1. A schematic diagram of a magnet with a magnetic dipole moment, m , at an angle, $\theta + \psi$, to a magnetic field, B , produced by a coil with current I . The coil has a cross-sectional area of $7.2 \times 10^{-5} \text{ m}^2$.

$$U = -\vec{m} \cdot \vec{B} = -mB_{\zeta} \cos(\theta + \psi), \quad (2)$$

where θ is a fixed angle between the dipole moment and the ζ -axis and ψ is a small angle whose mean is zero, as indicated in Fig. 1. We ignore changes in the axial force due to small radial displacements and rotations of the magnet. For $\psi = 0$, the force on the magnetic moment is given by

$$F = -\frac{\partial U}{\partial \zeta} = m \frac{\partial B_{\zeta}}{\partial \zeta} \cos \theta. \quad (3)$$

Taking the magnetic field to be a maximum at the center of the coil, $\partial B_{\zeta} / \partial \zeta$ is negative, so for a magnetic moment that is aligned with the field, the force will be attractive and reach a maximum negative value at some axial distance. The stiffness in the ζ direction is given by

$$k_{\zeta\zeta} = -\frac{\partial F}{\partial \zeta} = -m \frac{\partial^2 B_{\zeta}}{\partial \zeta^2} \cos \theta. \quad (4)$$

In order for this stiffness to be positive, leading to a passively stable system, we need the product of the cosine term and $\partial^2 B_{\zeta} / \partial \zeta^2$ to be negative. At the peak force, $\partial^2 B_{\zeta} / \partial \zeta^2$ is zero, so, in principle, we can choose the sign of the linear stiffness by selecting the axial location of the magnet. If the magnet is closer to the coil than the location of the peak force (where $\partial^2 B_{\zeta} / \partial \zeta^2 < 0$) and $\theta = 0$, we can achieve a stable system. Equally we can achieve a stable system by selecting a position of the magnet that is further away from the coil than the peak force position and $\theta = \pi$. Now, consider the angular stiffness given as, again in the case where $\psi = 0$,

$$k_{\theta\theta} = \frac{\partial^2 U}{\partial \theta^2} = mB_{\zeta} \cos \theta. \quad (5)$$

Clearly, here the choice of θ will also determine the stability of the system. We will see below, where we consider the stiffness of the whole float given by the actuators that control the float, that it is advantageous to make the angular stiffness negative and we therefore select $\theta = \pi$. If we desire a system that is stable for linear motion, according to Eq. (4), we therefore need to position the magnets further from the coils than the position of maximum force. This, in turn, implies that the force between the magnet and coil is repulsive. We should note that any unstable system can be servo-controlled; however, in practice, servo control is more easily achieved with an intrinsically stable system.

Now, we consider our experimental setup with 8 such coil-magnet pairs arranged around the float, as shown in Fig. 2 (with a new global coordinate system). We can now define the potential energy of a single magnet/coil pair in terms of the coordinates and angles given in Fig. 2, say, for magnet 5 and coil e. We define the separation of the magnet from its opposing coil that is due to the rotation of the float, ψ , as $f(\psi)$. The separation due to its simple translation is defined as y , which is along the y -axis in Fig. 2. Figure 3 highlights the changing separation of this magnet and coil and the relevant geometric terms. We can therefore define

$$\zeta = f(\psi) + y \quad (6)$$

for this magnet/coil pair, and it can be easily shown from Fig. 3 that

$$f(\psi) = b - a \cos \psi + x_0 \sin \psi. \quad (7)$$

The potential energy of the magnet/coil pair can then be written as




FIG. 2. A plan view schematic drawing showing the float, coils 1–8, and magnets A–H. New global coordinates centered on the float are defined: the x and y axes in the horizontal plane and the z -axis normal to this plane, with rotation around these axes.

$$U = -mB(f(\psi) + y) \cos(\theta + \psi), \quad (8)$$

where we have dropped the subscript ζ on the magnetic field. This can be applied to any magnet/coil pair with a suitable change in coordinates. We have used a similar nomenclature as in Ref. 13 and, on comparing Eq. (6) with Eq. (16) for $g(\psi)$ in Ref. 13, we note that the magnetic/coil actuators are qualitatively different from the superconducting and electrostatic actuators previously discussed.




FIG. 3. A diagram highlighting the geometry of the changing separation between magnet 5 and coil e (from Fig. 2) as described in Eq. (7). As the float rotates around its geometric center from position 1 to position 2, the separation between the magnet and coil, $f(\psi)$, is a function of the amount of rotation, ψ . This is independent of any separation change due to simple translation, y , along the y -axis, as described in Eq. (6). The amount of rotation shown here is exaggerated for clarity, and as before, we assume that radial displacements along the x -axis are comparatively small and can be ignored. Here, we have used a similar nomenclature as in Ref. 13.

This is because the points of application of the forces to the float (the magnets) by the coils now move with the rotation of the float rather than being defined by the fixed location of the coils. It is convenient to express the magnetic field as a Taylor series around the equilibrium position where we define the equilibrium spacing between the magnet and coil as g_0 . The magnetic field can then be written as

$$B = B_0 + \left(\frac{dB}{d\zeta}\right)_{g_0} \cdot (f(\psi) + y - g_0) + \frac{1}{2} \left(\frac{d^2B}{d\zeta^2}\right)_{g_0} \cdot (f(\psi) + y - g_0)^2. \quad (9)$$

By substituting this into Eq. (8) and setting $y = 0$, we can find the contribution to the total rotational stiffness of the float from one (the i th) coil,

$$K_{\psi\psi}^i = \frac{d^2U}{d\psi^2} = m \left(\frac{d^2B}{d\zeta^2} x_0^2 + \frac{dB}{d\zeta} a - B_0 \right), \quad (10)$$

where we have assumed again that the magnet is anti-aligned with the coil field. The upper case notation for the stiffness constant refers to the complete float rather than an individual coil/magnet pair [as compared to Eqs. (4) and (5)]. We can see that in order to create a zero stiffness configuration, we require that the quantities in the bracket sum to zero. Noting that in our configuration, the second and third terms are both negative (see the discussion above regarding the signs of the derivatives of the magnetic field), in principle, this is possible. However, we did not pursue this in the work described here but chose a value of x_0 and the positions of the magnets relative to the coils in order to achieve a nominally zero stiffness. The total rotational stiffness is given as the sum of the terms in Eq. (10) from all the actuators, which approximately multiplies it by a factor of eight. In Sec. IV, we compare measurements of float's rotational stiffness with this prediction.

We define the nominal center-of-buoyancy (NCB) as the point in the horizontal plane where the moments of all the forces acting on the float are zero when the coils carry their nominal currents. If we consider the x direction, the position of the center-of-buoyancy with respect to the NCB of the float, x_{cb} , can be modified by changing the stiffnesses of some actuators relative to others. For example, using Eq. (12) from Ref. 13, we have

$$x_{cb} = \frac{K_{y\psi}}{K_{yy}}. \quad (11)$$

The term in the denominator is the sum of the linear stiffnesses that are given for each coil by Eq. (4). The term in the numerator is a cross term from the stiffness matrix describing the static behavior of the float and in the symmetrical case is zero. We can compute the individual contributions from each coil/magnet pair to this cross term with the help of Eqs. (7)–(9) to find

$$k_{y\psi} = x_0 m \frac{d^2B}{dy^2}. \quad (12)$$

The change in center-of-buoyancy in the x direction is then

$$x_{cb} = \frac{K_{y\psi}^a + K_{y\psi}^b + K_{y\psi}^e + K_{y\psi}^f}{K_{yy}^a + K_{yy}^b + K_{yy}^e + K_{yy}^f} \quad (13)$$

or

$$x_{cb} \approx x_0 \frac{(I^b + I^e - I^a - I^f)}{4I}, \quad (14)$$

where I is the average bias current in the coils labeled a , b , e , and f in Fig. 2 and clearly x_{cb} is proportional to the difference in the currents flowing in the respective coils. This implies the theoretical maximum accuracy to which the center-of-buoyancy can be tuned in the case presented here depends on the precision with which the actuators' strength can be changed, which, in turn, depends on the current noise of the coil drivers.

When the bearing is tilted by an angle, ξ , from the horizontal plane, the total torque acting on the float from the actuators can then be described by the following equation:

$$\Gamma = \kappa \frac{\Delta I}{4I} + Mg\xi \left(x_0 \frac{\Delta I}{4I} + x_{cm} \right), \quad (15)$$

where the first term in the brackets corresponds to x_{cb} as given in Eq. (14), M is the mass of the float, g is the acceleration due to gravity, and x_{cm} is the position of the float's center-of-mass with respect to the NCB. Here, κ represents the torque due to any asymmetry of the magnetic actuators and their positions around the float. In the ideal case, the process we employ for changing the bias currents to tune the center-of-buoyancy should not apply a torque, but in any real system, such a term does exist. Any instability in the currents applied to the float to achieve a center-of-buoyancy tuning will introduce noise into the actual measurement via the κ parameter, so clearly it is desirable to reduce its magnitude as much as possible. A similar expression to Eq. (15) describes the center of mass tuning in the y direction.

In order to measure the κ parameter and to check how the current tuning shifts the center-of-buoyancy, we need to eliminate x_{cm} from Eq. (15). We do this by adjusting the physical center-of-mass using balance weights, as described below, until it coincides with the NCB. Then, dividing by the current ratio, we find

$$\tau = \Gamma / \frac{\Delta I}{4I} = \kappa + Mg\xi x_0. \quad (16)$$

This equation states, if the torque on the float is measured over a range of tilt angles for a given set of bias currents in the coils, the torque due to the actuator asymmetry can be calculated at $\xi = 0$. Furthermore, it predicts a linear response of τ to the tilt angle. This gradient allows the change in center-of-buoyancy described in Eq. (14) to be experimentally determined. This comparison was verified by experiment.

III. EXPERIMENTAL SETUP

Our experimental setup consisted of 8 coil-magnet pairs, two on each side of the square shaped float as shown in Fig. 2. The float and bearing were made of aluminum alloy. The coils themselves were based on the OSEM coils that have been developed for LIGO,¹⁷ and each consisted of 500 turns of copper wire and had a mean radius of 18 mm. The magnets used were grade N38 neodymium iron boron magnets with a magnetic dipole moment of 0.775 N m/T and were cylindrical with a radius and depth of 5 mm. They were attached to the float using contact adhesive. The bearing's flat top surface consisted of 0.5 mm diameter holes in a 10 mm grid under the entire bottom surface of the float through which compressed air was pumped at a constant pressure to provide a lift force to the float.

Figure 2 shows a schematic drawing of the float, coils, and magnets, with a coordinate system centered on the float. A summary

TABLE I. A summary of the various dimensions of the float, bearing, and magnetic actuator setup. The dimension label corresponding to the equations in Sec. II is stated where applicable.

Parameter	Length (mm) \pm 0.5 mm
Coil-magnet axial distance (ζ)	10.0
Coil's mean radius (R)	18.0
Coil-magnet radial distance (r)	0.0
Coil's cross-sectional length	8.0
Coil's cross-sectional width	9.0
Actuator arm length (x_0)	42.0
Magnet radius and depth	5.0
Float side length	115.0
Float depth	10.0
Bearing tilt length	190.0
Photodiode-mirror distance	70.0

of the dimensions of all the relevant components is given in Table I, where the stated measurement uncertainty is used to propagate through to the uncertainties on all measured torques and stiffnesses in Sec. IV. The rotation of the float was measured with an optical lever arrangement with a laser reflecting off a small mirror attached to the center of the float and a position sensitive photodiode. This diode and its associated electronic circuit were then connected to a computer via an analog-to-digital converter (ADC). The computer was connected to the coils through a DAC. This allowed the coil currents to be actively controlled via a proportional-integral-differential (PID) control loop in LabVIEW software, and hence, the float was kept stable relative to a reference null position on the photodiode. A general diagram of this setup is shown in Fig. 4.

A photograph of the float, bearing, micrometers, mirror, compressed air input, coils, and magnets is shown in Fig. 5. Figure 6 shows a schematic drawing of how the bearing can be tilted from the horizontal plane, where this can be done along the float's x and y axes in the horizontal plane. Tilt in the y -axis by a known tilt angle, ξ , is depicted in Fig. 6.




FIG. 4. A general diagram of the full experimental setup.




FIG. 5. A photograph of the setup. A: float, B: bearing, C: micrometers, D: mirror, E: compressed air input, F: a coil, and G: a magnet.




FIG. 6. A side-on schematic drawing showing the float, bearing, and the micrometer used to tilt the whole apparatus by an amount, ξ , around the x -axis from the horizontal plane where $\xi = 0$. The vertical plane is defined as the direction of the free-fall acceleration due to gravity, g . Positive and negative values of ξ correspond to raising and lowering the setup from the horizontal plane, respectively. The global coordinate system defined in Fig. 2 is shown here. The coils have been omitted for clarity.

IV. RESULTS

The torque acting on the float could be measured from the PID servo output, whose control loop is shown in Fig. 7. The PID servo applied a torque, Γ , to the float by adding or subtracting from the bias currents in the appropriate coils. All eight coils were used for this purpose, and the bias currents are those described in Eqs. (14)–(16). The magnitude of this change in current from the bias currents could then be used in conjunction with Eq. (3) multiplied by the actuator arm length term, x_0 , from Table I to give

$$\Gamma = -8x_0m \frac{\partial B_\zeta}{\partial \zeta} \tag{17}$$

for eight coils. This measured torque is assumed to be equal to that described by Eq. (15). The rotational stiffness could be measured by recording the change in torque, $\Delta\Gamma$, from Eq. (17) applied by the PID servo to the float after an offset equivalent to a known angle, $\Delta\psi$, was added to the input of the controller. The rotational stiffness would then be given by

$$K_{\psi\psi} = -\frac{\Delta\Gamma}{\Delta\psi}, \tag{18}$$

where this rotational stiffness is assumed to be equal to that described by Eq. (10) once it had been summed over all the coils. The linear transverse stiffness of the float in the x - and y -axis of the




FIG. 7. A block diagram showing the PID servo control. The PID introduces a torque, Γ , to the float after a known offset angle, ψ , is added to the controller input. There will also be additional noise torques, γ , acting on the float.

horizontal plane, as described in Fig. 2, could be calculated using Eq. (4) summed over the bias currents of the four coils in each respective axis. As stated in Sec. II, the positioning of the magnets was chosen to give $\theta = \pi$.

Before torque measurements could be made, the float’s center-of-mass displacement from its NCB, x_{cm} from Eq. (15), had to be made zero. This was done by placing small masses on the float in precise positions such that when tilting it from the horizontal plane, the PID servo torque did not change within the limit of the servo readout noise. Doing this, while keeping the bias currents in all the coils equal such that ΔI was zero in Eq. (15), implied that x_{cm} was equal to zero.

The PID servo torque on the float was then measured over a range of tilt angles from the horizontal plane for a given set of bias currents. This was done in both the x - and y -axis of the float as shown in Fig. 2. The measurements in these axes are shown in Figs. 8 and 9, respectively.

The total mass of the float, with the additional small masses used for minimizing x_{cm} , was 338.66 g. This, along with the value of x_0 of (42.0 ± 0.5) mm from Table I, implied that the expected gradient of the plots from Eq. (16) should be (0.139 ± 0.003) N m/rad. The gradients from Figs. 8 and 9 are (0.137 ± 0.006) N m/rad and (0.138 ± 0.006) N m/rad, respectively. With the bias currents used and Eq. (14), the expected change of the center-of-buoyancy of the float in both the x - and y -axis was (7.00 ± 0.17) mm. Using the




FIG. 8. The torque, divided by the current ratio term, acting on the float over a range of tilt angles from the horizontal plane in the x -axis.




FIG. 9. The torque, divided by the current ratio term, acting on the float over a range of tilt angles from the horizontal plane in the y-axis.

gradients from Figs. 8 and 9, in conjunction with Eqs. (14) and (16), the measured changes in the center-of-buoyancy in the x- and y-axis were calculated to be (6.90 ± 0.32) mm and (6.95 ± 0.32) mm, respectively. This shows that we had succeeded in tuning the float’s center-of-buoyancy to an accuracy of ± 0.3 mm.

With the bias currents in all coils set to 0.075 mA, and Eq. (10) summed over all eight coils, the float’s rotational stiffness was calculated to be (-14.73 ± 0.41) $\mu\text{N m/rad}$. Using Eq. (18), it was measured to be (-15.11 ± 2.05) $\mu\text{N m/rad}$. With these bias currents, the float’s transverse stiffness in each horizontal axis, using Eq. (4) summed over the four coils in each axis, was calculated to be (1.11 ± 0.16) N/m. This gave a natural oscillation frequency in each axis of (0.29 ± 0.04) Hz. All the results are summarized in Table II.

TABLE II. A summary of the measurement results.

Attribute	Result
Plot gradients	
Float x-axis	(0.137 ± 0.006) N m/rad
Float y-axis	(0.138 ± 0.006) N m/rad
Equation (16) prediction	(0.139 ± 0.003) N m/rad
Center-of-buoyancy change	
Float x-axis	(6.90 ± 0.32) mm
Float y-axis	(6.95 ± 0.32) mm
Equation (14) prediction	(7.00 ± 0.17) mm
Float rotational stiffness	
Measurement	(-15.11 ± 2.05) $\mu\text{N m/rad}$
Equation (10) prediction	(-14.73 ± 0.41) $\mu\text{N m/rad}$
Float transverse stiffness and frequencies	
Equation (4) x- and y-axis calculation	(1.11 ± 0.16) N/m
Oscillation frequency x- and y-axis	(0.29 ± 0.04) Hz

V. DISCUSSION

The measurements in Figs. 8 and 9 change linearly with the tilt angle as expected from Eq. (16). The gradients of these plots were expected to be 0.139 N m/rad, and the measured gradients all lie within 1.5% of this value and within their uncertainty ranges. The calculated and measured changes in the center-of-buoyancy displacement of the float from Figs. 8 and 9 were also within 1.5% of each other and within each other’s uncertainty ranges. Thus, all the results are in excellent agreement with the theory. This demonstrates that it is possible to tune *in situ* the center-of-buoyancy of a suspended object and hence also possible to decouple its rotational motion from ground tilt and horizontal accelerations.

The measured and calculated rotational stiffnesses of the float were within 2.5% of each other and within their respective uncertainty bounds. The magnitude of these values is an order of magnitude lower than the rotational stiffness of a torsion balance used in a recent determination of Newton’s constant of gravitation, which had a stiffness of approximately 218 $\mu\text{N m/rad}$.¹⁵ Additional actuators could increase the float’s transverse stiffness from the calculated values to make the float more transversely stable and also allow the adjustment of its rotational stiffness. Preliminary measurements with this setup show that this is possible.¹³ The natural oscillation frequencies of the float in the transverse plane were calculated. With the float being levitated through an air bearing, it can be assumed that its motion was highly damped. As such in this case, there was no risk of the oscillations significantly affecting the float’s motion.

The design could be improved to allow greater precision in the placement and adjustment of the different components to give more accurate results. The noise from the electronics of the positional photodiode was measured to be 8.4×10^{-7} N m/ $\sqrt{\text{Hz}}$, while the total noise of the air bearing system was measured to be 2.7×10^{-6} N m/ $\sqrt{\text{Hz}}$. This was the largest contribution to the data point uncertainties in Figs. 8 and 9. This gave an error on the gradients of these plots, which was then propagated onto the experimental estimation of x_{cb} from Eq. (14). No attempt was made to put the device in a protective enclosure or shield the photodiode from environmental light sources. As such the error on x_{cb} of ± 0.3 mm could be reduced by lowering the noise of the air bearing. This could be done by using a different method of levitation other than a pressurized-air suspension, such as an electrostatic or a superconducting suspension.¹³ From Eq. (14), the theoretical maximum accuracy of center-of-buoyancy tuning that could be attained with the setup presented here, taking into account the current noise from the coil drivers of the magnetic actuators averaged over a second, is $\pm 1.6 \times 10^{-8}$ m. This corresponds to a possible improvement in accuracy of a factor of over 18 000.

VI. CONCLUSION

An air bearing suspension that levitates a float, with its motion in the horizontal plane of the laboratory controlled using magnetic actuators, was constructed. The observed behavior of the float was compared to the predictions of a detailed model of the statics of the float-actuator system, and they were found to be consistent. The results from Figs. 8 and 9 demonstrate the *in situ* electromagnetic tuning of the float’s center-of-buoyancy to an accuracy of ± 0.3 mm.

This result implies that it is practical to decouple the rotational mode of a suspended object from tilt and horizontal accelerations due to seismic noise, by tuning its center-of-buoyancy to lie at its nominal center-of-buoyancy (NCB).

This result paves the way for other, more sensitive, experiments to be designed with a view to performing weak-force measurements at sub-mm ranges. Work is ongoing on a superconducting torsion balance.^{13,18} The aim here is to develop an instrument that exhibits the same advantages as the air bearing where it can be tuned *in situ* to be rotationally decoupled from ground tilt, in addition to allowing the *in situ* tuning of its rotational stiffness. This combined with a transverse stiffness provided by the superconducting magnetic actuators should allow measurements of the inverse square law of gravity down to mass separations of the order of $10\ \mu\text{m}$. The Newtonian torque signal, given a day's integration, requires a fundamental noise level of less than $1 \times 10^{-14}\ \text{N m}/\sqrt{\text{Hz}}$. Given a typical seismic noise acceleration spectral density of $5 \times 10^{-7}\ \text{m/s}^2/\sqrt{\text{Hz}}$ and a suspended mass of 338.66 g, we would need to match the center-of-buoyancy and mass to an accuracy of about $1 \times 10^{-7}\ \text{m}$. So an improvement in tuning accuracy of a factor of around a thousand would be required, compared with what is achieved here. The uncertainty on the matching of the center-of-mass and center-of-buoyancy is limited by the overall noise in the air bearing system, so this goal may be achievable with a superconducting or other type of suspension. We should also mention that a possible downside of this technique is the way that the actuation system can introduce noise into the measurement through its asymmetries [the κ parameter introduced in Eq. (15)]. Any strategy of reducing this factor would include making such asymmetries as small as possible in the first instance, thus ensuring that the actuation torques themselves are as small as possible.

ACKNOWLEDGMENTS

We would like to thank Dr. Chris Collins for helpful discussions. We would also like to thank Masters Project students, Joseph Parker, Jack Joynson, Hugo Yamaguchi, and Arthur O'Leary, for

their help in this project. We also thank John Bryant and David Hoyland in helping to design and build the PID servo system in LabVIEW and the ADC/DAC setup for the experiment. We are grateful to UK STFC (Grant No. ST/F00673X/1) who initially supported this work. We are also grateful to Leverhulme (Grant No. RPG-2012-674) for financial support.

REFERENCES

- ¹H. Cavendish, *Philos. Trans. R. Soc. London* **88**(5-6), 469 (1798).
- ²R. D. Newman, E. C. Berg, and P. E. Boynton, *Space Sci. Rev.* **148**, 175 (2009).
- ³J. Murata and S. Tanaka, *Classical Quantum Gravity* **32**, 033001 (2015).
- ⁴S. K. Lamoreaux, *Rep. Prog. Phys.* **68**, 201 (2005).
- ⁵T. A. Wagner, S. Schlamminger, J. H. Gundlach, and E. G. Adelberger, *Classical Quantum Gravity* **29**, 184002 (2012).
- ⁶G. T. Gillies and R. C. Ritter, *Rev. Sci. Instrum.* **64**, 283 (1993).
- ⁷E. G. Adelberger, J. H. Gundlach, B. R. Heckel, S. Hoedl, and S. Schlamminger, in *Progress in Particle and Nuclear Physics*, edited by A. Faessler (Elsevier Science B.V., Amsterdam, The Netherlands, 2009), Vol. 62, No. 1, pp. 102–134.
- ⁸D. J. McManus, P. W. F. Forsyth, M. J. Yap, R. L. Ward, D. A. Shaddock, D. E. McClelland, and B. J. J. Slagmolen, *Classical Quantum Gravity* **34**, 135002 (2017).
- ⁹T. Shimoda and M. Ando, *Classical Quantum Gravity* **36**, 125001 (2019).
- ¹⁰C. M. Mow-Lowry and D. Martynov, *Classical Quantum Gravity* **36**, 245006 (2019).
- ¹¹C. C. Speake and D. B. Newell, *Rev. Sci. Instrum.* **61**, 1500 (1990).
- ¹²J. Harms and K. Venkateswara, *Classical Quantum Gravity* **33**, 234001 (2016).
- ¹³C. C. Speake and C. J. Collins, *Phys. Lett. A* **382**, 1069 (2018).
- ¹⁴J. Joynson, "Air-bearing suspension with passive magnetic stabilisation and tuning," M.S. thesis, University of Birmingham, 2019.
- ¹⁵T. Quinn, C. C. Speake, H. Parks, and R. Davis, *Philos. Trans. R. Soc. A* **372**, 20140032 (2014).
- ¹⁶W. R. Smythe, *Static and Dynamic Electricity*, 3rd ed. (Taylor & Francis, 1989).
- ¹⁷A. Aston and D. Hoyland, "OSEM interface control document," Advanced LIGO UK Internal Document LIGO-T050111-06-K, University of Birmingham, 2004.
- ¹⁸E. C. Chalkley, S. Aston, C. C. Collins, M. Nelson, and C. C. Speake, in *The XLVth Rencontres de Moriond and GPhys Colloquium 2011* (The Gioi Publishers, 2011), pp. 207–210.