

## High-throughput platelet spreading analysis

Khan, Abdullah; Maclachlan, Annabel; Lowe, Gillian; Nicolson, Pip; Al Ghaithi, Rashid Hafidh Rashid; Thomas, Steven; Watson, Steve; Pike, Jeremy; Morgan, Neil

DOI:

[10.3324/haematol.2019.225912](https://doi.org/10.3324/haematol.2019.225912)

License:

None: All rights reserved

*Document Version*

Publisher's PDF, also known as Version of record

*Citation for published version (Harvard):*

Khan, A, Maclachlan, A, Lowe, G, Nicolson, P, Al Ghaithi, RHR, Thomas, S, Watson, S, Pike, J & Morgan, N 2020, 'High-throughput platelet spreading analysis: a tool for the diagnosis of platelet-based bleeding disorders', *Haematologica*, vol. 105, no. 3, pp. e124-e128. <https://doi.org/10.3324/haematol.2019.225912>

[Link to publication on Research at Birmingham portal](#)

### General rights

Unless a licence is specified above, all rights (including copyright and moral rights) in this document are retained by the authors and/or the copyright holders. The express permission of the copyright holder must be obtained for any use of this material other than for purposes permitted by law.

- Users may freely distribute the URL that is used to identify this publication.
- Users may download and/or print one copy of the publication from the University of Birmingham research portal for the purpose of private study or non-commercial research.
- User may use extracts from the document in line with the concept of 'fair dealing' under the Copyright, Designs and Patents Act 1988 (?)
- Users may not further distribute the material nor use it for the purposes of commercial gain.

Where a licence is displayed above, please note the terms and conditions of the licence govern your use of this document.

When citing, please reference the published version.

### Take down policy

While the University of Birmingham exercises care and attention in making items available there are rare occasions when an item has been uploaded in error or has been deemed to be commercially or otherwise sensitive.

If you believe that this is the case for this document, please contact [UBIRA@lists.bham.ac.uk](mailto:UBIRA@lists.bham.ac.uk) providing details and we will remove access to the work immediately and investigate.

### High-throughput platelet spreading analysis: a tool for the diagnosis of platelet-based bleeding disorders

Bleeding disorders are an extremely heterogeneous group of conditions presenting a unique diagnostic challenge. While the gold-standard platelet function assay is light transmission aggregometry, platelet defects are only detected in approximately 50% of patients with a clinical history of bleeding consistent with a platelet disorder. We tested a cohort of patients recruited to the UK-GAPP (Genotyping and Phenotyping of Platelets) study with a suspected platelet function disorder using a high-throughput platelet spreading assay. We detected a platelet spreading defect in 32 out of 55 patients tested

(58%), and of these, 16 presented with normal lumiaggregometry results despite a significant Bleeding Assessment Tool (BAT) score. Furthermore, a family identified through this approach was subsequently identified as carrying a rare genetic variant of *TUBB1*, a gene linked to macrothrombocytopenia. This work suggests that morphological defects detected through a high-content platelet spreading approach can identify platelet dysfunctions not detected by lumiaggregometry.

Platelet function disorders are a heterogeneous group of conditions whose clinical and laboratory diagnosis is complicated by the range of reported bleeding symptoms, as well as functional redundancy within platelet signaling pathways. To address this, patients with suspected inherited bleeding disorders are often recruited for

**Table 1.** Phenotypes of all patients recruited to the GAPP study who were included in the platelet spreading investigations.

Individual/ Family	Age	Gender	Platelet count ( $\times 10^9/L$ ) <sup>a</sup>	Mean platelet volume (fL) <sup>b</sup>	ATP secretion (nmol ATP/ $1 \times 10^8$ platelets)*	Platelet defect based on lumiaggregometry findings	BAT score <sup>c</sup>	Bleeding phenotype
1	78	M	91		0.45	Thrombocytopenia	2	2, 4
2	20	F	182	12.5	2.35	Gi	5	4, 5, 11
3	72	M	113	10.2	1.05	Thrombocytopenia	13	1, 7, 8, 9
4	80	F	204	11.6	0.74	No defect	9	1, 2, 4, 5
5	6	F	230	10.3	0.39	Secretion	NA	1, 2
6.1	7	F	287	8.8	1.01	No defect	NA	
6.2	73	F	150	10.5	0.5	Secretion		
7	6	M	104	8.9	N+	Thrombocytopenia	NA	2, 3
8.1		F	103	15.0	0.21	Thrombocytopenia/Gq	8	1, 7, 9
8.2	35	F	107	15.1	0.26	Thrombocytopenia/Gq	4	1, 9
9	77	F	417	10.1	1.72	No defect	23	1, 2, 4, 6, 8, 11, 12
10	45	F	375	10	1.22	Gi	8	4, 5, 6, 7
11	54	F	239	11.9	0.74	No defect	16	2, 7, 8, 10, 11
12	49	F	289	11.5	1.3	No defect	6	4, 5, 11
13		F	270	12	1.11	No defect	7	2, 4, 8, 9, 11
14		F	352	10.5	1.27	No defect	15	4, 5, 7, 9, 11
15	12	F	355	8.3	0.75	No defect	NA	
16	50	M	270	11.4	1.39	Gi		1
17	22	F	323	11.5	0.66	No defect	3	1, 5, 6
18	19	F	269	11.9	0.73	No defect	16	1, 2, 4, 5, 7, 9, 11
19	26	F	295	10.8	0.89	No defect	3	1, 5
20	29	F	225	12.1	0.55	No defect	10	2, 4, 5, 10, 11
21	32	F	122		1.47	Thrombocytopenia/Gi	10	2, 4, 11, 12
22	49	F	195	10.4	0.63	No defect	14	2, 4, 5, 7, 11, 12
23	38	F	101	12.1		Thrombocytopenia		1, 2
24.1	23	M	428	9.1	0.39	Secretion	10	1, 4, 5, 7, 9
24.2	20	F	345	10.2	0.59	COX	12	2, 4, 5, 6, 11
24.3	17	M	329	10.9	0.25	Secretion	3	2, 5
25	76	M	120	14.6	0.97	Thrombocytopenia + COX	3	
26	43	F	331	10	0.56	Gi	14	1, 2, 4, 5, 11, 12
27	29	F	225	12.1	0.55	No defect	10	2, 4, 5, 11, 12
28	18	F	235	11	0.97	No defect	11	1, 2, 4, 5, 10, 11
29	53	F	298	11.1	0.72	No defect	13	2, 4, 7, 9, 11
30	32	F	292	9.7	0.61	COX		1, 2, 3, 7, 8, 12

continued on the next page

continued from the previous page

31	40	F	232	10.2	0.77	No defect	6	2, 7, 11
32		F	371	10.5	0.86	No defect	5	4, 5, 10, 11
33		F	281	11.3	1.04	No defect	12	2, 4, 5, 6, 8, 11
34	7	F	170	10.5	0.55	Secretion	NA	
35	68	F	43		0.34	Thrombocytopenia		
36.1	30	M	59	13		Thrombocytopenia	8	1, 4, 5, 7
36.2	52	M	183	12	0.66	No defect	2	5, 7
37	6	M	140	10.1	0.53	Thrombocytopenia/ secretion	NA	2, 3
38	36	F	221	11.7	1.13	No defect	13	1, 2, 4, 11, 12
39	57	F	261	9.3	0.82	COX	12	1, 5, 6, 11
40	22	F	291	11	1.03	No defect	13	2, 4, 8, 11
41.1	41	F	221	13.9		Secretion	NA	
41.2	66	M	107	large		Thrombocytopenia	NA	
42	57	F				Thrombocytopenia	NA	
43	52	F			1.49	No defect	16	1, 4, 5, 11, 12
44	20	F				No defect		2, 5, 9
45	54	F	238	10.9	0.92	No defect	7	2, 4, 5, 11, 12
46	27	F	256	11.1	0.75	No defect	9	1, 2, 5, 6, 7, 11
47	60	F	184	12.2	0.46	COX	21	1, 2, 4, 5, 7, 9, 10, 11,
48	23	F	204	12.4	0.59	Secretion	8	1, 2, 5, 10, 11
49	72	F	270	10.1	0.57	Secretion	13	1, 2, 6, 11

Families with multiple affected individuals are labeled .1 and .2 etc. \*Mean platelet count (normal reference range, 150-400x10<sup>9</sup>/L.), thrombocytopenia is defined as a platelet count <150x10<sup>9</sup>/L. \*Mean platelet volume (normal reference range, 7.83-10.5 fL). \*ATP secreted in response to 100  $\mu$ M PAR-1 receptor-specific peptide SFLLRN (5<sup>th</sup> centile in healthy volunteers is 0.65 nmol/1x10<sup>8</sup> platelets). Lumiaggregometry was used to simultaneously test platelet aggregation and secretion and platelet defects were categorized as cyclo-oxygenase deficiency (COX), Gi receptor signaling defect (Gi), Gq receptor signaling defect (Gq), secretion defect, thrombocytopenia and no defect. \*International Society for Thrombosis and Haemostasis Bleeding Assessment Tool (BAT) score, 95<sup>th</sup> percentile (score of 4) was calculated from healthy volunteers.<sup>7</sup> F: female; M: male; NA: not applicable e.g. for children. Code for bleeding symptoms: 1: epistaxis; 2: cutaneous bleeding; 3: petechiae; 4: oral cavity bleeding; 5: bleeding from minor wounds; 6: bleeding after tooth extraction; 7: bleeding after surgery; 8: hematuria; 9: gastrointestinal bleeding; 10: ovulation bleeding; 11: menorrhagia; 12: postpartum hemorrhage; 13: muscle hematomas.

platelet function tests (PFT) following initial assessment.<sup>1</sup> The main consideration when designing a panel of PFT is the best use of a limited quantity of sample. Despite the multitude of PFT available, a platelet defect is not always detected in patients with a significant bleeding history. Indeed, in patients with patterns of bleeding consistent with platelet disorders, we and others only find a platelet defect in approximately 50% of patients. To address the limitations of current approaches, the GAPP study performed a range of additional PFT, including impedance aggregometry (Multiplate<sup>®</sup>), 96-well platelet aggregation, aggregation on collagen at arterial shear rates (1000 s<sup>-1</sup>) and clot retraction assays, on a subset of patients.<sup>2,3</sup> Unfortunately this use of multiple PFT has not increased the number of patients identified with a platelet function disorder. Here we investigated a high-throughput analysis of platelets spread on fibrinogen to interrogate morphological and platelet defects.

The full methods are provided in the *Online Supplement*. Briefly, patients with bleeding of unknown cause were recruited to the GAPP study.<sup>4</sup> Peripheral blood collected from patients and healthy volunteers was used to generate platelet-rich plasma for lumiaggregometry and washed platelets at a concentration of 2x10<sup>7</sup>/mL for spreading on fibrinogen-coated coverslips. Spreading samples were incubated for 45 min before fixation, immunostaining and fluorescence microscopy. Captured images were subjected to a semi-automated image analysis pipeline using the open source software platforms KNIME and Ilastik. This produced platelet segmentations

and corresponding measurements for area and circularity.<sup>5,6</sup>


Patients were recruited with a variety of bleeding symptoms consistent with a platelet defect (n=55). The International Society on Thrombosis and Haemostasis (ISTH)-BAT score was used to provide a quantitative evaluation of the patients' bleeding history; scores were available for 40 individuals with the exception of pediatric patients. The overall mean BAT score was 9.825 (range, 2-23), with scores  $\geq 4$  being considered to be indicative of excessive bleeding.<sup>7</sup>

Lumiaggregometry was carried out as previously described.<sup>8,9</sup> On the basis of this, the platelet defects in the cohort were split into six categories; cyclo-oxygenase (COX) deficiency (n=4), Gi receptor signaling defect (n=4), secretion defect (n=8), thrombocytopenia (n=8), multiple defects (n=5) and no defects (n=26) (Table 1). Upon recruitment, platelet samples from patients were also subjected to spreading on fibrinogen. Samples were fixed and stained with phalloidin to delineate cell morphology simply and accurately with the high signal/noise ratio required for robust high-throughput imaging and analysis. A large scale dataset was generated by taking six fields of view from a representative area of each slide. To establish a control dataset, platelets from five healthy volunteers were treated identically. Images generated from these assays were subjected to a semi-automated segmentation workflow to measure area and circularity for each platelet, which were plotted with the median of the dataset (Figure 1). Area and circularity measurements

from the five representative healthy controls were used to define normal ranges. Patients' data which lay outside of these ranges were defined as indicating a potential defect.

Overall, abnormal spreading morphologies were found in 32 of 55 patients tested (58%). This consisted of patients showing defects in platelet area only (n=6),

platelet circularity only (n=13), both area and circularity (n=13) or no defect (n=23) (Figure 1). Abnormal area and circularity are consistent with round, unspread platelets, and such double-positive hits were considered robustly indicative of morphological spreading defects. Once identified as displaying a potential platelet spreading defect, samples were correlated to existing PFT and clinical data



**Figure 1. High-throughput measurement of platelet area and circularity.** Area and circularity measurements were generated using a semi-automated image analysis workflow and abnormalities were identified by comparison to a control range established by spreading data from healthy volunteers (red box) (n=5). Patients' samples with medians and confidence intervals outside of the control range were classified as positive hits (blue boxes) for either a defect in platelet area or circularity, while samples which fell outside both normal ranges were classified as double-positives (orange boxes).





one such patient a heterozygous *TUBB1* missense variant (p.Arg359Trp) was identified (patient 41.2) but not in the patient's sibling, 41.1. This genetic defect is also in keeping with the reduced platelet count only observed in patient 41.2 ( $107 \times 10^9/L$ ) and not in the sibling ( $221 \times 10^9/L$ ). *TUBB1* defects are known to cause macrothrombocytopenia<sup>11-14</sup> and the identification of this variant through platelet spreading analysis supports the efficacy of this tool in the study of suspected platelet function disorders. Interestingly, this finding is consistent with previous reports showing a minimal effect of cytoskeletal inhibition on platelet aggregation. This suggests that the absence of a defect in aggregometry in *TUBB1*-defective patients is likely due to a minor effect which overlaps with controls. In this instance, platelet spreading effectively detects a defect classically masked in aggregation experiments.<sup>15-17</sup>

This study systematically correlated high-throughput platelet spreading analysis with clinical and lumiaggregometry data to determine the former's utility in the investigation of platelet function disorders. We found that patients positive for both circularity and area abnormalities, consistent with the morphology of abnormally small and round platelets, are both positive and negative for aggregation defects. Eighty percent of patients with a spreading defect, but negative for a defect by aggregometry, had a significant bleeding score. Furthermore, within this cohort we identified *TUBB1* as the causative genetic variant in an individual with thrombocytopenia.

Of particular interest is that a large number of patients positive for platelet spreading defects and with a history of excessive bleeding are negative for an aggregation defect, suggesting that applying platelet spreading as part of the platelet testing panel could reveal morphological defects which would otherwise be missed by lumiaggregometry alone. Furthermore this platelet spreading assay is not dependent upon platelet count, as is the case with aggregation-based PFT<sup>8,9</sup> and could therefore be a useful tool to test for platelet function defects in patients with thrombocytopenia. Our approach validates the use of high-throughput platelet spreading analysis as an analytical tool to improve and augment existing PFT. As high-content imaging approaches become increasingly accessible and commonplace, applying high-throughput platelet spreading strategies is likely to be an invaluable addition to the current arsenal of PFT.

Abdullah O. Khan,<sup>1</sup> Annabel Maclachlan,<sup>1</sup> Gillian C. Lowe,<sup>1</sup> Phillip L.R. Nicolson,<sup>1</sup> Rashid Al Ghaithi,<sup>1</sup> Steven G. Thomas,<sup>1,2</sup> Steve P. Watson,<sup>1,2</sup> Jeremy A. Pike<sup>1,2</sup> and Neil V. Morgan<sup>1</sup> on behalf of the UK GAPP Study Group<sup>3</sup>

<sup>1</sup>AOK and AM contributed equally to this work.

*Funding: the work in the author's laboratories is supported by the British Heart Foundation (PG/13/36/30275; FS/15/18/31317; PG/15/14/31945) The authors would like to thank the Imaging Suite and COMPARE Core facilities at the University of Birmingham. The authors also acknowledge the help and support from Iain Styles at the University of Birmingham.*

<sup>1</sup>Institute of Cardiovascular Sciences, College of Medical and Dental Sciences, University of Birmingham; <sup>2</sup>Centre of Membrane and Protein and Receptors (COMPARE), University of Birmingham and

University of Nottingham, Midlands, UK

Correspondence: NEIL V. MORGAN  
N.V.Morgan@bham.ac.uk

doi:10.3324/haematol.2019.225912

Information on authorship, contributions, and financial & other disclosures was provided by the authors and is available with the online version of this article at [www.haematologica.org](http://www.haematologica.org).

## References

1. Bolton-Maggs PH, Chalmers EA, Collins PW, et al. A review of inherited platelet disorders with guidelines for their management on behalf of the UKHCDO. *Br J Haematol*. 2016;135(5):603-633.
2. Lordkipanidze M, Lowe GC, Kirkby NS, et al. Characterization of multiple platelet activation pathways in patients with bleeding as a high-throughput screening option: use of 96-well optima assay. *Blood*. 2014;123(8):e11-22.
3. Al Ghaithi R, Drake S, Watson SP, Morgan NV, Harrison P. Comparison of multiple electrode aggregometry with lumi-aggregometry for the diagnosis of patients with mild bleeding disorders. *J Thromb Haemost*. 2017;15(10):2045-2052.
4. Watson SP, Lowe GC, Lordkipanidze M, Morgan NV. Genotyping and phenotyping of platelet function disorders. *J Thromb Haemost*. 2013;11 (Suppl 1):351-363.
5. Sommer C, Straehle C, Köthe U, Hamprecht FA. Ilastik: Interactive learning and segmentation toolkit. Eighth IEEE International Symposium on Biomedical Imaging (ISBI). 2011;230-233.
6. Berthold MR, Cebron N, Dill F, et al. KNIME - the konstanz information miner: version 2.0 and beyond. *Explor Newsl*. 2009;11(1):26-31.
7. Lowe GC, Lordkipanidze M, Watson SP. Utility of the ISTH bleeding assessment tool in predicting platelet defects in participants with suspected inherited platelet function disorders. *J Thromb Haemost*. 2013;11(9):1663-1668.
8. Dawood BB, Wilde J, Watson SP. Reference curves for aggregation and ATP secretion to aid diagnosis of platelet-based bleeding disorders: effect of inhibition of ADP and thromboxane A<sub>2</sub> pathways. *Platelets*. 2007;18(5):329-345.
9. Dawood BB, Lowe GC, Lordkipanidze M, et al. Evaluation of participants with suspected heritable platelet function disorders including recommendation and validation of a streamlined agonist panel. *Blood*. 2012;120(25):5041-5049.
10. Johnson B, Lowe GC, Futterer J, et al. Whole exome sequencing identifies genetic variants in inherited thrombocytopenia with secondary qualitative function defects. *Haematologica*. 2016; 101(10):1170-1179.
11. Kunishima S, Kobayashi R, Itoh TJ, Hamaguchi M, Saito H. Mutation of the beta1-tubulin gene associated with congenital macrothrombocytopenia affecting microtubule assembly. *Blood*. 2009;113(2):458-461.
12. Burley K, Westbury SK, Mumford AD. *TUBB1* variants and human platelet traits. *Platelets*. 2018;29(2):209-211.
13. Kunishima S, Nishimura S, Suzuki H, Imaizumi M, Saito H. *Tubb1* mutation disrupting microtubule assembly impairs proplatelet formation and results in congenital macrothrombocytopenia. *Eur J Haematol*. 2014;92(4):276-282.
14. Fiore M, Goulas C, Pillois X. A new mutation in *TUBB1* associated with thrombocytopenia confirms that c-terminal part of  $\beta 1$ -tubulin plays a role in microtubule assembly. *Clin Genet*. 2017;91(6):924-926.
15. Natarajan P, May JA, Sanderson HM, Zabe M, Spangenberg P, Heptinstall S. Effects of cytochalasin h, a potent inhibitor of cytoskeletal reorganisation, on platelet function. *Platelets*. 2000; 11(8):467-476.
16. Torti M, Festetics ET, Bertoni A, Sinigaglia F, Balduini C. Agonist-induced actin polymerization is required for the irreversibility of platelet aggregation. *Thromb Haemost*. 1996;76(3):444-449.
17. Lefebvre P, White JG, Krumwiede MD, Cohen I. Role of actin in platelet function. *Eur J Cell Biol*. 1993;62(2):194-204.
18. Postma M, Goedhart J. Plots of data-a web app for visualizing data together with their summaries. *PLoS Biol*. 2009;17(3):e3000202.