

Accuracy of Vitalograph lung monitor as a screening test for COPD in primary care

Dickens, Andy; Fitzmaurice, David; Adab, Peymane; Sitch, Alice; Riley, Richard D.; Enocson, Alexandra; Jordan, Rachel

DOI:

[10.1038/s41533-019-0158-2](https://doi.org/10.1038/s41533-019-0158-2)

License:

Creative Commons: Attribution (CC BY)

Document Version

Publisher's PDF, also known as Version of record

Citation for published version (Harvard):

Dickens, A, Fitzmaurice, D, Adab, P, Sitch, A, Riley, RD, Enocson, A & Jordan, R 2020, 'Accuracy of Vitalograph lung monitor as a screening test for COPD in primary care', *NPJ Primary Care Respiratory Medicine*, vol. 30, no. 2, pp. 1-8. <https://doi.org/10.1038/s41533-019-0158-2>

[Link to publication on Research at Birmingham portal](#)

General rights

Unless a licence is specified above, all rights (including copyright and moral rights) in this document are retained by the authors and/or the copyright holders. The express permission of the copyright holder must be obtained for any use of this material other than for purposes permitted by law.

- Users may freely distribute the URL that is used to identify this publication.
- Users may download and/or print one copy of the publication from the University of Birmingham research portal for the purpose of private study or non-commercial research.
- User may use extracts from the document in line with the concept of 'fair dealing' under the Copyright, Designs and Patents Act 1988 (?)
- Users may not further distribute the material nor use it for the purposes of commercial gain.

Where a licence is displayed above, please note the terms and conditions of the licence govern your use of this document.

When citing, please reference the published version.

Take down policy

While the University of Birmingham exercises care and attention in making items available there are rare occasions when an item has been uploaded in error or has been deemed to be commercially or otherwise sensitive.

If you believe that this is the case for this document, please contact UBIRA@lists.bham.ac.uk providing details and we will remove access to the work immediately and investigate.

ARTICLE OPEN

Accuracy of Vitalograph lung monitor as a screening test for COPD in primary care

A. P. Dickens^{1*}, D. A. Fitzmaurice², P. Adab^{1*}, A. Sitch¹, R. D. Riley³, A. Enocson¹ and R. E. Jordan¹

Microspirometry may be useful as the second stage of a screening pathway among patients reporting respiratory symptoms. We assessed sensitivity and specificity of the Vitalograph® lung monitor compared with post-bronchodilator confirmatory spirometry (nDD Easy on-PC) among primary care chronic obstructive pulmonary disease (COPD) patients within the Birmingham COPD cohort. We report a case-control analysis within 71 general practices in the UK. Eligible patients were aged ≥ 40 years who were either on a clinical COPD register or reported chronic respiratory symptoms on a questionnaire. Participants performed pre- and post-bronchodilator microspirometry, prior to confirmatory spirometry. Out of the 544 participants, COPD was confirmed in 337 according to post-bronchodilator confirmatory spirometry. Pre-bronchodilator, using the LLN as a cut-point, the lung monitor had a sensitivity of 50.5% (95% CI 45.0%, 55.9%) and a specificity of 99.0% (95% CI 96.6%, 99.9%) in our sample. Using a fixed ratio of $FEV_1/FEV_6 < 0.7$ to define obstruction in the lung monitor, sensitivity increased (58.8%; 95% CI 53.0, 63.8) while specificity was virtually identical (98.6%; 95% CI 95.8, 99.7). Within our sample, the optimal cut-point for the lung monitor was $FEV_1/FEV_6 < 0.78$, with sensitivity of 82.8% (95% CI 78.3%, 86.7%) and specificity of 85.0% (95% CI 79.4%, 89.6%). Test performance of the lung monitor was unaffected by bronchodilation. The lung monitor could be used in primary care without a bronchodilator using a simple ratio of FEV_1/FEV_6 as part of a screening pathway for COPD among patients reporting respiratory symptoms.

npj Primary Care Respiratory Medicine (2020)30:2; <https://doi.org/10.1038/s41533-019-0158-2>

INTRODUCTION

Chronic obstructive pulmonary disease (COPD) is one of the most common long-term respiratory conditions with rising burden and mortality worldwide.^{1–3} It is characterised by increasing breathlessness and decline in lung function, punctuated by episodes of acute exacerbations that often lead to hospital admission and result in poor prognosis and gradual deterioration of quality of life.⁴ Annual healthcare and societal costs of COPD in Europe are estimated to be €48.4 billion.⁵ Despite the high burden of disease, the large majority of patients with COPD remain undiagnosed⁶ while experiencing significant morbidity,⁷ resulting in calls to improve early diagnosis.^{8,9} Early diagnosis could focus smoking cessation support and allow prescription of treatments that have been shown to reduce risk of exacerbation in those with COPD, thus has the potential to slow disease progression.

Screening programmes are not yet recommended, partly because of lack of evidence of the long-term benefits,^{10,11} a view which is upheld in the most recent UK National Screening Committee report.¹² However, there are also uncertainties around the performance of available screening tests, including symptom or risk assessment questionnaires and lung function-based measures, alone or in combination.^{12,13} A recent study compared different screening strategies among current smokers, against post-bronchodilator spirometry. This concluded that microspirometry or peak flow meters had the best performance, but interpretation was limited by a small sample size and low-quality spirometry data.¹⁴ Microspirometers are small relatively inexpensive handheld devices that measure forced expiratory volume in 1 s (FEV_1) and in 6 s (FEV_6). While this is not a substitute for confirmatory spirometry, which is more time consuming and measures FEV_1 and forced vital capacity (FVC), usually after bronchodilation, the FEV_1/FEV_6 ratio could be used as a pragmatic

initial screening test to identify patients requiring confirmatory spirometry. Microspirometry can be undertaken in office settings and requires less time and patient effort.^{15–18}

Over the past decade, several studies have explored the accuracy of microspirometers in detecting airflow obstruction.^{14,19–27} However, none of the studies considered the use of microspirometers as the second stage of a screening pathway. Microspirometry as a screening tool is usually performed without bronchodilation, as this contributes to time savings and avoids the need for Salbutamol. However, it remains uncertain how microspirometry performance differs when conducted pre- and post-bronchodilator. Finally, there is little consensus regarding the optimal FEV_1/FEV_6 cut-point for referral to confirmatory spirometry, with recent studies suggesting ratios of < 0.73 ,²² < 0.75 ²¹ and < 0.78 .¹⁹

To address the current evidence gaps, we conducted a study in primary care patients with existing respiratory symptoms, including those pre-screened in our linked trial. We aimed to assess the test performance of a microspirometer (Vitalograph Lung monitor) against confirmatory post-bronchodilator spirometry (nDD Easy on-PC) and explore the effect of using pre- or post-bronchodilator microspirometer data, the impact of using different airflow obstruction criteria and optimal cut-points.

RESULTS

Follow-up assessments were booked for 1633 participants. Out of the 1500 participants who attended the assessment, 551 took part in the case-control study. Lung monitor and spirometry test data were available for a total of 544 participants (Fig. 1).

Of those, 349 (64.2%) were male, the mean age was 69.6 (9.1) years, 517 (96.3%) were White British, 382 (74.3%) were

¹Institute of Applied Health Research, University of Birmingham, Birmingham, UK. ²Warwick Medical School – Health Sciences, University of Warwick, Coventry, UK. ³Centre for Prognosis Research, Research Institute for Primary Care and Health Sciences, Keele University, Keele, UK. *email: a.p.dickens@bham.ac.uk; p.adab@bham.ac.uk




Fig. 1 Flow of the participants.

overweight/obese and 472 (88.1%) had a positive smoking history. Similar proportions of participants reported Medical Research Council (MRC) Dyspnoea scores of 1–2 and 3–5, one fifth (19.7%) had CAT scores representing high impact on daily life and over half the sample (57.3%) were retired (Table 1).

A total of 337 (62.0%) participants had airflow obstruction according to the reference test, with over three quarters ($n = 264$, 78.3%) representing GOLD stage I and II i.e. mild/moderate COPD (Table 1). In comparison with controls, the cases were slightly older (70.4 vs 68.2 years), more likely to be male (68.3% vs 57.5%), a higher proportion had a positive smoking history (90.4% vs 84.3%) and MRC Dyspnoea scores of 3–5 (54.5% vs 40.1%) (Table 1).

Nearly half of the participants (45.5%) reported exacerbations in the past 12 months and 37 (6.8%) reported a respiratory hospitalisation in the past 2 years (Table 1). Cases reported approximately twice as many exacerbations as controls (54.2% vs 31.4%) in the past 12 months.

Screening accuracy of the pre-bronchodilator lung monitor

Using lower limit of normal (LLN; i.e. the lower 5th percentile) as the cut-off for a positive result, the pre-bronchodilator lung monitor had sensitivity of 50.5% (95% confidence interval (CI) 45.0, 55.9) and specificity of 99.0% (95% CI 96.6, 99.9) (Table 2). The positive predictive value of the lung monitor was estimated to be 76.9% (95% CI 55.7, 89.8) for a population prevalence of 6%, dropping to 61.8% (95% CI 37.8, 81.1) for a population prevalence of 3% and rising to 85.3% (95% CI 68.6, 93.9) for a population prevalence of 10%.

FEV₁ measurements from both devices were highly correlated ($r = 0.97$; $p < 0.001$), with the Bland–Altman plot demonstrating good agreement (Supplementary Fig. 1). Comparison of FEV₆ from both devices again revealed high correlation ($r = 0.95$; $p < 0.001$), though agreement was lower, indicating that the lung monitor systematically underestimated FEV₆ values by -0.37 litres (Supplementary Fig. 2).

The lung monitor had high discriminatory accuracy ($C = 0.90$; 95% CI 0.88, 0.93) between cases and controls according to confirmatory spirometry FEV₁/FVC < LLN (Supplementary Fig. 3).

Screening accuracy of the post-bronchodilator lung monitor

Using post-bronchodilator data for the lung monitor, the sensitivity was 46.6% (95% CI 41.2%, 52.1%) and specificity was 97.1% (95% CI 93.8%, 98.9%) for detecting airflow obstruction (Table 3). The positive predictive value of the lung monitor was estimated to be 50.6% (95% CI 42.6, 58.6) for a population prevalence of 6%.

Comparison of pre- and post-bronchodilator lung monitor test accuracy revealed a borderline significant difference favouring pre-bronchodilator of 3.9% in sensitivity (7.8%, -0.05%); $p = 0.04$, but no statistically significant evidence of a difference in specificity (1.9% (-0.9% , 4.7%)); $p = 0.10$.

Lung monitor and confirmatory spirometry tests were also highly correlated for post-bronchodilator FEV₁ ($r = 0.97$; $p < 0.001$), with the Bland–Altman plot again demonstrating good agreement (Supplementary Fig. 4). Comparison of post-bronchodilator FEV₆ again revealed high correlation ($r = 0.97$; $p < 0.001$), though agreement was lower, indicating that the lung monitor

Table 1. Description of analysis sample, stratified by cases and controls.

	Analysis sample (n = 544)	Cases; <LLN (n = 337)	Controls; ≥LLN (n = 207)
Demographic and lifestyle information			
Sex; n (%) male	349/544 (64.2)	230/337 (68.3)	119/207 (57.5)
Age, mean (SD)	69.6 (9.1)	70.4 (7.9)	68.2 (10.6)
Ethnicity; n (%)			
White British	517/537 (96.3)	322/332 (97.0)	195/205 (95.1)
Other	20/537 (3.7)	10/332 (3.0)	10/205 (4.9)
Weight; n (%)			
Underweight (BMI <18.5)	9/514 (1.8)	7/318 (2.2)	2/196 (1.0)
Healthy (≥18.5 and <25)	123/514 (23.9)	85/318 (26.7)	38/196 (19.4)
Overweight (≥25 and <30)	194/514 (37.7)	118/318 (37.1)	76/196 (38.8)
Obese (≥30)	188/514 (36.6)	108/318 (34.0)	80/196 (40.8)
BMI, mean (SD) n = 514	28.6 (5.4)	28.1 (5.4)	29.4 (5.4)
Smoking status; n (%)			
Current smoker	111/536 (20.7)	79/332 (23.8)	32/204 (15.7)
Ex-smoker	361/536 (67.4)	221/332 (66.6)	140/204 (68.6)
Never smoker	64/536 (11.9)	32/332 (9.6)	32/204 (15.7)
Employment status; n (%)			
Employed	88/534 (16.5)	45/334 (13.5)	43/200 (21.5)
Unemployed	140/534 (26.2)	94/334 (28.4)	46/200 (23.0)
Retired	306/534 (57.3)	195/334 (58.4)	111/200 (55.5)
Airflow obstruction			
GOLD stage if <LLN			
1 (FEV ₁ ≥80% pred)	87/337 (25.8)	87/337 (25.8)	n/a
2 (50–79%)	177/337 (52.5)	177/337 (52.5)	n/a
3 (30–49%)	59/325 (18.1)	61/337 (18.1)	n/a
4 (<30%)	12/325 (3.6)	12/337 (3.6)	n/a
Self-reported health and healthcare usage			
MRC; n (%)			
1–2	219/427 (51.3)	116/255 (45.5)	103/172 (59.9)
3–5	208/427 (48.7)	139/255 (54.5)	69/172 (40.1)
CAT impact level; n (%)			
Low (0–9)	180/406 (44.3)	100/245 (40.8)	80/161 (49.7)
Medium (10–20)	146/406 (36.0)	93/245 (38.0)	53/161 (32.9)
High (21–30)	69/406 (17.0)	47/245 (19.2)	22/161 (13.7)
Very high (31–40)	11/406 (2.7)	5/245 (2.0)	6/161 (3.7)
Asthma; n (%) yes	198/538 (36.8)	129/334 (38.6)	69/204 (33.8)
CVD ^a ; n (%) yes	159/538 (29.6)	94/334 (28.1)	65/204 (31.9)
Other comorbidities ^b ; n (%) ≥1	302/540 (55.9)	184/336 (54.8)	111/204 (57.8)
Exacerbations in last 12 m; n (%) yes	247/543 (45.5)	182/336 (54.2)	65/207 (31.4)
Respiratory hospitalisation in last 2 years; n (%) yes	37/544 (6.8)	27/337 (8.0)	10/207 (4.8)

LLN = below the 5th percentile of the predicted FEV₁/FVC ratio using the NHANES III equations

^aCVD was defined as self-reported coronary heart disease, heart failure or other heart condition

^bOther comorbidities were defined as self-reported clinician diagnosis of diabetes, osteoporosis, depression, fracture

Table 2. Pre-BD lung monitor (FEV₁/FEV₆ < LLN) against post-BD confirmatory spirometry (FEV₁/FVC < LLN).

	Spirometry +ve	Spirometry –ve	
Lung monitor +ve	170 (50.5%)	2 (1.0%)	172
Lung monitor –ve	167 (49.6%)	205 (99.0%)	372
	337	207	544

Table 3. Post-BD lung monitor (FEV₁/FEV₆ < LLN) against post-BD confirmatory spirometry (FEV₁/FVC < LLN).

	Spirometry +ve	Spirometry –ve	
Lung monitor +ve	157 (46.6%)	6 (2.9%)	163
Lung monitor –ve	180 (53.4%)	201 (97.1%)	381
	337	207	544

Table 4. Screening accuracy of pre-bronchodilator lung monitor FEV₁/FEV₆ cut-points, against post-BD confirmatory spirometry (FEV₁/FVC < LLN).

	TP	FP	TN	FN	Sens (95% CI)	Spec (95% CI)	PPV (3% prevalence)	PPV (6% prevalence)	PPV (10% prevalence)	Proportion referred to diagnostic test ^a	Proportion of true cases missed ^a
<0.4	6	0	207	331	1.79 (0.7, 3.8)	100 (98.2, 100)	—	—	—	1.1%	98.2%
<0.5	34	0	207	303	10.1 (7.1, 13.8)	100 (98.2, 100)	—	—	—	6.3%	89.9%
<0.6	88	1	206	249	26.1 (21.5, 31.1)	99.5 (97.3, 100)	0.63 (0.19, 0.92)	0.78 (0.33, 0.96)	0.86 (0.46, 0.98)	16.4%	73.9%
<0.7	197	3	204	140	58.5 (53.0, 63.8)	98.6 (95.8, 99.7)	0.56 (0.39, 0.70)	0.72 (0.57, 0.83)	0.82 (0.70, 0.90)	36.8%	41.5%
<0.71	207	4	203	130	61.4 (56.0, 66.6)	98.1 (95.1, 99.5)	0.50 (0.38, 0.61)	0.67 (0.56, 0.77)	0.78 (0.69, 0.85)	38.8%	38.6%
<0.72	220	7	200	117	65.3 (59.9, 70.4)	96.6 (93.2, 98.6)	0.37 (0.31, 0.44)	0.55 (0.48, 0.62)	0.68 (0.62, 0.74)	41.7%	34.7%
<0.73	234	9	198	103	69.4 (64.2, 74.3)	95.7 (91.9, 98.0)	0.33 (0.29, 0.38)	0.50 (0.45, 0.56)	0.64 (0.59, 0.69)	44.7%	30.6%
<0.74	244	12	195	93	72.4 (67.3, 77.1)	94.2 (90.1, 97.0)	0.28 (0.25, 0.31)	0.44 (0.41, 0.48)	0.58 (0.54, 0.62)	47.1%	27.6%
<0.75	253	17	190	84	75.1 (70.1, 79.6)	91.8 (87.2, 95.1)	0.22 (0.20, 0.24)	0.37 (0.34, 0.39)	0.50 (0.48, 0.53)	49.6%	24.9%
<0.76	264	25	182	73	78.3 (73.6, 82.6)	87.9 (82.7, 92.0)	0.17 (0.16, 0.18)	0.29 (0.28, 0.31)	0.42 (0.40, 0.44)	53.1%	21.7%
<0.77	273	29	178	64	81.0 (76.4, 85.1)	86.0 (80.5, 90.4)	0.15 (0.14, 0.16)	0.27 (0.26, 0.28)	0.39 (0.38, 0.41)	55.5%	19.0%
<0.78	279	31	176	58	82.8 (78.3, 86.7)	85.0 (79.4, 89.6)	0.15 (0.14, 0.15)	0.26 (0.25, 0.27)	0.38 (0.37, 0.39)	57.0%	17.2%
<0.79	287	41	166	50	85.2 (80.9, 88.8)	80.2 (74.1, 85.4)	0.12 (0.11, 0.12)	0.22 (0.21, 0.22)	0.32 (0.31, 0.33)	60.3%	14.8%
<0.8	295	48	159	42	87.5 (83.5, 90.9)	76.8 (70.5, 82.4)	0.10 (0.10, 0.11)	0.19 (0.19, 0.20)	0.30 (0.29, 0.30)	63.1%	12.5%
<0.9	333	182	25	4	98.8 (97.0, 99.7)	12.1 (8.0, 17.3)	0.03 (0.03, 0.03)	0.07 (0.07, 0.07)	0.11 (0.11, 0.11)	94.7%	1.2%

TP true positives, FP false positives, TN true negatives, FN false negatives, Sens sensitivity, Spec specificity, PPV positive predictive value
^aAssuming 6% COPD prevalence

systematically underestimated FEV₆ values by −0.28 litres (Supplementary Fig. 5).

Discriminatory accuracy of the post-bronchodilator lung monitor was identical to that based on pre-bronchodilator data ($C = 0.90$; 95% CI 0.87, 0.92; Supplementary Fig. 6).

Sensitivity analyses: optimal cut-points for lung monitor FEV₁/FEV₆ ratio, relative to confirmatory spirometry FEV₁/FVC < LLN

In light of comparable test accuracy of the lung monitor based on pre- and post-bronchodilator data, we explored optimal FEV₁/FEV₆ cut-points using pre-bronchodilator tests (Table 4).

Using an FEV₁/FEV₆ cut-point of <0.7 to define a positive test for the lung monitor, sensitivity increased to 58.5% (95% CI 53.0, 63.8), specificity was 98.6% (95% CI 95.8, 99.7) and the positive predictive value increased to 72.0% (95% CI 57.4, 83.1) for a population prevalence of 6% (Table 4). However, using a fixed ratio had little effect on the discriminatory accuracy of the lung monitor ($C = 0.91$; 95% CI 0.89, 0.94; Supplementary Fig. 7).

In our sample, an FEV₁/FEV₆ cut-point of <0.78 had the best overall test performance with sensitivity of 82.8% (95% CI 78.3%, 86.7%) and specificity of 85.0% (95% CI 79.4%, 89.6%). Using this cut-point would result in the lung monitor only missing 17.2% of true positives and correctly identifying the majority of patients without the disease. Furthermore, this cut-point would result in 57% of those screened requiring confirmatory spirometry. The positive predictive value for a population COPD prevalence of 6% was estimated to be 26.1% (95% CI 25.0, 27.2) meaning that around one in four patients referred for confirmatory spirometry would result in a diagnosis.

The above pattern was broadly similar when analyses were repeated using the fixed ratio to define obstruction for confirmatory spirometry (FEV₁/FVC < 0.7), though sensitivity was slightly lower at each cut-point and specificity remained at 100% until FEV₁/FEV₆ > 0.7 (Supplementary Table 1). These analyses may reflect the test performance when using the simpler criterion for the lung monitor in countries defining airflow obstruction as FEV₁/FVC < 0.7, such as the UK.²⁸

DISCUSSION

We found that the lung monitor has high discriminatory accuracy among patients with existing chronic respiratory symptoms. This supports its suitability, either alone or perhaps in combination with a symptom questionnaire, as a screening test prior to confirmatory spirometry. We further demonstrated that using a bronchodilator with the lung monitor as part of screening offers no performance advantage.

Importantly, the lung monitor demonstrated good test performance despite being delivered with minimal coaching and only requiring a maximum of three blows, rather than the possible six blows to achieve repeatability with confirmatory spirometry.

Using pre-bronchodilator FEV₁/FEV₆ < LLN, the lung monitor missed half of COPD cases identified by FEV₁/FVC < LLN from confirmatory spirometry but detected virtually all non-COPD cases correctly. When using pre-bronchodilator FEV₁/FEV₆ < 0.70, the lung monitor detected a higher proportion of true positives, the same proportion of true negatives and the discriminatory accuracy remained constant ($C = 0.90$ vs $C = 0.91$). Given the added complexity of applying LLN to the lung monitor as it is not connected to computer software, it appears justifiable to apply an FEV₁/FEV₆ fixed ratio to the lung monitor for purposes of screening, while maintaining the LLN for diagnosing and monitoring COPD.^{29–32}

Test performance varied considerably depending on the specified cut-point of the pre-bronchodilator FEV₁/FEV₆ ratio. Our proposed optimal cut-point of <0.78 was similar to previous studies, which had suggested using cut-points of <0.75,²¹ <0.78¹⁹ and <0.80.²⁰ The sensitivity and specificity of the lung monitor in our sample was acceptable for a screening test, missing <20% of COPD cases, while 1 in 4 patients of the 57% referred for confirmatory spirometry were true positives and therefore would be eligible for diagnosis and relevant treatment. While FEV₁/FEV₆ < 0.78 appeared the most efficient in our sample, if the lung monitor were to be used as a screening test the cut-point could be modified according to the balance of acceptable false negative rates and availability of resources.

We have assessed the screening test performance of one type of microspirometer. One factor affecting accuracy may be the

different lung function indices being measured: FEV₆ by the lung monitor vs FVC by the ndd device (confirmatory spirometry). We assessed test performance of both devices using FEV₁/FEV₆ < LLN as the cut-off for a positive result, relative to confirmatory spirometry FEV₁/FVC < LLN. The ndd device had sensitivity of 80.4% and specificity of 98.1%, compared with the lung monitor sensitivity of 50.5% and specificity of 99.0%. This suggests that the difference in indices only partly affects performance. Another important difference to consider is the type of sensor used in the two devices for flow/volume measurement (turbine in the lung monitor vs ultrasonic in the ndd), as evidence suggests a degree of inaccuracy in turbine devices.^{33,34}

Our analysis sample had fewer controls than determined by our sample size calculation, containing 207 instead of 248. While the precision around specificity was reduced, the precision around sensitivity estimates was unaffected; the latter being arguably more important in the context of screening.

Using the LLN criteria to define cases in our primary analysis ensured an accurate assessment of lung function, without added 'noise' from misdiagnosed patients which can be introduced when using the FEV₁/FVC < 0.7 ratio.³² As the majority of previous microspirometry test accuracy studies used the fixed ratio definition of obstruction,^{19–21,24,25,27,35} our study has made a valuable contribution to the evidence base.

Owing to the case–control study being nested within a larger COPD cohort study, the analysis sample had a higher prevalence of COPD and possibly more advanced disease than would be observed in an undiagnosed primary care population reporting respiratory symptoms. Therefore, our study is at potential risk of spectrum bias, as the reported sensitivities and specificities may not fully reflect the test performance of the lung monitor if used as a screening tool within symptomatic patients with lower prevalence of COPD. However, by using Bayes' Theorem the reported post-test estimates were based on current UK COPD prevalence of 3–10%, mitigating against this risk.

Nearly a third of our sample was a screened population, suggesting that our findings will resonate with potential screening processes, as patients could be selected for microspirometry on the basis of symptom- or risk-based screening tests. Furthermore, the fact that we included patients with chronic respiratory symptoms and a range of lung function severities means that our results may apply to an undiagnosed population with a similar symptom profile. In addition, our study was not restricted to ever-smokers, unlike previous studies.^{14,19,21,22,25}

For practical reasons, the same researcher administered both the lung monitor and confirmatory spirometry. Although researchers only recorded raw FEV₁ and FEV₆ lung monitor values and did not calculate obstruction from this first test, it is possible that researchers were not entirely blind when administering the confirmatory spirometry to the patient. While this introduced a risk of review bias, this was minimised as researchers received standardised training to give only brief instruction for lung monitor tests and proper coaching for confirmatory spirometry.

Most previous studies have either used only pre-bronchodilator microspirometry^{19–21,23–26} or post-bronchodilator microspirometry,²⁷ and the only study to measure pre- and post-bronchodilator microspirometry did not report comparative test accuracy.²² By demonstrating the comparability of test performance irrespective of bronchodilation, our study supports the continued use of pre-bronchodilator microspirometers for screening purposes.

Participants performed three blows using the lung monitor, irrespective of blow quality as indicated by the device's in-built quality alert, with the highest recorded readings being used for analyses. While this follows some previous studies,^{23,25,26} had we required all lung monitor blows to be technically valid^{19,20,22} we may have obtained greater FEV₁ or FEV₆ values for some participants. Furthermore, like most studies we did not assess

within-participant repeatability across blows on the lung monitor, though this has been done in at least one study.²¹

The observed test performance of the lung monitor suggests that it could be reliably used as a screening tool in patients perceived to be at risk of COPD, to select those requiring confirmatory spirometry. The efficiency of the diagnostic spirometry test could therefore be substantially increased, by patients highly unlikely to have airflow obstruction being screened out in advance. Screening at-risk symptomatic patients with a lung monitor rather than referring all patients for confirmatory spirometry also represents financial savings, with the handheld device being approximately one tenth of the cost of diagnostic spirometers. Resource savings could be realised in practices irrespective of whether they conduct confirmatory spirometry 'in house' or refer patients to a lung function unit, as both models would reduce the number of patients performing this diagnostic test.

The ability to use a fixed ratio for the lung monitor rather than the LLN to assess airflow obstruction represents a time saving for clinicians, who would otherwise need to use software to refer to reference equations. The comparable test performance of the lung monitor irrespective of bronchodilation supports the use of pre-bronchodilator tests, further contributing to the efficiency and ease of the screening test, a key consideration in the context of time-pressured primary care consultations.

The lung monitor could potentially be administered by any member of a primary care team, as it is a simple device requiring minimal training. This would be beneficial in general practice where staff may be unfamiliar with the device¹⁹ and the simplicity may minimise the risk of becoming de-skilled in using the lung monitor, in contrast to confirmatory spirometry where clinicians' skills can reduce over time if they do not perform the test regularly.³⁶

The simplicity of the lung monitor, the minimal number of required blows and its good test performance suggests that it could be particularly useful as a screening test in patients with poor coordination or lower cognitive ability. Furthermore, our Patient Advisory Group preferred the lung monitor over other microspirometer models suggesting that it may be more acceptable to patients.

While we have suggested optimal cut-points based on the balance of sensitivity and specificity, in practice, the optimal cut-point would be determined by the clinical setting in which the lung monitor was being used. For example, in settings where access to quality confirmatory spirometry may not be available, particularly in low-resource settings, specificity of the lung monitor may be prioritised. In these settings, using thresholds with higher specificity could effectively exclude the majority of those with respiratory symptoms who do not have COPD, thus preventing overdiagnosis.

In addition to use as a screening tool, the accurate measurement of FEV₁ may indicate that the device could be used to monitor obstruction severity or lung function decline among diagnosed COPD patients, for example during annual reviews. Further research would be needed to explore this, but the potential time and cost savings afforded by using the lung monitor instead of confirmatory spirometry may be attractive to General Practitioner practices, who would still obtain annual FEV₁ values as recommended by bodies such as the National Institute for Health and Care Excellence in the UK.²⁸

Future research could build on preliminary evidence regarding microspirometer screening strategies,^{13,14} which could be implemented in differing clinical or economic contexts. Using a combination of microspirometry and screening questionnaires for example may prove more efficient than microspirometry alone. Furthermore, rather than using one cut-point to identify patients requiring confirmatory spirometry, certain contexts may warrant using two cut-points to refer only those patients where there is

uncertainty about their diagnosis. For example, in low- and middle-income countries where availability of confirmatory spirometry may be limited, a three-tiered approach may be plausible whereby the top proportion of patients are defined test negative, the bottom proportion are defined as test positive and the middle proportion are referred for confirmatory spirometry.

Our results show that the Vitalograph lung monitor, which is a cheap and simple device, has acceptable accuracy for use within a screening pathway for undiagnosed COPD among primary care patients with respiratory symptoms. We have established that the test performance of the lung monitor is unaffected by bronchodilation, and our optimum cut-point of $FEV_1/FVC < 0.78$ supports previous studies, with no observed advantage of using LLN for this screening test. Our paper makes a valuable contribution to the evidence base concerning potential COPD screening tests, though more work is required to inform the need for a formal screening programme.

METHODS

Study design

We conducted a prospective case–control study to evaluate the screening performance of the Vitalograph® lung monitor (Vitalograph Ltd, Buckingham, UK), nested within a large COPD Cohort study.

Participant recruitment

Study participants were drawn from those attending for their 3-year follow-up assessment as part of the Birmingham COPD Cohort Study, which has been reported in detail elsewhere.³⁷ In brief, participants were primary care patients aged ≥ 40 years, who either had previously clinically diagnosed COPD or had reported chronic respiratory symptoms as part of a case-finding trial.³⁸ Participants from the case-finding trial were invited to join the Cohort study, irrespective of their spirometry results, if they reported chronic cough or phlegm for ≥ 3 months for at least 2 years, wheeze in the past 12 months or dyspnoea of MRC grade ≥ 2 .

At the 3-year follow-up assessment visit, cohort participants were invited to take part in the additional tests for this case–control study (Fig. 2) and those who agreed were asked to sign a consent form. Those who declined to participate completed the standard Cohort assessment. The National Research Ethics Service Committee West Midlands, Solihull provided approval for both the Birmingham COPD Cohort (11/WM/0304) and the case-finding trial (11/WM/0403).

Data collection and clinical measures

In addition to the lung function tests described below, participants underwent the standard Cohort follow-up assessment, which included

various physiological and anthropometric measurements (height, weight, grip strength, exercise capacity) as well as completing questionnaires.

Index test: lung monitor microspirometry

Participants received pre- and post-bronchodilator microspirometry with the Vitalograph lung monitor prior to confirmatory post-bronchodilator spirometry (Fig. 1). The lung monitor measured FEV_1 and FEV_6 in litres. In contrast to confirmatory spirometry, participants received minimal explanation or coaching when using the lung monitor. Researchers told participants to take a deep breath until lungs were full and blow into the mouthpiece as hard and fast as they could until being told to stop. Researchers demonstrated the correct technique once and then allowed the participant to perform the blows themselves, without additional coaching or encouragement. Participants performed three blows pre-bronchodilator and three blows post-bronchodilator. Technically unsatisfactory blows identified by the in-built quality assessment were recorded on the case report form, but participants were not asked to repeat the blow. The best FEV_1 and FEV_6 blows were used for analyses, irrespective of quality and which blow attempt they came from.

Positive test results were defined as being below the 5th percentile of the predicted pre-bronchodilator FEV_1/FEV_6 ratio (i.e. the LLN) using the NHANES III equations.³⁹ Alternative positive test results were also pre-specified, including post-bronchodilator FEV_1/FEV_6 below the LLN, and various cut-points of the FEV_1/FEV_6 ratio.

Reference test: post-bronchodilator confirmatory spirometry

Post-bronchodilator confirmatory spirometry was conducted according to American Thoracic Society and European Respiratory Society 2005 guidelines⁴⁰ by trained researchers using the ndd Easy on-PC spirometer. Participants received 400 μg of Salbutamol and after waiting at least 20 min, performed a minimum of 3 (maximum of 6) blows until repeatability was achieved. Although the lung monitor and spirometry tests were administered by the same researcher, the tests were in effect administered blind of each other, as researchers did not record the FEV_1/FEV_6 ratio for the lung monitor before administering confirmatory spirometry.

Cases were defined as participants whose predicted FEV_1/FVC ratio was below the LLN using the NHANES III equations, according to confirmatory spirometry. Participants not meeting this criterion formed the controls.

Aims

The primary aim was to assess the pre-bronchodilator test accuracy (sensitivity and specificity) of the lung monitor (FEV_1/FEV_6) against post-bronchodilator confirmatory spirometry (FEV_1/FVC), using the LLN definition of airflow obstruction.

We also aimed to assess the correlation and agreement between lung function measures from both devices and to compare test accuracy of pre-




Fig. 2 Case–control study design.

and post-bronchodilator lung monitor data. Finally, to identify the threshold that optimised sensitivity and specificity, we explored the effect of using different FEV₁/FEV₆ thresholds, including the fixed ratio of <0.7, to define a positive test result on the lung monitor.

Sample size

We calculated that we required a sample size of 248 cases and 248 controls to detect an assumed sensitivity of 85%^{21,27} while ensuring the lower bound of the CI was >80%.

Statistical analysis

We evaluated the diagnostic test accuracy of the lung monitor (index test) for all participants with complete data for the index and reference tests. We estimated sensitivity and specificity of the lung monitor using pre-bronchodilator data. We compared test accuracy of pre- and post-bronchodilator lung monitor blows, using McNemar's test. Using continuous test values, we assessed the discriminatory accuracy of FEV₁ and FEV₆ measured by the lung monitor via receiver operating characteristic curve analysis. We then conducted sensitivity analyses using a fixed ratio definition of obstruction to identify a lung monitor FEV₁/FEV₆ optimal threshold.

To account for the case-control study design, post-test probabilities (herein referred to as positive predictive values (PPV)) were calculated using Bayes' Theorem to reflect current COPD prevalence in the UK. For our tables and appendices, we calculated PPVs based on COPD prevalence among adults aged ≥40 years being 3–10%.^{1,41}

All analyses were conducted in Stata SE v15.

The paper was written according to the STARD guidance⁴² for reporting studies of diagnostic accuracy.

DATA AVAILABILITY

The data that support the findings of this study are available on request from the corresponding authors (A.P.D. or P.A.).

CODE AVAILABILITY

No custom code or mathematical algorithms were used in the analyses, which adopted only standard data manipulation and analysis within the statistical software.

Received: 14 May 2019; Accepted: 13 November 2019;

Published online: 03 January 2020

REFERENCES

- Halbert, R. J. et al. Global burden of COPD: systematic review and meta-analysis. *Eur. Respir. J.* **28**, 523–532 (2006).
- Chapman, K. R. et al. Epidemiology and costs of chronic obstructive pulmonary disease. *Eur. Respir. J.* **27**, 188–207 (2006).
- Lozano, R. et al. Global and regional mortality from 235 causes of death for 20 age groups in 1990 and 2010: a systematic analysis for the Global Burden of Disease Study 2010. *Lancet* **380**, 2095–2128 (2012).
- Global Initiative for Chronic Obstructive Lung Disease. *Global Strategy for the Diagnosis, Management, and Prevention of Chronic Obstructive Pulmonary Disease*. 2017 Report (Global Initiative for Chronic Obstructive Lung Disease, 2017).
- Gibson, G. J., Loddenkemper, R., Sibille, Y. & Lundback, B. Respiratory health and disease in Europe: the new European Lung White Book. *Eur. Respir. J.* **42**, 559–563 (2013).
- Mannino, D. M., Gagnon, R. C., Petty, T. L. & Lydick, E. Obstructive lung disease and low lung function in adults in the United States: data from the National Health and Nutrition Examination Survey, 1988–1994. *Arch. Intern. Med.* **160**, 1683–1689 (2000).
- Labonte, L. E. et al. Undiagnosed chronic obstructive pulmonary disease contributes to the burden of health care use. data from the CanCOLD study. *Am. J. Respir. Crit. Care Med.* **194**, 285–298 (2016).
- Soriano, J. B., Zielinski, J. & Price, D. Screening for and early detection of chronic obstructive pulmonary disease. *Lancet* **374**, 721–732 (2009).
- Lamprecht, B. et al. Determinants of underdiagnosis of COPD in national and international surveys. *Chest* **148**, 971–985 (2015).
- US Preventive Services Task Force. Screening for chronic obstructive pulmonary disease: US Preventive Services Task Force Recommendation Statement. *JAMA* **315**, 1372–1377 (2016).
- UK National Screening Committee. An evaluation of screening for COPD against the National Screening Committee criteria (2013).
- Solutions for Public Health. Screening for chronic obstructive pulmonary disease (COPD) in the general adult population. External review against programme appraisal criteria for the UK National Screening Committee (2018).
- Haroon, S., Jordan, R., Takwoingi, Y. & Adab, P. Diagnostic accuracy of screening tests for COPD: a systematic review and meta-analysis. *BMJ Open* **5**, e008133 (2015).
- Ronaldson, S. J. et al. Determining the optimal approach to identifying individuals with chronic obstructive pulmonary disease: the DOC study. *J. Eval. Clin. Pract.* **24**, 487–495 (2018).
- Enright, P. L., CONNETT, J. E. & BAILEY, W. C. The FEV₁/FEV₆ predicts lung function decline in adult smokers. *Respir. Med.* **96**, 444–449 (2002).
- Jing, J. Y., Huang, T. C., Cui, W., Xu, F. & Shen, H. H. Should FEV₁/FEV₆ replace FEV₁/FVC ratio to detect airway obstruction? A metaanalysis. *Chest* **135**, 991–998 (2009).
- Lamprecht, B. et al. Six-second spirometry for detection of airway obstruction: a population-based study in Austria. *Am. J. Respir. Crit. Care Med.* **176**, 460–464 (2007).
- Melbye, H., Medbo, A. & Crockett, A. The FEV₁/FEV₆ ratio is a good substitute for the FEV₁/FVC ratio in the elderly. *Prim. Care Respir. J.* **15**, 294–298 (2006).
- Labor, M., Vrbica, Z., Gudelj, I., Labor, S. & Plavec, D. Diagnostic accuracy of a pocket screening spirometer in diagnosing chronic obstructive pulmonary disease in general practice: a cross sectional validation study using tertiary care as a reference. *BMC Fam. Pract.* **17**, 112 (2016).
- Represas-Represas, C. et al. Screening for chronic obstructive pulmonary disease: validity and reliability of a portable device in non-specialized healthcare settings. *PLoS ONE* **11**, e0145571 (2016).
- Frith, P. et al. Simplified COPD screening: validation of the PiKo-6 in primary care. *Prim. Care Respir. J.* **20**, 190–198 (2011).
- van den Bemt, L. et al. Diagnostic accuracy of pre-bronchodilator FEV₁/FEV₆ from microspirometry to detect airflow obstruction in primary care: a randomised cross-sectional study. *NPJ Prim. Care Respir. Med.* **24**, 14033 (2014).
- Franco-Marina, F. et al. Efficient screening for COPD using three steps: a cross-sectional study in Mexico City. *NPJ Prim. Care Respir. Med.* **24**, 14002 (2014).
- Kim, J. K. et al. Active case finding strategy for chronic obstructive pulmonary disease with handheld spirometry. *Medicine (Baltimore)* **95**, e5683 (2016).
- Thorn, J. et al. Improved prediction of COPD in at-risk patients using lung function pre-screening in primary care: a real-life study and cost-effectiveness analysis. *Prim. Care Respir. J.* **21**, 159–166 (2012).
- Kaufmann, M., Hartl, S., Geyer, K., Breyer, M. K. & Burghuber, O. C. Measuring FEV₁ (6) for detecting early airway obstruction in the primary care setting. Quality and utility of the new PiKo-6 device. *Respiration* **78**, 161–167 (2009).
- Sichletidis, L. et al. A combination of the IPAG questionnaire and PiKo-6(R) flow meter is a valuable screening tool for COPD in the primary care setting. *Prim. Care Respir. J.* **20**, 184–189 (2011).
- National Institute for Health and Care Excellence. *Chronic Obstructive Pulmonary Disease in over 16 s: Diagnosis and Management*. Report No. ng115 (National Institute for Health and Care Excellence, 2018).
- Miller, M. R. & Levy, M. L. Chronic obstructive pulmonary disease: missed diagnosis versus misdiagnosis. *BMJ* **351**, h3021 (2015).
- van Dijk, W. et al. Clinical relevance of fixed ratio vs lower limit of normal of FEV₁/FVC in COPD: patient-reported outcomes from the CanCOLD cohort. *Ann. Fam. Med.* **13**, 41–48 (2015).
- American Thoracic Society. Lung function testing: selection of reference values and interpretative strategies. *Am. Rev. Respir. Dis.* **144**, 1202–1218 (1991).
- Miller, M. R. et al. Clinical characteristics of patients newly diagnosed with COPD by the fixed ratio and lower limit of normal criteria: a cross-sectional analysis of the TargetCOPD trial. *Int. J. Chronic Obstr. Pulm. Dis.* **13**, 1979–1986 (2018).
- Schermer, T. R. et al. Accuracy and precision of desktop spirometers in general practices. *Respiration* **83**, 344–352 (2012).
- Marshall, M., Jackson, J. & Cooper, B. G. Does it matter which type of spirometer is used to measure FEV₁, FVC and VC? *Thorax* **54**, A78 (1999).
- Duong-Quy, S. et al. Early detection of smoking related chronic obstructive pulmonary disease in Vietnam. *Rev. Mal. Respir.* **26**, 267–274 (2009).
- Poels, P. J. et al. Variation in spirometry utilization between trained general practitioners in practices equipped with a spirometer. *Scand. J. Prim. Health Care* **24**, 81–87 (2006).
- Adab, P. et al. Cohort profile: the Birmingham Chronic Obstructive Pulmonary Disease (COPD) cohort study. *Int. J. Epidemiol.* <https://doi.org/10.1093/ije/dyv350> (2016).

38. Jordan, R. E. et al. Targeted case finding for chronic obstructive pulmonary disease versus routine practice in primary care (TargetCOPD): a cluster-randomised controlled trial. *Lancet Respir. Med.* **4**, 720–730 (2016).
39. Hankinson, J. L., Odencrantz, J. R. & Fedan, K. B. Spirometric reference values from a sample of the general U.S. population. *Am. J. Respir. Crit. Care Med.* **159**, 179–187 (1999).
40. Miller, M. R. et al. ATS/ERS Task Force: standardisation of spirometry. *Eur. Respir. J.* **26**, 319–338 (2005).
41. Mercaldo, N. D., Lau, K. F. & Zhou, X. H. Confidence intervals for predictive values with an emphasis to case-control studies. *Stat. Med.* **26**, 2170–2183 (2007).
42. Bossuyt, P. M. et al. STARD 2015: an updated list of essential items for reporting diagnostic accuracy studies. *BMJ* **351**, h5527 (2015).

ACKNOWLEDGEMENTS

We acknowledge the patients who participated in our study, as well as the support of their general practices. We also thank our patient advisory group and programme steering committee. We also acknowledge Professor MR Miller and Professor BG Cooper for their involvement in the spirometry training and quality control elements of this study. Finally, we acknowledge all BLISS team members for facilitating the study and conducting the assessment visits. This paper summarises independent research funded by the National Institute for Health Research (NIHR) School for Primary Care Research (project no. 124) and the NIHR Programme Grants for Applied Research Programme (grant reference number RP-PG-0109-10061). The views expressed are those of the authors and not necessarily those of the NIHR or the Department of Health and Social Care. The Birmingham COPD Cohort study and TargetCOPD are part of the Birmingham Lung Improvement Studies: BLISS.

AUTHOR CONTRIBUTIONS

D.A.F. was the PI for the case–control study, and P.A. and D.A.F. are co-PIs for the BLISS programme. The case–control study was conceived by R.E.J., P.A. and D.A.F. with contributions from other BLISS co-investigators, while the methodology was developed by R.D.R., R.E.J. and P.A. A.E. oversaw programme management, and R.D.R. and A.S. advised on statistical aspects. A.D. recruited the general practices, supervised data collection and conducted analyses, supervised by A.S. A.D. wrote the first draft

with contribution from R.E.J. R.E.J. is guarantor for the paper. All authors reviewed and approved the final submission.

COMPETING INTERESTS

The authors declare no competing interests.

ADDITIONAL INFORMATION

Supplementary information is available for this paper at <https://doi.org/10.1038/s41533-019-0158-2>.

Correspondence and requests for materials should be addressed to A.P.D. or P.A.

Reprints and permission information is available at <http://www.nature.com/reprints>

Publisher's note Springer Nature remains neutral with regard to jurisdictional claims in published maps and institutional affiliations.



Open Access This article is licensed under a Creative Commons Attribution 4.0 International License, which permits use, sharing, adaptation, distribution and reproduction in any medium or format, as long as you give appropriate credit to the original author(s) and the source, provide a link to the Creative Commons license, and indicate if changes were made. The images or other third party material in this article are included in the article's Creative Commons license, unless indicated otherwise in a credit line to the material. If material is not included in the article's Creative Commons license and your intended use is not permitted by statutory regulation or exceeds the permitted use, you will need to obtain permission directly from the copyright holder. To view a copy of this license, visit <http://creativecommons.org/licenses/by/4.0/>.

© The Author(s) 2020