

# Myotonometry for the evaluation of Achilles tendon mechanical properties

Schneebeili, Alessandro; Falla, Deborah; Clijisen, Ron; Barbero, Marco

DOI:

[10.1136/bmjsem-2019-000726](https://doi.org/10.1136/bmjsem-2019-000726)

License:

Creative Commons: Attribution-NonCommercial (CC BY-NC)

*Document Version*

Publisher's PDF, also known as Version of record

*Citation for published version (Harvard):*

Schneebeili, A, Falla, D, Clijisen, R & Barbero, M 2020, 'Myotonometry for the evaluation of Achilles tendon mechanical properties: a reliability and construct validity study', *BMJ Open Sport & Exercise Medicine*, vol. 6, no. 1, e000726. <https://doi.org/10.1136/bmjsem-2019-000726>

[Link to publication on Research at Birmingham portal](#)

## General rights

Unless a licence is specified above, all rights (including copyright and moral rights) in this document are retained by the authors and/or the copyright holders. The express permission of the copyright holder must be obtained for any use of this material other than for purposes permitted by law.

- Users may freely distribute the URL that is used to identify this publication.
- Users may download and/or print one copy of the publication from the University of Birmingham research portal for the purpose of private study or non-commercial research.
- User may use extracts from the document in line with the concept of 'fair dealing' under the Copyright, Designs and Patents Act 1988 (?)
- Users may not further distribute the material nor use it for the purposes of commercial gain.

Where a licence is displayed above, please note the terms and conditions of the licence govern your use of this document.

When citing, please reference the published version.

## Take down policy

While the University of Birmingham exercises care and attention in making items available there are rare occasions when an item has been uploaded in error or has been deemed to be commercially or otherwise sensitive.

If you believe that this is the case for this document, please contact [UBIRA@lists.bham.ac.uk](mailto:UBIRA@lists.bham.ac.uk) providing details and we will remove access to the work immediately and investigate.

# Myotonometry for the evaluation of Achilles tendon mechanical properties: a reliability and construct validity study

Alessandro Schneebeli <sup>1,2</sup>, Deborah Falla <sup>2</sup>, Ron Clijisen <sup>1,3</sup>, Marco Barbero <sup>1</sup>

**To cite:** Schneebeli A, Falla D, Clijisen R, *et al.* Myotonometry for the evaluation of Achilles tendon mechanical properties: a reliability and construct validity study. *BMJ Open Sport & Exercise Medicine* 2020;**6**:e000726. doi:10.1136/bmjsem-2019-000726

Accepted 24 January 2020

## ABSTRACT

**Objective** This study evaluates the intra-rater and inter-rater reliability of the MyotonPRO and its construct validity for the assessment of Achilles tendon stiffness.

**Design** Reliability and construct validity study.

**Methods** Forty healthy participants were assessed using the MyotonPRO by two raters on two different occasions. Tendon was evaluated in three different positions (relaxed, 0° plantarflexion and standing) and during different isometric contractions (range 0–3 kg). Reliability was calculated using intraclass correlation coefficient (ICC and 95% CI) standard error of measurement and minimal detectable change. Construct validity was evaluated between the different positions and the different contraction intensities using Friedman test.

**Results** Intra-rater reliability was very high ICC<sub>2,k</sub> 0.87–0.98. The reliability of the 0.5 kg contraction was moderate with an ICC<sub>2,k</sub> of 0.59. Inter-rater reliability ranged from high to very high with an ICC<sub>2,k</sub> of 0.76–0.86. The reliability of the 0.5 kg, 1 kg contraction and the standing position was moderate with an ICC<sub>2,k</sub> of 0.55, 0.54 and 0.56 respectively. Inter-session reliability ranged from high to very high with an ICC<sub>2,k</sub> of 0.70–0.89. The reliability of the 0.5 kg contraction was moderate with an ICC<sub>2,k</sub> of 0.54. Construct validity was demonstrated between different contraction levels and different positions.

**Conclusion** MyotonPRO is a reliable tool for the evaluation of Achilles tendon stiffness during different contraction levels and in different positions. Construct validity was supported by changes of tendon stiffness during the explored conditions. MyotonPRO can be implemented, as a ready to use device, in the evaluation of tendon tissue mechanical properties.

## INTRODUCTION

The Achilles tendon (AT) is one of the most commonly injured structures in the lower limb, both in athletes and the general population.<sup>1 2</sup> Tendinopathy of AT can alter the mechanical, material and morphological properties of the tendon structure with a decrease of stiffness and young modulus and an increase of tendon cross-sectional area and diameter.<sup>3 4</sup>

Healthy tendon is highly responsiveness to mechanical loading.<sup>5</sup> Adaptation of tendon

## Summary box

- ▶ Achilles tendon stiffness can be measured with MyotonPRO. High intra-rater, inter-rater and inter-session reliability has been found in different position and during isometric contractions.
- ▶ The interest in measuring tendon stiffness is growing for its clinical and research relevance. A reliable and ready to use device has been presented in this study. MyotonPRO could be implemented in the evaluation of Achilles tendon mechanical properties.
- ▶ Stiffness of healthy Achilles tendons has been assessed, reliability in patient with tendinopathies could be different.

has been shown following different loading protocols, and the intensity of the exercise (ie, high loading magnitude) and the intervention duration (>12 weeks) appear to contribute the most to changes in the mechanical properties of the tendon. The change in tendon stiffness appears to be largely due to adaptations in the material rather than morphological properties of the tendon.<sup>5</sup> That is, early changes in young modulus is apparent following loading protocols whereas changes in tendon hypertrophy appear as longer-term adaptations to mechanical loading.<sup>6</sup>

Tendon responses to loading, unloading and ageing have been evaluated in a recent review revealing adaptability of the tendon tissue with an increase in stiffness when loaded and reduced stiffness when unloaded.<sup>3</sup> Nevertheless, conflicting results has been reported when evaluating the in vivo effects of ageing on tendon stiffness, with some authors suggesting a decrease in modulus<sup>3 7–11</sup> and others reporting unchanged tendon stiffness with ageing.<sup>12 13</sup>

Given the clinical and research interest in evaluate tendon stiffness as a parameter to describe tendon behaviour and tendon health, different measurement techniques have been developed. Once recent, non-invasive, ready to use technology is the



© Author(s) (or their employer(s)) 2020. Re-use permitted under CC BY-NC. No commercial re-use. See rights and permissions. Published by BMJ.

<sup>1</sup>Rehabilitation Research Laboratory 2rLab, Department of Business Economics, Health and Social Care, University of Applied Sciences and Arts of Southern Switzerland, Manno/Landquart, Switzerland

<sup>2</sup>Centre of Precision Rehabilitation for Spinal Pain (CPR Spine), School of Sport, Exercise Rehabilitation Sciences, College of Life Environmental Sciences, University of Birmingham, Birmingham, UK

<sup>3</sup>University College Physiotherapy, Thim van der Laan AG, Landquart, Switzerland

## Correspondence to

Alessandro Schneebeli; alessandro.schneebeli@sups.ch



MyotonPRO (Muomeetria, Tallinn, Estonia). This device is a handheld myotonometer producing a mechanical impulse to the skin overlying the target structure. The oscillation of the tissues underneath the probe permits calculation of the viscoelastic properties of the tissue.<sup>14–16</sup> One parameter that can be extracted with this device is dynamic stiffness (N/m), which has been used in different studies to characterise the behaviour of different musculoskeletal tissues. This device has recently been used to evaluate tendon stiffness<sup>17–20</sup> with a recent work showing decreased stiffness over the tendinopathic region of the tendon in people with Achilles tendinopathy and a significant difference in tendon stiffness between weight-bearing and non-weight-bearing positions.

The reliability of the MyotonPRO has been demonstrated for the evaluation of muscle stiffness,<sup>15 21–23</sup> however, the reliability of MyotonPRO for measuring AT stiffness requires further investigation,<sup>17 18 24 25</sup> especially regarding the capability of this device to measure the tendon in different conditions. In contrast to muscle, values of AT stiffness are extremely high and even higher when tested under load.<sup>19</sup> Being able to quantify AT mechanical properties with a ready to use and affordable device which provides a reliable and valid evaluation of the tendon under different conditions, would provide new opportunities to characterise the mechanical behaviour of the AT, for example, in people with Achilles tendinopathy or following an intervention.

This study evaluates the intra-rater and inter-rater reliability of using the MyotonPRO to measure stiffness of asymptomatic AT in different position and during increasing load and to determine the construct validity of this device for measuring tendon stiffness.

## METHODS

The study was conducted between January and July 2019 at the University of Applied Sciences and Arts of Southern Switzerland. The study reporting adheres to the Guidelines for Reporting Reliability and Agreement Studies.<sup>26</sup>

A convenience sample of 40 healthy participants (18 women, 22 men, aged 33.8±14.2 years) were prospectively recruited through public announcement at the University. Their mean (±SD) height and weight were 172.2 (±8.2) cm and 71.7 (±12) kg. Health-related physical activity was monitored using the International Physical Activity Questionnaires Short form (IPAQ)<sup>27</sup>; 23 participants were categorised as high, 15 as moderate and 2 as having low physical activity levels. Participants were excluded if they had a history of tendon or foot injury or surgery; connective tissue, metabolic or endocrine disease. Furthermore, participants with systemic inflammatory disorders, rheumatoid arthritis, spondyloarthritis or hypercholesterolaemia were excluded.

Sample size was estimated according to Walter *et al.*<sup>28</sup> using the following parameters: two replicates,  $p_0=0.6$ ,  $p_1=0.80$ ,  $\alpha=0.05$  and  $\beta=0.2$ . The optimal sample size was 39.1, and we rounded up to 40 participants.

## Procedures

The AT mechanical properties were randomly analysed by two independent operators who were both physiotherapists with a minimum of 4 hours of training with the MyotonPRO prior to the study. Two measurement sessions were planned within 7 days. During the first session, the two operators consecutively performed the measures in order to evaluate the intra-rater and inter-rater reliability. The second session was used to evaluate inter-session reliability.

Testing location and room temperature (21°C–23°C) was kept constant across sessions. Participants were asked to maintain the same activity level between the two testing days.

The resistance of the soft tissues to the deformation induced by the MyotonPRO is calculated from an acceleration graph representing the damped natural oscillation of the tissues. Dynamic stiffness, calculated in N/m, is computed by multiplying the maximum acceleration of the oscillation graph to the mass of the probe divided by the maximum displacement of the tissue; higher values represent a stiffer tissue.


Both the left and right AT were evaluated at the most central point of the tendon following a trajectory from the most central component of the medial malleolus as this point has been suggested as the most prevalent region for tendinopathy.<sup>29</sup> Furthermore, a previous study using MyotonPRO to evaluate stiffness of the AT in people with Achilles tendinopathy and asymptomatic controls showed that changes in mechanical parameters are more likely to be detected from 8 to 12 cm from the plantar aspect of the heel, corresponding to the free tendon part.<sup>19</sup>

MyotonPRO was setup with a multiscan sequence of 5 impulses (1 s apart), a short duration (15 ms) impulse involving minimal mechanical force with a precompression of 0.18 N and an impulse of 0.40 (total=0.58 n) was exerted on the tendon.

Measurements were taken in three different positions; standing, a relaxed position and during isometric contractions of the plantar flexors of increasing intensities, in order to evaluate construct validity (figure 1). As the triceps surae contracts to increase plantarflexor force, the passive tension within the tendon increases as shown in previous studies.<sup>30</sup> Since tendon stiffness and tendon displacement are directly related to plantarflexor strength<sup>31</sup> and ankle angle,<sup>32</sup> myotonometry values are expected to increase with increasing muscle contraction and with increased plantarflexion. To illustrate this construct, stiffness of the AT was compared across the five different contraction levels and between the three different positions. Testing in the three different positions as well as the sequence of the isometric contractions were randomised using a random number table.

## Isometric contractions

The participants were positioned in prone with the foot secured within a dynamometer with the ankle in



**Figure 1** Schematic of the experimental setup. (A) Isometric contractions: the foot was securely attached to a dynamometer with a wooden board at a fixed angle of 0° of dorsiflexion. The force measured by the load cell was proportional to the torque exerted at the ankle level. (B) Standing position: participants were asked to stand on a Wii balance board (Nintendo, Kyoto, Japan) with the centre of mass centred on the centre of board with the feet equally spaced. (C) Relaxed position: the feet were hanging freely over the edge of the examination bed. MyotonPRO was applied to the most central point of the tendon at the level of the medial malleolus.

0° of dorsiflexion/plantarflexion. Different intensities of isometric contractions of the plantar flexors were performed (0, 0.5, 1, 2 and 3 kg). The torque at the ankle joint was evaluated using a dynamometer connected to a force sensor operating in a 0–100 n range (CCT Transducers, Turin, Italy). The force sensor was attached to a wooden board connected to a hinge that enabled natural movement of the ankle joint in dorsiflexion and plantarflexion (figure 1A). The board was then fixed in order to permit isometric contractions only at a defined angle of 0° dorsiflexion. The force detected by the sensor was proportional to the torque exerted at the ankle level. Force signals were amplified using MISO-II (OT Bioelettronica, Turin, Italy) and visual feedback was provided to the participants.

A practice phase prior to the recording was conducted by asking the patient to familiarise with the visual feedback and perform contractions with the instruction to ‘push on the board as if you want to reach a tip-toe position’. During this practice phase, the lift of the heel was controlled. In order to prevent the detachment of the heel from the board, the foot and ankle were tightly secured in the dynamometer with two belts.

### Standing

For the analysis in the standing position, participants were asked to stand on a Wii-Balance board (Nintendo, Kyoto, Japan) with the centre of mass on the middle of the board with the feet equally spaced (figure 1B). Visual feedback was provided by the BrainBLOX Software (University of Colorado, Boulder, Colorado, USA) to the participant in order to ensure that the centre of mass remained in the same position. Participants were asked to rest their hands on a fixed box which ensured the stability of the standing position and avoided oscillation of the centre of mass during the measurement that could affect the tension of the triceps surae muscle-tendon complex.

### Relaxed position

The measurements in the relaxed position were performed with the participants in prone on an examination bed and the feet hanging freely over the edge (figure 1C). The participants were instructed not to move their feet during the measurement.

### Statistical analysis

Statistical Package for the Social Sciences (SPSS V.24) was used to perform statistical analysis. Descriptive statistics (mean±SD) were used to describe the demographic characteristics of the sample. The distribution of the values was assessed using the Shapiro-Wilk test. Given the non-normal distribution of the data, non parametric, independent samples, Mann-Whitney test was used to assess whether there were significant differences in AT mechanical properties between men and women, between participants with different physical activity levels based on their IPAQ score.

Intra-rater and inter-rater reliability was assessed using intraclass correlation coefficient ( $ICC_{2,k}$ ) and the absolute reliability was calculated using standard error of measurement (SEM) and minimal detectable change (MDC).<sup>33</sup>

Furthermore, a cut-off point of  $ICC > 0.6$  was used to consider the clinical relevance of the measurements.<sup>34 35</sup>

Construct validity was tested analysing the correlations between the five different contraction levels and the three different positions using Friedman test with post hoc pairwise comparison Bonferroni correction was applied and statistical significance was set to  $\alpha=0.05$ . Data are presented as median and IQR.

### Patient and public involvement statement

This research was done without patient involvement. Patients were not invited to comment on the study design and were not consulted to develop patient relevant outcomes or interpret the results. Patients were not invited to contribute to the writing or editing of this document for readability or accuracy.

### RESULTS

No significant difference in AT stiffness was observed between men and women ( $p=0.6$ ) or between participants with different physical activity levels based on their IPAQ score ( $p=0.9$ ), therefore, the data were pooled and further analysed across the sample. Table 1 reports the descriptive statistics for the sample in the different conditions.

A total of eighty AT of 40 healthy participants were assessed by the two physical therapists.

Intra-rater reliability was very high with an  $ICC_{2,k}$  of 0.95 for 0 kg, 0.87 for 1 kg, 0.95 for 2 kg, 0.94 for 3 kg, and 0.98 and 0.87 for the relaxed and standing positions respectively. The intra-rater reliability of measuring AT stiffness during the 0.5 kg contraction was moderate with an  $ICC_{2,k}$  of 0.59.

**Table 1** Descriptive statistics

	N	Sample (mean±SD)			Stiffness N/m (median IQR)						
		Age (year)	Height (cm)	Weight (kg)	0 kg	0.5 kg	1 kg	2 kg	3 kg	Relax	Standing
Total	40	33.8	172.2	71.7	1043.0	1096.5	1110.5	1180.5	1381.5	771.0	1596.0
		14.2	8.2	12.0	132.3	62.0	121.3	255.5	214.3	105.3	339.3
Men	22	35.6	177.7	77.4	1057.5	1089.5	1086.5	1117.5	1333.0	792.0	1577.5
		13.8	5.6	10.6	144.0	65.0	132.3	143.3	198.3	124.0	337.3
Women	18	32.4	167.6	67.0	1041.0	1099.5	1129.5	1284.0	1435.5	753.0	1616.5
		15.1	6.9	11.0	139.0	89.5	129.8	279.8	201.3	83.3	363.5

Sample characteristics and stiffness values.

Inter-rater reliability between the two raters ranged from high to very high with an ICC<sub>2,k</sub> of 0.76 for 0 kg, 0.79 for 2 kg, 0.81 for 3 kg and 0.86 for the relaxed position. The inter-rater reliability of measuring AT stiffness during the 0.5 kg, 1 kg contraction and the standing position was moderate with an ICC<sub>2,k</sub> of 0.55, 0.54 and 0.56, respectively.

Inter-session reliability ranged from high to very high with an ICC<sub>2,k</sub> of 0.88 for 0 kg, 0.70 for 1 kg, 0.86 for 2 kg, 0.78 for 3 kg, 0.89 and 0.75 for the relaxed and standing positions, respectively. The inter-session reliability of measuring AT stiffness during the 0.5 kg contraction was moderate with an ICC<sub>2,k</sub> of 0.54.

Table 2 reports the relative reliability with ICC and 95% CI and absolute reliability with SEM and MDC.

Changes in AT stiffness between the different conditions were calculated using Friedman test for related samples. Significant differences (p<0.01) were found between all contraction levels except for 0–0.5 kg (p=0.9); 0.5–1 kg (p=0.9); 1–2 kg (p=0.7) and between the different positions (figure 2).

## DISCUSSION

The present study evaluated the absolute and relative reliability of the myotonometry to measure the stiffness of AT in different conditions. Even though the MyotonPRO was developed to evaluate the stiffness and tone of skeletal muscle, the present study shows the transferability of this technology and that measuring stiffness of tendon

tissue is feasible and reliable. Intra-rater and inter-rater reliability ranged from high to very high in most of the tested conditions. Moreover, stiffness changes between the different contraction intensities and the different positions was detected, demonstrating the construct validity of this device.


One of the conditions in the current study was assessing AT stiffness with the person in a relaxed position with the feet hanging freely over the edge of the examination bed. A previous study<sup>25</sup> also showed high intra-rater reliability when the AT was evaluated in a relaxed position (ICC: 0.80 in healthy participants). A further study evaluating the effect of dorsiflexion of the ankle joint on tendon stiffness<sup>18</sup> showed high to very high intra-reliability and inter-reliability when measured in a relaxed position (ICC >0.88). Our results strengthen these previous observations by confirming high to very high intra and inter-rater reliability (ICC >0.86), now demonstrated in a larger sample. The stiffness values vary slightly between studies with an average of 873 N/m reported by Finnamore *et al.*,<sup>25</sup> 776.11±71.70 reported by Liu *et al.*,<sup>18</sup> and 771±105.3 N/m observed in the present study. Nevertheless, collectively the current and previous findings confirm that MyotonPRO is a reliable device to evaluate AT stiffness in a relaxed position on the same day, between different days and between different operators.

Stiffness values increased from 1043±132.3 N/m when tendon was assessed in a 0° ankle dorsiflexion

**Table 2** Absolute and relative reliability for the different conditions

	Intra-rater			Inter-rater			Inter-session		
	ICC (95% CI)	SEM	MDC	ICC (95% CI)	SEM	MDC	ICC (95% CI)	SEM	MDC
0 kg	0.95 (0.92 to 0.97)	19.3	53.5	0.76 (0.62 to 0.84)	43.2	119.6	0.88 (0.82 to 0.93)	30.5	84.6
0.5 kg	0.59 (0.36 to 0.74)	40	111	0.55 (0.30 to 0.72)	50.2	139.1	0.54 (0.29 to 0.71)	49.8	137.9
1 kg	0.87 (0.80 to 0.92)	36	99.8	0.54 (0.28 to 0.71)	66.2	183.5	0.70 (0.54 to 0.81)	56.1	155.6
2 kg	0.95 (0.91 to 0.97)	39.3	109	0.79 (0.67 to 0.87)	83.3	231	0.86 (0.78 to 0.91)	65.8	182.5
3 kg	0.94 (0.91 to 0.96)	47.7	132.1	0.81 (0.69 to 0.88)	85.9	238.2	0.78 (0.66 to 0.86)	92.7	256.9
Relax	0.98 (0.97 to 0.99)	10.8	30	0.86 (0.78 to 0.91)	30.5	84.6	0.89 (0.83 to 0.93)	24.7	68.4
Standing	0.87 (0.89 to 0.92)	70.7	196.1	0.56 (0.29 to 0.73)	130.1	360.5	0.75 (0.60 to 0.84)	101.5	281.3

CI, confidence interval; ICC, intraclass correlation coefficient; MDC, minimal detectable change; SEM, standard error of measurement.



**Figure 2** Construct validity. Box plot showing the stiffness median and IQR values of the entire sample for the different contraction levels (A) and for the different positions (B). \* $p < 0.01$ ; statistical significant difference between the different conditions.

position with 0 kg of contraction to  $1381 \pm 214$  N/m for the isometric contraction at 3 kg. The ICC values showed high to very high intra-rater and inter-rater reliability for the majority of the isometric contraction levels (see table 2). Moderate intra-rater and inter-rater reliability was found for the 0.5 kg contraction level and the level of reliability did not exceed the cut-off point of ICC = 0.6 considered as the threshold for clinical relevance.<sup>34</sup> The low reliability for the contraction at 0.5 kg is likely due to the difficulty in maintaining the force feedback stable during such a low contraction of the plantar flexors. A previous study<sup>18</sup> evaluating AT stiffness at different ankle angles showed similar results to the present study with an average stiffness of 1143 N/m at 0° of dorsiflexion compared with  $1043 \pm 132$  N/m found in the current study. Liu *et al*<sup>18</sup> did not examine reliability at different ankle angles, however, the data reported in this article demonstrated a statistically significant increase in tendon stiffness due to the change in ankle dorsiflexion angle from 1143 N/m at 0° to 1329 N/m at 30°. These results are in line with the present study, which showed a significant increase in tendon stiffness with contractions of increasing intensity. Thus, the MyotonPRO can reliably detect changes in AT stiffness whether the tendon is stretched from contraction of the triceps surae muscle or stretched as a result of a change in the angle of the ankle.

Another study<sup>20</sup> analysing the tissue properties of the AT and plantar fascia in healthy volunteers found slightly different stiffness values when the tendon was measured at 0° of dorsiflexion. The article reports a mean ( $\pm$ SD) of 848 ( $\pm 95.73$ ) N/m and 847.3 ( $\pm 101.5$ ) N/m for the left and right AT, respectively, whereas we observed higher values with an average stiffness of  $1043 \pm 132$  N/m.

Tendon structural and mechanical adaptations are strongly related to the load magnitude exerted to the musculotendinous complex<sup>3 5</sup> and the majority of interventions targeting pathological tendon tissue are

loading-based exercise regimens.<sup>36–40</sup> The current results support the evaluation of the tendon tissue behaviour during load with a reliable and easy to use device could be helpful to detect possible mechanical adaptation following different rehabilitation protocol.

Stiffness of the AT was also evaluated in standing position in this study to verify the feasibility of MyotonPRO to analyse the tendon in a weight-bearing position. Intra-rater reliability was very high and inter-session reliability was high. Conversely inter-rater reliability in the standing position was moderate. It is important to note, however, that in the standing position the average stiffness values were 1596 ( $\pm 339$ ) N/m which is approaching the upper limit that the device is able to detect. The lower reliability could be a consequence of the measurement threshold of the device such that when it is close to the higher limit, there were difficulties in providing a mechanical impulse strong enough to give a proper oscillation wave. Morgan *et al*<sup>19</sup> found similar results in a population of healthy participants in a weight-bearing position. Despite the possible technical limitation of analysing tendon stiffness in a weight-bearing position, the results demonstrate good intra-rater reliability but only moderate inter-rater reliability.

This is the first study to evaluate changes in AT stiffness between different contraction levels using the MyotonPRO. The results confirm a significant difference in tendon stiffness between the different contraction levels. Although stiffness changes can be detected across different contractions, the levels with similar intensity (0–0.5 kg ( $p = 0.9$ ); 0.5–1 kg ( $p = 0.9$ ); 1–2 kg ( $p = 0.7$ )) showed similar stiffness values likely as the difference between consecutive contraction levels was not sufficient to change the mechanical properties of the tendon. These results are consistent with those reported in a previous sonoelastography evaluation<sup>41</sup> showing an increase of tendon elastography colour scale index when the AT was

gradually put under tension by isometric contractions of increasing intensity. In further support of the construct validity, pairwise comparisons showed a difference in AT stiffness between the different positions (relaxed, 0° ankle angle and standing). Similarly, a previous study using shear wave sonoelastography<sup>42</sup> showed an increase in wave velocity, indicating a hardening of the AT, when the ankle angle changed from a plantarflexed position to 0° of dorsiflexion.

There are some limitations of this study which should be noted. Healthy participants were recruited and stiffness was evaluated in only one point of the tendon and it is therefore unknown whether variation may exist in different regions of the tendon. Another potential limitation was the selection of the contraction intensities. The upper limit of this ramp was chosen because after a certain amount of load (approximately 5 kg) the device cannot detect stiffness values. The small changes between the different contraction levels could be not sufficient to fully describe the behaviour of the tendon during load.

## CONCLUSIONS

MyotonPRO is a reliable tool to measure AT stiffness in different conditions including when the tendon is in a relaxed position and under load. Reliability is confirmed for the same rater, between different days and between two raters. Construct validity was supported by changes of tendon stiffness during the explored contractions and ankle positions. MyotonPRO can be implemented, as a ready to use device, in the evaluation of tendon tissue mechanical properties.

**Acknowledgements** We thank the students involved in the present study at the University of Applied Sciences and Arts of Southern Switzerland (SUPSI). We especially thank the Thim van der Laan Foundation, Landquart, Switzerland for financial support.

**Contributors** AS, DF, RC and MB designed and planned the study. AS and MB collected the data. All authors were involved in the drafting and revision of the manuscript and approved the final version of the article.

**Funding** The authors have not declared a specific grant for this research from any funding agency in the public, commercial or not-for-profit sectors.

**Competing interests** None declared.

**Patient consent for publication** Not required.

**Ethics approval** The study was approved by the Ethics Committee of Canton Ticino (ID:2018-01483/CE3391) and written informed consent was provided by the participants.

**Provenance and peer review** Not commissioned; externally peer reviewed.

**Data availability statement** Data are available on reasonable request.

**Open access** This is an open access article distributed in accordance with the Creative Commons Attribution Non Commercial (CC BY-NC 4.0) license, which permits others to distribute, remix, adapt, build upon this work non-commercially, and license their derivative works on different terms, provided the original work is properly cited, appropriate credit is given, any changes made indicated, and the use is non-commercial. See: <http://creativecommons.org/licenses/by-nc/4.0/>.

## ORCID iDs

Alessandro Schneebeli <http://orcid.org/0000-0002-8411-2012>  
 Deborah Falla <http://orcid.org/0000-0003-1689-6190>  
 Ron Clijsen <http://orcid.org/0000-0001-7125-5332>  
 Marco Barbero <http://orcid.org/0000-0001-8579-0686>

## REFERENCES

- de Jonge S, van den Berg C, de Vos RJ, et al. Incidence of midportion Achilles tendinopathy in the general population. *Br J Sports Med* 2011;45:1026–8.
- Józsa L, Kvist M, Bálint BJ, et al. The role of recreational sport activity in Achilles tendon rupture. A clinical, pathoanatomical, and sociological study of 292 cases. *Am J Sports Med* 1989;17:338–43.
- Magnusson SP, Kjaer M. The impact of loading, unloading, ageing and injury on the human tendon. *J Physiol* 2019;597:1283–98.
- Obst SJ, Heales LJ, Schrader BL, et al. Are the mechanical or material properties of the Achilles and Patellar tendons altered in tendinopathy? A systematic review with meta-analysis. *Sports Med* 2018;48:2179–98.
- Bohm S, Mersmann F, Arampatzis A. Human tendon adaptation in response to mechanical loading: a systematic review and meta-analysis of exercise intervention studies on healthy adults. *Sports Med Open* 2015;1:7.
- Kjaer M, Langberg H, Heinemeier K, et al. From mechanical loading to collagen synthesis, structural changes and function in human tendon. *Scand J Med Sci Sports* 2009;19:500–10.
- Couppé C, Svensson RB, Grosset J-F, et al. Life-Long endurance running is associated with reduced glycation and mechanical stress in connective tissue. *Age* 2014;36:9665.
- Kubo K, Kanehisa H, Miyatani M, et al. Effect of low-load resistance training on the tendon properties in middle-aged and elderly women. *Acta Physiol Scand* 2003;178:25–32.
- Onambele GL, Narici MV, Maganaris CN. Calf muscle-tendon properties and postural balance in old age. *J Appl Physiol* 2006;100:2048–56.
- Quinlan JI, Maganaris CN, Franchi MV, et al. Muscle and tendon contributions to reduced rate of torque development in healthy older males. *J Gerontol A Biol Sci Med Sci* 2018;73:539–45.
- Stenroth L, Peltonen J, Cronin NJ, et al. Age-Related differences in Achilles tendon properties and triceps surae muscle architecture in vivo. *J Appl Physiol* 2012;113:1537–44.
- Carroll CC, Dickinson JM, Haus JM, et al. Influence of aging on the in vivo properties of human Patellar tendon. *J Appl Physiol* 2008;105:1907–15.
- Couppé C, Hansen P, Kongsgaard M, et al. Mechanical properties and collagen cross-linking of the Patellar tendon in old and young men. *J Appl Physiol* 2009;107:880–6.
- Agyapong-Badu S, Warner M, Samuel D, et al. Measurement of ageing effects on muscle tone and mechanical properties of rectus femoris and biceps brachii in healthy males and females using a novel hand-held myometric device. *Arch Gerontol Geriatr* 2016;62:59–67.
- Fröhlich-Zwahlen AK, Casartelli NC, Item-Glatthorn JF, et al. Validity of resting myotonometric assessment of lower extremity muscles in chronic stroke patients with limited hypertonia: a preliminary study. *J Electromyogr Kinesiol* 2014;24:762–9.
- Gervasi M, Sisti D, Amatori S, et al. Muscular viscoelastic characteristics of athletes participating in the European master indoor athletics Championship. *Eur J Appl Physiol* 2017;117:1739–46.
- Feng YN, Li YP, Liu CL, et al. Assessing the elastic properties of skeletal muscle and tendon using shearwave ultrasound elastography and MyotonPRO. *Sci Rep* 2018;8:17064.
- Liu CL, Li YP, Wang XQ, et al. Quantifying the stiffness of Achilles tendon: intra- and Inter-Operator reliability and the effect of ankle joint motion. *Med Sci Monit* 2018;24:4876–81.
- Morgan GE, Martin R, Williams L, et al. Objective assessment of stiffness in Achilles tendinopathy: a novel approach using the MyotonPRO. *BMJ Open Sport Exerc Med* 2018;4:e000446.
- Orner S, Kratzer W, Schmidberger J, et al. Quantitative tissue parameters of Achilles tendon and plantar fascia in healthy subjects using a handheld myotonometer. *J Bodyw Mov Ther* 2018;22:105–11.
- Ortega-Cebrian S, Luchini N, Whiteley R. Dry needling: effects on activation and passive mechanical properties of the quadriceps, pain and range during late stage rehabilitation of ACL reconstructed patients. *Phys Ther Sport* 2016;21:57–62.
- Bizzini M, Mannion AF. Reliability of a new, hand-held device for assessing skeletal muscle stiffness. *Clin Biomech* 2003;18:459–61.
- Kelly JP, Koppenhaver SL, Michener LA, et al. Characterization of tissue stiffness of the infraspinatus, erector spinae, and gastrocnemius muscle using ultrasound shear wave elastography and superficial mechanical deformation. *J Electromyogr Kinesiol* 2018;38:73–80.
- Sohrad S, Wilson D, Waugh C, et al. Feasibility of using a hand-held device to characterize tendon tissue biomechanics. *PLoS One* 2017;12:e0184463.

- 25 Finnamore E, Waugh C, Solomons L, *et al.* Transverse tendon stiffness is reduced in people with Achilles tendinopathy: a cross-sectional study. *PLoS One* 2019;14:e0211863.
- 26 Kottner J, Audigé L, Brorson S, *et al.* Guidelines for reporting reliability and agreement studies (GRRAS) were proposed. *J Clin Epidemiol* 2011;64:96–106.
- 27 Craig CL, Marshall AL, Sjöström M, *et al.* International physical activity questionnaire: 12-country reliability and validity. *Med Sci Sports Exerc* 2003;35:1381–95.
- 28 Walter SD, Eliasziw M, Donner A. Sample size and optimal designs for reliability studies. *Stat Med* 1998;17:101–10.
- 29 Mahieu NN, Witvrouw E, Stevens V, *et al.* Intrinsic risk factors for the development of Achilles tendon overuse injury: a prospective study. *Am J Sports Med* 2006;34:226–35.
- 30 Kjaer M. Role of extracellular matrix in adaptation of tendon and skeletal muscle to mechanical loading. *Physiol Rev* 2004;84:649–98.
- 31 Magnusson SP, Hansen P, Aagaard P, *et al.* Differential strain patterns of the human gastrocnemius aponeurosis and free tendon, in vivo. *Acta Physiol Scand* 2003;177:185–95.
- 32 Arampatzis A, Karamanidis K, Albracht K. Adaptational responses of the human Achilles tendon by modulation of the applied cyclic strain magnitude. *J Exp Biol* 2007;210:2743–53.
- 33 Atkinson G, Nevill AM. Statistical methods for assessing measurement error (reliability) in variables relevant to sports medicine. *Sports Med* 1998;26:217–38.
- 34 Chinn S. The assessment of methods of measurement. *Stat Med* 1990;9:351–62.
- 35 Munro BH. *Statistical methods for health care research*. 5th edn. Philadelphia, Penn, London: Lippincott Williams & Wilkins, 2005.
- 36 Silbernagel KG, Thomeé R, Thomeé P, *et al.* Eccentric overload training for patients with chronic Achilles tendon pain—a randomised controlled study with reliability testing of the evaluation methods. *Scand J Med Sci Sports* 2001;11:197–206.
- 37 Silbernagel KG, Thomeé R, Eriksson BI, *et al.* Continued sports activity, using a pain-monitoring model, during rehabilitation in patients with Achilles tendinopathy: a randomized controlled study. *Am J Sports Med* 2007;35:897–906.
- 38 Mafi N, Lorentzon R, Alfredson H. Superior short-term results with eccentric calf muscle training compared to concentric training in a randomized prospective multicenter study on patients with chronic Achilles tendinosis. *Knee Surg Sports Traumatol Arthrosc* 2001;9:42–7.
- 39 Kongsgaard M, Kovanen V, Aagaard P, *et al.* Corticosteroid injections, eccentric decline squat training and heavy slow resistance training in Patellar tendinopathy. *Scand J Med Sci Sports* 2009;19:790–802.
- 40 Alfredson H, Lorentzon R. Chronic Achilles tendinosis: recommendations for treatment and prevention. *Sports Med* 2000;29:135–46.
- 41 Schneebeli A, Del Grande F, Falla D, *et al.* A novel application of strain sonoelastography can detect changes in Achilles tendon elasticity during isometric contractions of increasing intensity. *J Foot Ankle Res* 2019;12:30.
- 42 Aubry S, Nueffer J-P, Tanter M, *et al.* Viscoelasticity in Achilles tendonopathy: quantitative assessment by using real-time shear-wave elastography. *Radiology* 2015;274:821–9.